

HANS DINSLAGE GmbH	Title	EC Declaration of Conformity
	Date	2026-04-23

Manufacturer: Hans Dinslage GmbH (see address in footer)

SRN: DE-MF-000005488

Product category: Thermometers

Product type: SFT 09

Intended use: The thermometer is intended exclusively for the orally, rectally or axillary measurement of human body. The thermometer is only intended for private use in a domestic environment and not for commercial use.

The product specified above is in conformity with the following specifications.

(EU) 2017/745 Medical device regulation (MDR)

Basic-UDI-DI: 4066299SFT09YN

Classification/applied rule(s): Class IIa/rule 10

Conformity assessment procedure: Annex IX, Chapter I

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1326900052, valid to 2030-11-21

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Hans Dinslage GmbH

Place, date of issue: Ulm, 2026-04-23

Name, function, signature, stamp: Max Harder, Team Lead Regulatory Affairs

Hans Dinslage GmbH
Riedlinger Str. 28 • 88524 Uttenweiler