



beurer

IH 57



EN Portable mesh atomiser
Instructions for use

CE 0483



ENGLISH



Read these instructions for use carefully and keep them for later use. Make them accessible to other users and note the information they contain.

CONTENTS

1. Included in delivery	3
2. Signs and symbols	3
3. Intended Purpose	5
4. Warnings and safety notes	7
5. Device and atomiser set description	11
6. Initial use	12
7. Usage	13
8. Cleaning and maintenance	17
9. Accessories and/or replacement parts	21
10. What if there are problems?	21
11. Disposal	23
12. Technical specifications	23
13. Guarantee / Service	25

1. INCLUDED IN DELIVERY

Check that the exterior of the cardboard delivery packaging is intact and make sure that all contents are present. Before use, ensure that there is no visible damage to the device or atomiser set (mouthpiece, masks) and that all packaging material has been removed. If you have any doubts, do not use the device and contact your retailer or the specified Customer Services address.

Nebuliser, Mesh atomiser with medication container (hereinafter referred to as “medication container”), Mouthpiece, PVC adult mask, PVC children’s mask, Micro USB cable, Storage bag, These instructions for use

2. SIGNS AND SYMBOLS

The following symbols appear in these instructions for use.

WARNING

Indicates a potentially impending danger. If it is not avoided, there is a risk of death or serious injury.

CAUTION

Indicates a potentially impending danger. If it is not avoided, slight or minor injuries may result.

NOTICE

Indicates a potentially harmful situation. If it is not avoided, the device or something in its vicinity may be damaged.

The following symbols are used on the packaging and on the type plate for the device and atomiser set.



Observe the instructions

Read the instructions before starting work and/or operating devices or machines



CE labelling

This product satisfies the requirements of the applicable European and national directives.



Type BF applied part



Marking to identify the packaging material. A = Material code, B = Material number: 1-7 = Plastics, 20-22 = Paper and cardboard



Separate the product and packaging elements and dispose of them in accordance with local regulations.



The electronic device must not be disposed of with household waste



Do not dispose of batteries containing hazardous substances with household waste



Manufacturer



Date of manufacture



On/Off



Medical device










Item number



Serial number



Unique Device Identifier (UDI) for unique product identification

	Type number
	Device protected against foreign objects ≥ 12.5 mm and against water dripping at an angle
	Protection class II device
	Temperature limit
	Humidity limitation
	Swiss authorised representative
	Importer

3. INTENDED PURPOSE

Intended Use

Nebulisers (including compressor, ultrasonic and mesh nebulisers) are medical devices for the nebulisation of liquids and liquid medication (aerosols). This device produces aerosols by combining an oscillating mesh with holes and a liquid medication.

The aerosol treatment is suitable for treating the upper and lower airways.

By nebulising and inhaling the medication prescribed/recommended by your doctor, you can prevent diseases affecting the airways, or in the case that you contract such an illness, you can alleviate symptoms and speed up your recovery.

Intended users

The nebuliser is intended for use in the home healthcare environment, not in professional healthcare facilities.

The nebuliser can be used on persons over 2 years of age under supervision; it can be used for self-treatment by persons over 12 years of age.

The shape and size of the person's face determines whether or not the device can be used under supervision. As such, it may not be possible to use the device under supervision until they are older. When using a mask to inhale, take care to ensure the mask fits well and the eyes are unobstructed.

Clinical benefits

Inhalation is the most effective way to administer medication for most respiratory system conditions.

The benefits are:

- The medication is carried directly to the target organs
- The local bioavailability of the medication is significantly increased
- Systemic diffusion is reduced substantially
- Only very low doses of the medication are required
- Quick and effective therapeutic effect
- Side effects are significantly reduced compared to systemic administration
- Humidification of the airways
- Loosening and making (bronchial) secretions more liquid
- Releasing bronchospasms (spasmodic)
- Relieving swollen and inflamed bronchial mucosa
- Coughing up secretions
- Fighting viruses that affect the upper and lower airways

Indication

The nebuliser can be used for diseases of the upper and/or lower airways.

Examples of upper respiratory system diseases include:

- Nasal mucosal inflammation
- Allergic nasal mucosal inflammation

- Nasal sinus infection
- Pharyngitis
- Laryngitis

Examples of lower respiratory system diseases include:

- Bronchial asthma
- Bronchitis
- COPD (chronic obstructive pulmonary disease)
- Bronchiectasis
- Acute tracheobronchitis
- Cystic fibrosis
- Pneumonia

Contraindications

- The atomiser is not intended for the treatment of life-threatening conditions.
- This device must not be used by children under the age of 12 and by people with reduced physical, sensory (e.g. reduced sensitivity to pain) or mental abilities or a lack of experience or knowledge, unless they are supervised or have been instructed on how to use the device safely, and are fully aware of the resulting risks of use.
- Do not use the device on persons who are ventilated and/or unconscious.

- Check whether there are contraindications for use with the usual systems for aerosol treatment on the medication instruction leaflet.
- If the device does not work properly, or if you feel unwell or experience pain, stop using it immediately.

4. WARNINGS AND SAFETY NOTES

WARNING

- The atomiser has no significant effect on the efficacy and safety of the medication administered and is not intended for the treatment of life-threatening conditions.
- After proper preparation, the device can be reused. This preparation includes the replacement of all components, including the atomiser. We recommend that you replace the medication container and other components after one year at the latest. This replacement period is based on the device being used for max. 3 cycles per day and for ≤ 10 min per cycle, and provided it is regularly disinfected and cleaned in accordance with the instructions for use (8. Cleaning and maintenance). For hygiene reasons, it is essential that every user uses their own components.
- Always follow the instructions of your doctor or pharmacist regarding the type of medication to be used, the dosage and the frequency and duration of inhalation. When using with medications/medication nebulisation, observe the conditions and restrictions applicable to these medications. Notice: Use only the medications prescribed by your doctor for treatment according to the respective diagnosis.
- The device is not a substitute for medical consultation and treatment. Always consult your doctor first in the event of pain or illness.
- If you have health concerns of any kind, consult your GP!
- If the liquid you wish to use is not compatible with steel, PP, PC, silicone and PVC, do not use this liquid with our nebuliser. If the information provided with the liquid does not indicate whether the liquid is compatible with these materials, please contact the liquid manufacturer. Examples of compatible liquids include salbutamol solution and sodium chloride solution.
- The particle size distribution curve, MMAD, aerosol delivery and/or aerosol delivery rate may change if a different medication, suspension, emulsion or highly viscous solution is used than the one referenced in the technical specifications for the aerosol values.

- Please note the general hygiene measures when using the nebuliser.
- Before use, ensure that there is no visible damage to the device or components. If you have any doubts, do not use the device and contact your retailer or the specified Customer Services address.
- No modifications may be made to the device or its components.
- Use only accessories recommended by the manufacturer.
- The nebuliser may only be operated with compatible Beurer atomisers and with the appropriate Beurer components. The use of atomisers and components made by other companies may result in less efficient treatment and could damage the device.
- Keep the device away from your eyes when it is in use, as the mist of medication could be harmful.
- Never use the device near flammable gases, oxygen or nitrogen oxide.
- Store the device and components out of the reach of children and pets.
- Keep packaging material away from children (risk of suffocation).
- The device must be switched off and the plug of the micro USB cable pulled out before every cleaning and/or maintenance procedure.
- If the device was at the maximum storage temperature, a period of 4 hours is required before using it in order for the device to cool down to a level where it is ready for the intended use. Likewise, if the device was at the minimum storage temperature, a period of 4 hours is required before using it until the device is ready for the intended use.
- Never submerge the device in water and do not use it in the bathroom. Under no circumstances may liquid enter the device.
- If the device is dropped, exposed to high levels of moisture or suffers any other damage, it must no longer be used. If in doubt, contact Customer Services or the retailer.
- Power cuts, sudden interference or other unfavourable conditions could lead to the device becoming inoperable. We, therefore, recommend that you obtain a replacement device or medication (the latter should be agreed with your doctor).
- To avoid the risk of entanglement and strangulation, store the cable out of the reach of small children.
- The device must only be connected to the mains voltage that is specified on the type plate.
- Never touch the micro USB cable with wet hands, as you could get an electric shock.

- Do not pull the mains adapter out of the socket using the micro USB cable.
- Do not crush or bend the micro USB cable, pull it over sharp-edged objects or leave it dangling down, and protect it from sources of heat.
- We recommend that the micro USB cable is completely unrolled to avoid dangerous overheating.
- If the micro USB cable of this device is damaged, it must be disposed of. Please contact Customer Services or the retailer.
- If the device is opened, there is a risk of electric shock. The power supply network not disconnected unless the adapter is unplugged (and the micro USB cable has no other power connection).
- Store the device in a location protected from climatic effects. The device must be stored in the environmental conditions specified.
- Do not operate the device near electromagnetic transmitters.
- Protect the device from severe knocks.
- After the battery has been successfully charged, disconnect the device from the mains.
- Should you require an adapter or extension lead, it must meet the applicable safety requirements. The power limit and the maximum output specified on the adapter must not be exceeded.

CAUTION

- Never store the device (or the micro USB cable) near to sources of heat.
- Do not use the device in a room in which a spray has previously been used. Air the room before carrying out the treatment.
- Never use the device if it is making an abnormal sound.

Prior to initial use

- Remove all packaging material before using the device.
- Protect the device from dust, dirt and humidity and never cover the device while it is in use.
- Do not operate the device in a very dusty area.
- Switch the device off immediately if it is faulty or not working properly.
- The manufacturer is not liable for damage resulting from improper or incorrect use.

Repairs

- Under no circumstances should you open or repair the device yourself. If you do so, it may

no longer function correctly. Failure to comply with this instruction will void the guarantee.

- The device is maintenance-free.
- For repairs, please contact Customer Services or an authorised retailer.

Notes on handling batteries

WARNING

- **Risk of explosion! Risk of fire!** Failure to comply with the points mentioned can result in personal injury, overheating, leakage, venting, breakage, explosion, or fire.
- Always use the correct or supplied charging cable/charger/mains adapter for charging.
- Avoid continuous charging or overcharging. Unplug the charger when charging is complete.
- Charge the device under supervision, paying attention to any heat generated, deformation, or release of gases. If in doubt, stop charging.
- If batteries/charging cables/chargers are defective, stop using them and dispose of them properly as soon as possible (see chapter on disposal).
- Do not throw the device or batteries into a fire.
- Never forcibly discharge, heat, disassemble, open, crush, deform, encapsulate, modify, or knock the device or batteries.

- Never short-circuit batteries or the connections of the battery-powered device.
- Protect the device or batteries from direct sunlight, rain, heat, and water.
- Exposure of batteries to an environment with extremely high temperatures or an extremely low air pressure may result in explosion or leakage of flammable liquids and gases.
- If fluid from a battery comes into contact with your skin or eyes, wash the affected areas with water and seek medical assistance.

NOTICE

- This device contains a battery that is not replaceable. Once a battery has reached the end of its service life, the device must be disposed of properly (see chapter on disposal).

Notes on electromagnetic compatibility

- The device is suitable for use in all environments listed in these instructions for use, including domestic environments.
- The use of the device may be restricted if electromagnetic disturbance occurs. This could result in issues such as the failure of the device.

- Avoid using this device directly next to other devices or stacked on top of other devices, as this could lead to faulty operation. If, however, it is necessary to use the device in the manner stated, this device as well as the other devices must be monitored to ensure they are working properly.
- The use of an atomiser set other than that specified or provided by the manufacturer of this device can lead to an increase in electromagnetic emissions or a decrease in the device's electromagnetic immunity; this can result in faulty operation.
- Keep portable RF communication devices (including peripheral equipment, such as antenna cables or external antennas) at least 30 cm away from all device parts, including all cables included in delivery.
- Failure to comply with the above could impair the performance of the device.

5. DEVICE AND ATOMISER SET DESCRIPTION

The associated drawings are shown on page 3.

Device overview

- 1 Cover for medication container

- 2 Mesh atomiser with medication container (hereinafter referred to as "medication container")
- 3 Housing
- 4 On/Off switch
- 5 LED indicator
- 6 Button for unlocking the medication container ("PRESS")
- 7 Micro USB port
- 8 Micro USB cable

Atomiser set overview

- 9 PVC adult mask
- 10 PVC children's mask
- 11 Mouthpiece

Atomiser set

Only use the atomiser set recommended by the manufacturer. If this is not respected, the device may not function safely.

⚠ CAUTION

Leakage protection

When pouring the medication into the medication container, ensure that you only fill it up to the maximum mark (8 ml). It is recommended to fill the device up to between 0.5 and 8 ml.

The substance to be nebulised must be in contact with the mesh for nebulisation to take place. If this is not the case, the nebulisation will stop automatically.

So, please try to hold the device as vertically as possible.

6. INITIAL USE

Before using the device for the first time

NOTICE

- Clean and disinfect the medication container [2] and atomiser set before using them for the first time. See section “Cleaning and maintenance” on this process.

Assembly

Take the device out of the packaging.

If the medication container [2] has not been assembled, slide it onto the housing [3] horizontally from the front. **A**

Charging the nebuliser battery

To charge the nebuliser battery, proceed as follows:

- Connect the micro USB cable [8] included in delivery to the micro USB port [7] on the bottom of the nebuliser and a USB mains adapter (not included in delivery; the mains adapter should comply with protection class 2 and have been tested in accordance with the European standard EN 60601-1). Insert the USB mains adapter into a suitable socket. **B**
Lie the nebuliser flat on its back during the charging process.
- While charging, the LED indicator [5] will light up continuously orange. As soon as the battery is fully charged, the LED indicator will stay on green.
- It usually takes approx. 4 hours to fully charge the battery.
- As soon as the battery is completely charged, remove the micro USB cable [8] with the USB mains adapter from the socket and the micro USB port [7] on the nebuliser.
- When the battery is fully charged, the battery life is approx. 4 hours. If the LED indicator [5]

lights up orange while you are using the device, this indicates that the battery level is low.

- The way in which a low battery level is indicated depends on how you are using the device. If you have opted for a high nebulisation rate, the LED indicator [5] will stay on orange; if you have chosen a low nebulisation rate, the LED indicator [5] will flash orange slowly.
- The nebuliser cannot be used while it is being charged.

7. USAGE

7.1 Preparing the nebuliser

- For hygiene reasons, it is essential to clean and regularly disinfect the medication container [2] and the atomiser set after each treatment. For this purpose, see section “Cleaning and maintenance”.
- If the therapy involves inhaling several different medications one after the other, please be aware that the medication container [2] must be rinsed under warm tap water after each use. See section “Cleaning and maintenance” on this process.

7.2 Filling the medication container

- Open the cover of the medication container [2] by lifting up the flap [1] on the rear [C] and fill the medication container with an isotonic saline solution or pour the medication straight in. Avoid overfilling.
- The maximum recommended filling quantity is 8 ml.
- Use medication only as instructed by your doctor and ask about the appropriate inhalation period and quantity for you.
- If the prescribed quantity of medication is less than 0.5 ml, it must be topped up with isotonic saline solution. Dilution is also necessary with viscous medications (viscosity < 3). Please observe the instructions of your doctor in this case, too.

7.3 Closing the medication container

- Close the cover of the medication container [2] by pressing the flap on the rear down and locking it. [D]

7.4 Connecting the atomiser set

- Securely connect the part of the atomiser set that you would like to use (mouthpiece [11], adult mask [9] or children’s mask [10]) to the

mounted medication container **2** **E**. If you are using the adult mask **9** or children's mask **10**, you may attach one of the straps included. Ensure that you securely fix the strap on the mask with a knot.

- Now move the device to your mouth and put your lips firmly around the mouthpiece. When using the mask(s), please place it over your nose and mouth.
- Ensure that you have added fluid before switching on the device. Start the device by pressing the On/Off switch **4**.
- If the spray mist is flowing out of the device and the LED indicator **5** is lit up continuously blue, the device is operating correctly.

NOTICE

The most effective way of nebulising is by using the mouthpiece. Nebulisation using a mask is only recommended if it is not possible to use a mouthpiece (e.g. for children who are not yet able to inhale using a mouthpiece). When using a mask to inhale, take care to ensure the mask fits well and the eyes are unobstructed.

The device medication flow rate is set to “high” in the factory settings. You can adjust the medication flow rate during use by pressing and holding the On/Off switch **4** for 3 seconds until the LED

indicator **5** flashes blue. You can switch from the high to the low nebulisation rate and back again. If the LED indicator **5** lights up continuously blue, this indicates a high nebulisation rate of ≥ 0.4 ml/min; if the LED indicator **5** flashes blue slowly, this indicates a low nebulisation rate of ≥ 0.2 ml/min.

Adjust the medication flow rate according to your requirements.

7.5 Inhaling correctly

Breathing technique

- It is important to use the correct breathing technique to ensure that the particles are distributed as widely as possible throughout the airways. To ensure that the particles reach your airways and lungs, you must breathe in slowly and deeply, then hold your breath briefly (5 to 10 seconds) and then breathe out quickly.
- Nebulisers should only be used for the treatment of respiratory diseases after consultation with your doctor. They will advise you about choosing the medications suitable for inhalation therapy, and at what dose and how to use them.
- Certain medications are only available with a prescription.

NOTICE

Certain medications (such as budesonide) might foam up during their application. If foam builds up within the medication container, or if the fill level is low and nebulisation can no longer occur, switch off the device. Automatic switch-off may no longer function reliably due to the foam; the mesh might continue to vibrate and eventually break. You should hold the device as vertically as possible. However, a slight angle will not affect usage, as the container is leak proof. For the device to work properly, ensure when using it that the medication is in contact with the mesh. If the device is tipped gently by 45 degrees with a maximum volume of 8 ml of liquid, the device will work as before.

⚠ CAUTION

For health reasons, essential oils, cough syrups, gargling solutions and drops to be used as a rub or in a steam bath are totally unsuitable for inhalation with a nebuliser. These additives are also often viscous and can impair the correct functioning of the device and therefore the effectiveness of usage in the long term.

For individuals with a hypersensitive bronchial system, medications containing essential oils may in some cases cause an acute bronchospasm (a sudden cramp-like restriction of the bronchi with

shortness of breath). Talk to your doctor or pharmacist about this.

7.6 Stopping inhalation

a) Manually ending nebulisation:

- When you have finished the treatment, switch off the device using the On/Off switch **4**.
- The LED indicator **5** will go off.

b) Automatically ending nebulisation:

- If the inhalant has been nebulised, the LED indicator **5** will flash blue quickly and the device will switch itself off. For technical reasons, a small amount of medication will remain in the medication container. Do not use this.

7.7 Cleaning


See section “Cleaning and maintenance”.


7.8 Automatic switch-off


The device has an automatic switch-off function. If there is only a very small amount of medication or fluid left or if the substance to be nebulised is no longer in contact with the mesh, the device will switch off automatically to avoid causing damage to the mesh. If the nebulisation stops automatically, the automatic switch-off has been activated. The LED indicator **5** will flash blue quickly and the device will then switch off. After 15 minutes

of nebulisation time, the device switches off automatically.

Do not operate the device when the medication container **2** is empty. The device detects when there is no medication in the medication container and will switch off automatically.

Colour of the LED indicator	Explanation
Green 	Fully charged: As soon as the battery is fully charged, the LED indicator will light up continuously green.

Colour of the LED indicator	Explanation
Orange 	Charging process: While the battery is being charged, the LED indicator will light up orange continuously. Low battery level: If the LED indicator lights up orange whilst you are using the device, this indicates that the battery level is low. <ul style="list-style-type: none">• At a high nebulisation rate: it will light up orange continuously• At a low nebulisation rate: it will flash orange slowly Self-cleaning mode: If you press and hold the On/Off switch for 3 seconds while the device is switched off, the LED indicator will flash orange quickly and the device will go into self-cleaning mode.

Colour of the LED indicator	Explanation
Blue 	<p>Operation and medication flow:</p> <p>The blue LED indicator indicates that the device is working correctly.</p> <ul style="list-style-type: none"> • High nebulisation rate of ≥ 0.4 ml/min: the LED indicator will light up blue continuously • Low nebulisation rate of ≥ 0.2 ml/min: the LED indicator will flash blue slowly <p>Too little medication and automatic switch-off:</p> <p>If the inhalant has been nebulised, the LED indicator will flash blue quickly and the device will switch itself off.</p>

8. CLEANING AND MAINTENANCE

⚠ WARNING

Follow these hygiene instructions to avoid health risks.

- The medication container **2** and atomiser set are intended for multiple use. Please note that different areas of use involve different cleaning and hygiene preparation requirements.

NOTICE

- Do not clean the mesh or the atomiser set mechanically using a brush or similar device, as this could cause irreparable damage and optimum treatment results may no longer be guaranteed.
- Please consult your doctor about the additional requirements for hygiene preparation (hand care, handling of medication/inhalation solutions) for high-risk groups (e.g. patients with cystic fibrosis).
- The cleaning instructions mention water later on; it should be as soft as possible. Distilled water is suitable for all cleaning steps except for self-cleaning.

Disassembling

- Hold the device with one hand. The button **6** on the rear of the device should be facing you. Press the button **6** using your thumb and horizontally pull off the medication container **2** forwards using your free hand. **F**
- Remove the mouthpiece **11** or the mask **9** or **10** from the medication container **2**.

- Open the cover of the medication container **2** by lifting up the flap **1** on the rear **G**.
- Reassembly is carried out in reverse order.

Cleaning

After each use, the medication container **2** and the parts of the atomiser set that were used, such as the mouthpiece **11**, mask (**9** or **10**), must be cleaned for approx. 5 minutes using (ideally distilled) warm water at a max. temperature of 40°C. To clean the medication container, please follow the self-cleaning instructions described below. Dry the parts carefully using a soft cloth. Put the parts together again once they are completely dry and place them in a dry, sealed container or disinfect them.

- Please perform the self-cleaning function afterwards (see “Self-cleaning function” section).

CAUTION

Never hold the whole device under running water to clean it.

When cleaning, ensure that any residue is removed and dry all parts carefully.

Never use any substances for cleaning that could potentially be toxic if they were to come into contact with the skin or mucous membranes, or if they were swallowed or inhaled.

If required, clean the housing of the device using a slightly damp cloth, which you can moisten with a mild soapy solution.

- If there is any residue from medical solutions or contamination on the silver/gold-coloured contacts of the device or medication container **2**, clean with a cotton bud moistened with ethyl alcohol.
- Also clean the outer sides of the mesh and medication container with a cotton bud moistened with ethyl alcohol.
- The device must be switched off and allowed to cool down each time before cleaning.
- Do not use any abrasive cleaning products.
- Make sure that no water gets inside the device!
- Do not clean the device or atomiser set in the dishwasher!
- Do not allow water to splash the device. Only operate the device if it is completely dry.

We recommend carrying out inhalation continuously since the inhalant may clog the medication container **2** if the treatment is paused for a significant time, and the effectiveness of the administration may be impaired as a result.

Self-cleaning function

To completely remove medication residue and any blockages, you can use the device's self-cleaning function.

To do this, proceed as follows:

1. Disassemble the device as described above.
2. Open the cover of the medication container **1** and pour out the remaining fluid.
3. Pour a sufficient amount (3-6 ml) of tap water (soft water) into the medication container **2**. Then close the cover. Shake the container slightly so that the remaining fluid mixes properly with the cleaning fluid.
4. Open the cover of the medication container, pour out all the liquid and then add enough tap water (soft water) again.
5. Place the medication container on the housing **3**.
6. Press and hold the On/Off switch **4** for 3 seconds when the device is switched off. The LED indicator **5** will then flash orange quickly and the device will go to self-cleaning mode. The device will automatically run for 3 minutes. Please make sure there is sufficient fluid in the medication container.

Disinfection

CAUTION

Ensure that you do not touch the mesh if possible, as this may destroy it. The masks must not be placed in hot water!

Please observe the following points when disinfecting your medication container **2** and the atomiser set.

Regular disinfection of the individual parts is recommended. First, clean the medication container **2** and atomiser set as described in the "Cleaning" section.

The details of the disinfection method are listed below.

Parts	Boiling	Ethyl alcohol
Medication container	✓	✓
Cover for medication container	✓	✓
Mouthpiece	✓	✓
Masks		✓

Use only the method specified for each component.

Disinfection method	Instructions
Boiling	Place the parts in question in boiling water for 15 minutes. Ideally, distilled water should be used to avoid contamination and limescale formation. The parts should not come into contact with the hot base of the pan.
Ethyl alcohol (70-75%)	Pour 8 ml of ethyl alcohol into the medication container. Close the container. Leave the alcohol in the container for at least 10 minutes. For better disinfection, shake the unit gently back and forth. Once this is complete, pour the ethyl alcohol out of the container. Repeat the process, but this time using water. Position the medication container so that you can drip a few drops of ethyl alcohol onto the mesh from the outside. Leave this to work for 10 minutes. Place the masks and mouthpiece in a container full of ethyl alcohol so that the parts are covered with the alcohol.

	Leave the parts in the container for 10 minutes. Then clean the masks and the mouthpiece with water.
--	--

Drying

- Dry the parts carefully using a soft cloth.
- Gently shake the medication container **2** a few times so that any water is removed from the tiny holes in the mesh.
- Place the individual parts on a clean, dry and absorbent surface and leave them to dry completely (for at least 4 hours).
- The medication container **2** must not be placed in the microwave.
- We guarantee 50 disinfection cycles with boiling water per medication container.

NOTICE

Please ensure that the parts are completely dried after cleaning, otherwise there is a high risk of bacterial growth.

When completely dry, reassemble the parts and place in a dry, sealed container.

Durability of materials

Like all plastic parts, the medication container **2** and the atomiser set are subject to a certain degree of wear and tear if used frequently and prepared hygienically. Over time, this may lead to

a change in the aerosol, which may reduce the effectiveness of the treatment. We therefore recommend that you replace the medication container **2** and atomiser set after one year at the most.

Storage

- Do not store in damp conditions (such as in a bathroom) and do not transport with any damp items.
- Protect from direct sunlight during storage and transport.

9. ACCESSORIES AND/OR REPLACEMENT PARTS

To purchase accessories and/or replacement parts, visit www.beurer.com or contact the corresponding service address (as per the service address list) for your country. Accessories and/or replacement parts are also available from retailers.

Name (Material)	REF
Mouthpiece (PP), adult mask (PVC), children's mask (PVC), mesh atomiser with medication container (PC)	601.43

10. WHAT IF THERE ARE PROBLEMS?

Problems/questions	Possible cause/remedy
The nebuliser produces no or too little aerosol.	1. Too little medication in the medication container 2 .
	2. Nebuliser not held vertically.
	3. Unsuitable medication fluid added for nebulisation (e.g. too viscous. Viscosity must not be greater than 3. Check with the pharmacy regarding viscosity.) The medication fluid should be prescribed by a doctor.
	4. The battery is flat. Charge it.

Problems/ questions	Possible cause/remedy
The output is too low.	1. There are air bubbles in the medication container, which are preventing continuous contact with the mesh. Please check this and remove any air bubbles.
	2. Particles on the mesh are hindering output. In this case, you should pause the inhalation process and remove the mask or mouthpiece. Then clean and disinfect the medication container (see "Cleaning and maintenance" chapter).
	3. The mesh is worn. Replace the medication container 2 .

Problems/ questions	Possible cause/remedy
Which medications are suitable for inhaling?	Naturally, only your doctor can decide which medication should be used for treating your condition. Please consult your doctor about this. With the IH 57, you can nebulise medications with a viscosity of less than 3. Medications containing oil (particularly essential oils) should not be used, as these attack the material of the medication container, which can lead to faults.
There is inhalation solution residue in the medication container 2 .	This is normal and is due to technical reasons. Stop the inhalation as soon as the atomiser starts to make a notably different sound or the device switches off automatically due to a lack of inhalant.

Problems/ questions	Possible cause/remedy
What should be taken into account when using the device with infants and children?	<ol style="list-style-type: none"> 1. For infants, the mask should cover the nose and mouth to ensure effective inhalation. 2. For children, the mask should also cover both the nose and mouth. It makes no sense to use the nebuliser on someone who is sleeping, since not enough of the medication will reach the lungs. <p>Notice: Inhalation should only be carried out under the supervision of an adult and with their assistance and the child should not be left alone.</p>
Should each person have their own atomiser set?	Yes, this is absolutely essential for hygiene reasons.

11. DISPOSAL

Battery disposal

- Empty, completely flat batteries must be disposed of in specially designated collection boxes, at recycling points or at electronics retailers. You are legally required to dispose of the batteries.
- The codes below are printed on batteries containing harmful substances:
Pb = battery contains lead
Cd = battery contains cadmium
Hg = battery contains mercury



General disposal

For environmental reasons, do not dispose of the device in household waste at the end of its service life. Dispose of the device at a suitable local collection or recycling point in your country. Dispose of the device in accordance with EC Directive – WEEE (Waste Electrical and Electronic Equipment). If you have any questions, please contact the local authorities responsible for waste disposal.



12. TECHNICAL SPECIFICATIONS

Type


IH 57

Dimensions (LxWxH)	46 x 78 x 108 mm
Weight	0,109 kg ± 0,005 kg
Input	5 V DC; 1 A
Energy consumption	< 1.5 W
Filling volume	min. 0.5 ml max. 8 ml
Medication flow rate	Low nebulisation rate: ≥ 0.2 ml/min High nebulisation rate: ≥ 0.4 ml/min
Vibration frequency	110 kHz ± 10 kHz
Operating conditions	Temperature: +5°C to +40°C Relative humidity: 15-90% non-condensing Ambient pressure: 700 to 1060 hPa
Storage and transportation conditions	Temperature: -25°C to +70°C Relative humidity: ≤ 90% non-condensing
Expected service life of the device	Information on the service life of the product can be found on the homepage

Battery:	
Capacity	1,800 mAh
Nominal voltage	3.7 V
Type designation	Lithium-ion
Aerosol values according to EN 27427:2019 based on adult breathing patterns with salbutamol:	<ul style="list-style-type: none"> • Aerosol delivery: 0.90 ± 0.03 ml • Aerosol delivery rate: 0.21 ± 0.01 ml/min • Fill volume dispensed in percent per min.: 28.71 ± 3.12% • Residual volume: 0.004 ± 0.001 ml • Particle size (MMAD): 4.519 ± 0.09 µm • GSD (geometrical standard deviation): 2.054 ± 0.04 • RF (respirable fraction < 5 µm): 55.04 ± 1.21% • Large particle range (> 5 µm): 44.96 ± 1.21% • Medium particle range (2 to 5 µm): 41.03 ± 0.46% • Small particle range (< 2 µm): 14.01 ± 1.03%
Sound pressure	≤ 45 dB(A)
Software version	V1.1.2.3

The serial number is located on the device.
Subject to technical changes.
(see “Cleaning and maintenance” chapter).

Particle size diagram



Measurements were performed using a salbutamol solution with a “Next Generation Impactor” (NGI).

The diagram may, therefore, not be applicable to suspensions or highly viscous medications. You can obtain more detailed information from the manufacturer of your medication. Aerosol values

are based on adult breathing patterns and are likely to vary for paediatric or infant populations.

NOTICE

If the device is not used according to the specifications, it may not work correctly! We reserve the right to make technical changes to improve and develop the product.

This device and its atomiser set comply with the European standard EN 60601-1-2 (Group 1, Class B, in accordance with CISPR 11, IEC 61000-3-2, IEC 61000-3-3, IEC 61000-4-2, IEC 61000-4-3, IEC 61000-4-4, IEC 61000-4-5, IEC 61000-4-6, IEC 61000-4-7, IEC 61000-4-8, IEC 61000-4-11) and is subject to particular precautions with regard to electromagnetic compatibility.

13. GUARANTEE / SERVICE

More information on the guarantee and guarantee conditions can be found in the guarantee leaflet supplied.

Notification of incidents

For users/patients in the European Union and identical regulation systems, the following applies: If a major incident occurs during or through use of the product, notify the manufacturer and/ or their representative of this as well as the re-

spective national authority of the member state in which the you are located.



UK Importer: Beurer UK Ltd., Suite 16, Stonecross Place, Stonecross Lane North,
WA3 2SH Lowton, United Kingdom



Beurer GmbH • Söflinger Str. 218 • 89077 Ulm, Germany
• www.beurer.com • www.beurer-healthguide.com
www.beurer-gesundheitsratgeber.com

