

	Title	EU Declaration of Conformity
	Date	2026-08-10

Manufacturer: Beurer GmbH (see address in footer)

SRN: DE-MF-000005422

Product category: Thermometers

Product type: FT 58 (Model: Medel Ear Temp)

Intended use: The thermometer can be used to measure body temperature in the ear canal in the home environment. The temperature is monitored in °C (or °F) to one decimal place.

The product specified above is in conformity with the following specifications.

(EU) 2017/745 Medical device regulation (MDR)

Basic-UDI-DI: 4211125FT58QB

Classification/applied rule(s): Class IIa/rule 10

Conformity assessment procedure: Annex IX, Chapter I

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700065, valid to 2030-11-15

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)


EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 2026-08-10

Name, function, signature, stamp: Max Harder, Team Lead Regulatory Affairs


Soeflinger Straße 218 • 89077 Ulm