

beurer	Title	EU Declaration of Conformity
	Date	2026-10-22

Manufacturer: Beurer GmbH (see address in footer)

SRN: DE-MF-000005422

Product category: Thermometers

The products specified below are in conformity with the following specifications.

Product name (Model/Type)	Basic-UDI-DI	intended use
FT 09-1	4211125FT09/16W	The thermometer is intended exclusively for the orally, rectally or axillary measurement of human body. The thermometer is only intended for private use in a domestic environment and not for commercial use.
FT 16	4211125FT16PT	
FT 17 (Type: FT 16)	4211125FT16PT	
FT 58	4211125FT58QB	The thermometer can be used to measure body temperature in the ear canal in the home environment. The temperature is monitored in °C (or °F) to one decimal place.
Medel Ear Temp (Type: FT 58)	4211125FT58QB	
FT 65	4211125FT65Q8	The thermometer is intended exclusively for measuring human body temperature in the ear canal (ear temperature mode) or on the forehead (forehead temperature mode). The thermometer is only intended for private use at home, not for commercial use.
FT 66	4211125FT66QA	
FT 85	4211125FT85QE	The thermometer is intended exclusively for measuring human body temperature on the forehead. The thermometer is only intended for private use at home, not for commercial use.
FT 93	4211125FT93QD	

(EU) 2017/745 Medical device regulation (MDR)

Classification/applied rule(s): Class IIa/rule 10

Conformity assessment procedure: Annex IX, Chapter I

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700065, valid to 2030-11-15

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)
EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 2026-10-22

Name, function, signature, stamp: Max Harder, Team Lead Regulatory Affairs

Beurer GmbH
Soeflinger Straße 218 • 89077 Ulm