

Fluid Fighter® | Suction Strip

Instructions for Use

Pre-Surgery

- 1. Remove the Fluid Fighter® Suction Strip from its packaging and place it on the operating room floor with the label sticker facing up.
- Position the suction strip with the suction tubing facing closest to the suction source.
- 3. Ensure the suction strip is placed flat on the floor.
- 4. Connect the suction strip's pre-attached tubing to the suction source.
- 5. Power on the vacuum source.

Post-Surgery

- Following the procedure, run the vacuum source for a few minutes to allow residual fluid in the suction strip to be vacuumed into the suction source.
- 2. Power off the suction source and disconnect the tubing from the suction source.
- 3. Dispose of the suction strip per your facility's protocol and place it in an appropriate waste receptacle.