



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance
for Weighing and Measuring Devices

For:

Load Cell
Compression, Beam
Model: PBW
 n_{max} : 4000, Single Cell, Class III
Capacity: See table below

Submitted By:

Flintec UK Ltd.
Caxton House, Caxton Place
Cardiff, South Wales CF23 8HG United Kingdom
Tel: +44 (0)2920 797959
Contact: Nick Jones
Email: nick.j@flintec.com
Website: www.flintec.com

Standard Features and Options

- Specific load cell capacities and v_{min} values covered by this certificate are listed in the table below.
- Nominal output: 0.9 mV/V
- Stainless Steel material
- 4 wire design
- Minimum Dead Load: 0 kg

Model	Capacity (lb)	v_{min} Class III (lb)
PBW	12.5	0.001
	18.75	0.002
	25	0.003
	37.5	0.004
	50	0.006
	100	0.008
	240	0.018

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of *Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices*. Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.

Ivan Hankins
Chair, NCWM, Inc.

Hal Prince
Chair, NTEP Committee
Issued: February 21, 2022

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend, or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Flintec UK Ltd.
Load Cell / PBW

Application: The load cells may be used in Class III scales for single cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the v_{\min} value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions (n_{\max}) and with greater v_{\min} values than those listed on the certificate. However, the load cells must be marked with the appropriate n_{\max} and v_{\min} for which the load cell may be used.

Identification: A pressure sensitive identification label located on the cell, states manufacturer name, model, serial number. Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

Test Conditions: A Model PBW, 12.5 lb and 100 lb capacity load cells were tested by the NMi Certin B.V. at the Netherlands facility. Testing was conducted in accordance with the OIML-CS for OIML R60 Certificate System, signed by the NCWM as a utilizing participant of load cell test data. Testing was conducted using deadweights as the reference standard. The load cells were tested over a temperature range of -10 °C to 40 °C with tests run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The data was analyzed for single load cell applications. OIML R60 selection criteria was used to determine which load cell capacities were tested.

Evaluated By: E. van der Grinten, M.M.J. Meijer (NMI)

Type Evaluation Criteria Used: *Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2020 Edition. *NCWM Publication 14: Weighing Devices*, 2021 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: D. Flocken (NCWM)

Example(s) of Device:

