

SINGLE USE EXAMINATION GLOVE, ACCELERATOR FREE, 100% NITRILE, POWDER FREE, 240MM LONG, 0.07MM THICK

High performance, powder-free, accelerator free nitrile single use examination glove. Alternative for people suffering allergies or reactions to accelerators contained in standard nitrile single use gloves

FEATURES

- › High quality nitrile, soft and durable
- › Palm thickness 0.07mm (2.7 mil), weight : 3.5g size M
- › Accelerator free to avoid skin irritation to people sensitive to accelerators contained in standard nitrile single use gloves
- › Textured finish on fingertips for improved grip and touch
- › Powder-free, latex free, non-sterile, AQL 1.5
- › Chlorinated inside to offer more comfort and reduce friction when donning
- › Antistatic properties due to inherent antistatic properties of nitrile
- › Approved for food handling
- › Dual labelling: PPE and medical
- › Medical device compliant (MDD 93/42/EEC), conforms to EN455 (all parts)
Tested against chemotherapy drugs according to ASTM D6978-05
Compliant with ASTM D6319-2019 and ASTM D5151-2019



SUITABLE FOR

Typical Industries

- Healthcare
- Electronics
- Food
- Pharmaceutical
- Manufacturing
- Oil and Gas
- Repair and Finishing

Suitable Applications

- Examination
- Food Handling
- Handling Oily Components
- Laboratory Analysis
- Light Assembly
- Paint Spraying

CERTIFICATION

CAT III

EN ISO 374-1:2016/ Type B

EN ISO 374-5:2016

CE 2777

KPT

Virus

MD MEDICAL DEVICE

See overleaf for explanation

**PRODUCTION INFORMATION**

MATERIALS	LINER:	Nitrile
	COATING:	–
COLOUR	Blue	
LENGTH (mm)	240	
CUFF STYLE	Beaded	

ORDERING DETAILS

SIZE	CODE	PACKAGING
7/S	SKG00310DD	100 singles per box 10 boxes per case
8/M	SKG00310DF	
9/L	SKG00310DH	
10/XL	SKG00310DJ	

RECOMMENDATIONS FOR USE

- **USE:** Single use glove. Not suitable for mechanical, thermal, heat, electrical, heavy chemical protection. Do not use near moving machines if there is a risk of entanglement
- **STORAGE:** Store in dry conditions in the original packaging and away from direct sunlight
- **CLEANING:** To clean, wipe with a damp cloth. Note: The performance characteristics of worn and laundered gloves may differ from the results shown. Inspect the gloves to ensure no damage is present
- **LIFETIME:** Service life depends on the glove application and therefore cannot be specified. It is the responsibility of user to ensure the glove is suitable for its intended use

CERTIFICATION LEGENDS**RESISTANCE TO CHEMICAL PERMEATION - EN ISO 374:2016**

CODE	CHEMICAL	CODE	CHEMICAL
A	Methanol	J	n-Heptane
B	Acetone	K	Sodium hydroxide 40%
C	Acetonitrile	L	Sulphuric acid 96%
D	Dichloromethane	M	65% Nitric acid
E	Carbon Disulfide	N	99% Acetic acid
F	Toluene	O	25% Ammonium hydroxide
G	Diethylamine	P	30% Hydrogen peroxide
H	Tetrahydrofuran	S	40% Hydrofluoric acid
I	Ethyl acetate	T	37% Formaldehyde

TYPE OF GLOVES	BREAKTHROUGH TIME
A	≥30 min for at least 6 chemicals
B	≥30 min for at least 3 chemicals
C	≥10 min for at least 1 chemical

**PROTECTION AGAINST MICRO-ORGANISMS EN 374-5**

VIRUS = Glove has passed ISO 16604:2004 (method B)

**FOOD CONTACT DIRECTIVE 2002 / 72 EC****GLOBUS****EUROPE**www.globusgroup.com**E:** sales@globus.co.uk**T:** +44 (0)161 877 4747**F:** +44 (0)161 877 4746**MIDDLE EAST & AFRICA**www.globusgroup.com/gcc**E:** gcc@globusgroup.com**T:** +971 4 882 9962**F:** +971 4 882 9963**GLOBUS AMERICAS**www.globusgroup.com**E:** americas@globusgroup.com

Globus Group, T2 Trafford Point, Twining Road, Trafford Park, Manchester, M17 1SH, UNITED KINGDOM
 Globus EMEA FZE, Jafza One, Tower A, Office 2201, Jebel Ali, PO Box 61195, Dubai, UNITED ARAB EMIRATES
 Globus Americas LLC, 45 Rockefeller Plaza, Suite 2000, New York City, New York, 10111, USA

EU Type-Examination Certificate issued by SATRA Technology Europe Limited, Bracetown Business Park, Clonee, D15YN2P, Republic of Ireland. (Notified Body No. 2777)