

NX16



16" Extended Cuff Nitrile Examination Gloves NON-LATEX, POWDER FREE, NON-STERILE

These long and strong powder-free nitrile disposable gloves are ideal where submersion is required above the wrist. They feature a textured surface to aid control and perform well in wet and dry situations.



Key Features and Benefits

Extended Cuff

Enhanced protection against fluids to the hand and forearm.

Secure Control

A textured surface enhances slip resistance and allows for a secure grip.

Beaded Cuff

Durable beaded cuff design enables the glove to be donned and removed efficiently and securely.

Powder-free and Latex-free

Eliminates powder-induced irritation and makes it ideal for people with a Type I latex sensitivity.

Chemical Resistant

Tested in accordance with EN374-1:2016 for resistance to chemical substances.

Premium Nitrile Butadiene Rubber

Intrinsic anti-static properties.


Recommended For

- ✓ Biotechnology
- ✓ Chemical Handling
- ✓ Electronics
- ✓ Laboratory Analysis
- ✓ Medical Device Manufacturing
- ✓ Pharmaceutical

Material Information

- Premium performance Nitrile Butadiene Rubber (NBR)
- Powder-free
- Beaded cuff
- Light blue

Certification



Packaging



Technical and Ordering Data

TYPE	Non-sterile, powder-free single use examination gloves
MATERIAL	Nitrile Butadiene Rubber (NBR)
COLOUR	Light Blue
GLOVE SHAPE	Flat with a beaded cuff
OVERALL LENGTH / SIZE (AS PER EN455-2)	16 inches / 400mm XS / S / M / L / XL
WALL THICKNESS (MEASURED IN SINGLE LAYER OF THE PALM AREA)	Min. 0.12 mm
IMPERMEABILITY (AS PER EN 455-1)	AQL 1.5
FORCE AT BREAK (N) (AS PER EN 455-2)	7.0 (before ageing) 6.0 (after ageing)
SHELF LIFE (IN ORIGINAL PACKAGING WHEN STORED AS PER DIN 7716, ISO 2230)	5 years
PACKAGING	100 single gloves per dispenser 10 dispensers per case
COUNTRY OF ORIGIN	Malaysia

ORDER NUMBERS	HAK0160 HAK0161 HAK0162 HAK0163 HAK0164	XS S M L XL
----------------------	---	-------------------------

General Information

Conformance:

HAIKA™ NX16 medical examination gloves have been tested and found to comply with all parts of EN455.

Registering Authority:

Products conforms to the requirements of: UK Regulation 2016/425 on PPE, brought into UK law and amended & EU regulation 2016/425 of the European Parliament of the Council as Personal Protective Equipment (PPE).

Type examination certificates (Module B) issued by:

UK: ITS Testing Services UK Ltd, Centre Court, Meridian Business Park, Leicester, LE19 1WD, United Kingdom (Approved Body No. 0362).

EU: INTERTEK ITALIA SPA, Via Guido Miglioli, 2/A - 20063 Cernusco sul Naviglio (MI) - Italia (Notified Body No. 2575).

They are also in compliance with Medical Device Regulation 2017/745 Medical Device Class 1 (non-sterile).

Note: Globus assumes no responsibility for the suitability of a user's glove selection for a specific application.

More information:

www.globusgroup.com

E: sales@globusgroup.com

T: +44 (0)161 877 4747

F: +44 (0)161 877 4746

