



# Certificate of Compliance

**Certificate:** 1828961 (LR80259)

**Master Contract:** 157918

**Project:** 2369435

**Date Issued:** January 14, 2011

**Issued to:** BW Technologies by Honeywell

2840 2nd Ave SE  
Calgary, AB T2A 7X9  
Canada  
Attention: Thomas A. Crawford

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Alan McLaughlin*

**Issued by:** Alan McLaughlin, P.Eng.

## **PRODUCTS**

**CLASS 2252 05** - PROCESS CONTROL EQUIPMENT

**CLASS 2252 85** - PROCESS CONTROL EQUIPMENT - Certified to US Standards

Gas detector charging station, Model 'X' MultiCharger, rated 6.0 Vdc, 2.5 A; powered from external AC power supply adaptor, input rated 100-240 Vac, 0.6 A, 50/60 Hz, output rated 6.0 Vdc, 2.5 A.

Note:

1. Where 'X' is 'GasAlert MicroClip' or 'GasAlert MicroClip XT' or 'GasAlertMax XT' or 'GasAlertQuattro'
2. The above model is Equipment Class III, Pollution Degree 2, Installation Category I.



**Certificate:** 1828961 (LR80259)

**Master Contract:** 157918

**Project:** 2369435

**Date Issued:** January 14, 2011

---

**APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 0-10 - General Requirements- Canadian Electrical Code, Part II

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements