

Benchmark Scientific Inc. 116 Corporate Blvd. South Plainfield, NJ 07080 USA

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Benchmark BioClaveTM 16 and BioclaveTM Mini Research Autoclaves

Device Model Number(s): B4000-16-E and B4000-M-E Private label versions of these devices.

Relevant EC Directives: Council Directive 93/42/EEC

Harmonized Standards: EN 285:2006+A1:2008

EN 13060:2004+A1:2009

Manufacturer's Name: Benchmark Scientific, Inc.

PO Box 709 Edison, NJ 08818

Date: February 20, 2017

Marketing Manager

Phone: (908) 769-5555 Fax: (908) 222-1864 Email: info@BenchmarkScientific.com