

# Declaration of Conformity

## Techne Biological Stirrers, Models MCS-101, MCS-102, MCS-104S, MCS-104L, MCS-104XL

These products comply with the requirements of the EU Directives listed below:

**2004/108/EC**      **EMC Directive.**  
**2006/95/EC**      **Low voltage Directive (LVD)**

Compliance with the requirements of these Directives is claimed by meeting the following standards:

**EN 61326-1:2006 (Electrical Equipment for Measurement, Control and Laboratory use).**  
**EN 61010-1: 2001**  
**(Safety Requirements Electrical Equipment for Measurement, Control and Laboratory use)**  
**EN61010-2-051: 2003 (Particular Requirements for Laboratory Equipment for Mixing and Stirring)**

Compliance Certificates and Equipment Specification.


The full product specifications, listed in report numbers:

RETS(E)1496/A/1

RETS(L)1496/A/2

The above certificates and reports, from an independent test house, are available upon request.

**CE mark affixed '96.**

Signed:  (Mr C. Warren)

Date: 027 09

Authority: Technical Manager  
Bibby Scientific Ltd

 Bibby Scientific

Bibby Scientific Ltd - Stone - Staffs - ST15 0SA - UK  
Tel: +44 (0) 1785 812121 - Fax +44 (0) 1785 813748  
DofC\_88\_Rev01