THIS IS ISSUED IN STRICT CONFIDENCE THAT IT IS NOT USED AS A BASIS FOR MANUFACTURE OR SALE AND THAT IS NOT COPIED, REPRINTED OR DISCLOSED TO A THIRD PARTY EITHER WHOLLY OR IN PART WITHOUT THE PRIOR WRITTEN CONSENT OF CAL TEST ELECTRONICS. INC.

1. 4mm SAFETY JACK w/M4 STUD & (2) NUTS.

2. SPECIFICATIONS:

INSULATOR: POLYAMIDE (NYLON). CONTACT: BRASS, NICKEL PLATED LOCKING NUT: BRASS, NICKEL PLATED

3. RATING:

VOLTAGE: 1,000 V CAT III/600 V CAT IV P2 (IEC61010-031)

CURRENT: 36 A MAX. RESISTANCE: <5 m $\Omega$ .


TEMPERATURE:  $-20^{\circ}$  C TO  $+80^{\circ}$  C.

4. MOUNTING HOLE: REF. DIAGRAM.

5. INSULATOR THREAD, M12 X 0.75. MAX. TORQUE 1.7 ft-lb (2.3 Nm) AT 20° C.

6. MAXIMUM PANEL THICKNESS: .40 (10).

🕇 7. RECOMMENDED ASSEMBLY TOOL: CT2246A.



-COLOR: 0 - BLACK

1 - BROWN

2 - RED

3 - ORANGE (SPECIAL)

4 - YELLOW 5 - GREEN

6 - BLUE

7 - PURPLE

8 - GRAY 9 - WHITE

CONFORMITY TO RoHS 2002/95/EC

REVISIONS


REV. DESCRIPTIONS DATE APPROVED

- CLASS A RELEASED 7/12/99

A ADDED CAGE CODE 6/10/05

B NOTE 2, MAT'L WAS POLYACETAL 5/19/06

C COLOR ORANGE ADDED AS SPECIAL; REVISED NOTE 5 WITH NUT TORQUE SPECIFICATION; NOTED RoHS COMPLIANCE.



THIRD ANGLE PROJECTION		UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE:	CONTRACT NO.		Cal Test Electronics 22820 Savi Ranch Parkway Yorba Linda, CA 92887-4610					
		FRACTIONS DECIMALS ± 1/32 .XX ± .02 .XXX ±.005 DO NOT SCALE DRAWING	APPROVALS DATE		TITLE					-4610
			DRAWN WDH	7/12/99	1	4mm			CK, PANEL	
		w/M4 STUD						D		
		FINISH	ENGR.		SIZE CAGE CODE		DWG. NO.	P223600 REV		REV.
NEXT ASSY	USED ON		DESIGN ACTIVITY		A	43F45		1 223000		
APPLICATION					SCALE	1:1			SHEET 1 OF 1	