



Werkskalibrier-Zertifikat Company-calibration certificate

Gegenstand object Referenzöl für Frittieröltester
Reference oil for frying oil tester

Hersteller manufacturer Testo AG

Typ type 0554.2650

Chargen Nr. serial no. 1343129022

Datum der Kalibrierung 21.10.2013
date of calibration

Hiermit bestätigen wir, dass neben genanntes Messsystem unter Beachtung eines nach **DIN EN ISO 9001:2008** zertifizierten Qualitätssicherungssystems kalibriert wurde. Bei den für die Kalibrierung verwendeten Messeinrichtungen handelt es sich um werkseigene Normale.

Hereby we confirm that the beside mentioned measuring system was calibrated in compliance with an accredited quality assurance system, which has been certified to **DIN EN ISO 9001:2008**. The measuring installations used for calibration are company-internal measurements standards.

Messergebnisse measuring results

Referenzwert reference	Temperatur temperature	Messunsicherheit uncertainty of measuring
5,4 %TPM	51,2 °C	±0,5 %TPM

Dieser Kalibrierschein darf nur vollständig weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory.

Calibration certificates without signature and seal are not valid.

Stempel seal



Laboraufsicht supervisor

Helmut Isele / P-F

Bearbeiter person responsible

Georg Winterhalter

testo AG

Testo-Strasse 1
79853 Lenzkirch

Telefon 07653 681 0
Telefax 07653 681 100

www.testo.de
info@testo.de