

Manual Supplement

Manual Title: 1587/1577 Users
Part Number: 2401027
Print Date: April 2005
Revision/Date: 2, 6/09

Supplement Issue: 1
Issue Date: 3/10
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 54524

On page 34, replace the **Ohms Measurement** table with the following:

Ohms Measurement

Range	Resolution	Accuracy 1587 and 1587T ^[1] +(% of Rdg+Digits)	Accuracy 1577 ^[1] +(% of Rdg+Digits)
600.0 Ω	0.1 Ω	±(0.9 % + 2)	±(1.2 % + 2)
6.000 kΩ	0.001 kΩ		
60.00 kΩ	0.01 kΩ		
600.0 kΩ	0.1 KΩ		
6.000 MΩ	0.001 MΩ		
50.0 MΩ ^[2]	0.01 MΩ	±(1.5 % + 3)	±(2.0 % + 3)
<p>[1] Accuracies apply from 0 to 100% of range.</p> <p>[2] Up to 80 % relative humidity.</p>			

