

# Manual Supplement

Manual Title: 1760 Users  
Print Date: June 2006  
Revision/Date: 2, 5/10

Supplement Issue: 1  
Issue Date: 2/11  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 1760  
CD Rev. & Date: 5, 7/2010  
CD PN: 2583518


**T**equipment  
\_\_\_\_\_.NET



205 Westwood Ave  
Long Branch, NJ 07740  
1-877-742-TEST (8378)  
Fax: (732) 222-7088  
salesteam@Tequipment.NET

## Change #1, 57261, 57516

On page 1-7, replace the first Figure with the following:



On page 5-3, under **General Specifications**, replace the Power Supply with the following:

Power Supply Nominal	100 to 240 V ac, 50/60Hz, 100 to 350Vdc
Power Supply Operating	83 to 264 V ac, 45 to 65 Hz, 100 to 375Vdc, 35W/70 VA

On page 6-5, under **Flexible Current Probes for AC**, replace the Model No. Product No. column with:

Model No. Product No.
TPS FLEX 18-TF 3889490
TPS FLEX 24-TF 3889503
TPS FLEX 36-TF 3889515

On page 6-18, under **General Specification**, add the Measurement head cross section specifications and replace Weight, and Order-number with:

Measurement head cross section: 8 mm (0.3 in)  
 Weight: Approx. 0.3 kg (0.66 lb)  
 Order-number: 3889490

On page 6-21, under **General Specification**, add the Measurement head cross section specifications and replace Weight, and Order-number with:

Measurement head cross section: 8 mm (0.3 in)  
 Weight: Approx. 0.35 kg (0.66 lb)  
 Order-number: 3889503

On page 6-25, under **General Characteristics**, add the Measurement head cross section specifications and replace Weight, and Order-number with:

Measurement head cross section: 8 mm (0.3 in)  
 Weight: Approx. 0.4 kg (0.88 lb)  
 Order-number: 3889515