



Manual Supplement

Manual Title:	287/289 Getting Started	Supplement Issue:	3
Part Number:	2748860	Issue Date:	6/10
Print Date:	June 2007	Page Count:	2
Revision/Date:	1, 3/09		

This supplement contains information necessary to ensure the accuracy of the above manual.


Change #1, 54932

In the Specification section on page 6, under **Electromagnetic Compatibility Standards (EMC)**, replace the **Certifications** with following:

Certifications.....UL, CE, CSA,  (N10140), 

Change #2, 55189, 55213, 55269

On page 3, Table 1, replace the TUV symbol with the following:

From: 

To: 

In the Specification section on page 11, under **Resistance Specifications**, change the **Accuracy** for 500 k Ω :

From: 0.05 % + 2

To: 0.05 % + 15

In the Specification section on page 13, under **Frequency Counter Specifications**, add a footnote 3 to the Duty Cycle row.

[3] For 10 μ s < pulse width <25 μ s add 1%. For 2 μ s < pulse width \leq 10 μ s add 3.5 %.

Under Pulse Width, change the **Accuracy** for 0.1000 ms:

From: 0.002 ms + 3 counts

To: 0.002 ms + 30 counts

In the Specification section on page 15, under **Typical Short Circuit Current**, change **5 MΩ**:

From: 0.3 μA
To: 1 μA

Change #3, 55354

In the Specification section on page 5, under the **General Specifications**, replace the **Safety Standards**, with the following:

Safety Standards

- US ANSI ANSI/ISA 61010-1 2004 (82.02.01):2004
- CSA CAN/CSA-C22.2 No 61010-1-04
- UL UL 61010-1:2004
- CE European IEC/EN 61010-1 2nd Edition
- Pollution Degree 2



Equipment.NET

**205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Equipment.NET**