205 Westwood Ave Long Branch, NJ 07740 1-877-742-TEST (8378) Fax: (732) 222-7088 salesteam@Tequipment.NET

## **Manual Supplement**

Manual Title: 323/324/325 Users

Part Number: 4045153 Print Date: May 2012

Revision/Date:

Supplement Issue: 1

Issue Date: 6/12

Page Count:

This supplement contains information necessary to ensure the accuracy of the above manual.



323/324/325 Users Manual Supplement

## Change #1, 62269

On page 3, add the following Warning:

 Do not use a current measurement as an indication that a circuit is safe to touch. A voltage measurement is necessary to know if a circuit is hazardous.

On page 10, under **Specifications**, AC Current (Jaw), change Accuracy for the 324:

Add the following Note:

Note

Add 2 % for position sensitivity.



6/12