

# Manual Supplement

Manual Title:	33X Calibration	Supplement Issue:	1
Part Number:	1618765	Issue Date:	5/07
Print Date:	June 2001	Page Count:	1
Revision/Date:	3, 4/06		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1

On page 8, under **5520A Output**,  $\tilde{\text{A}}$  Amps AC,

Change:

10 A 50 Hz (333)
------------------

To:

8.0 A 50 Hz (333)
-------------------

Under **5520A Output**, \*INRUSH,

Change:

*INRUSH	**5 A 60 Hz
---------	-------------

To:

*INRUSH	**0.1 A 60 Hz
---------	---------------

On Page 9, under **5520A Output**, **Lower Limit** and **Upper Limit**,  $\tilde{\text{A}}$  Amps AC, delete the entire row:

10 A 50 Hz	187.5 A AC	212.5 A AC
------------	------------	------------

Under **5520A Output** \*INRUSH and Frequency (337),

Change:

*INRUSH	**5 A 60 Hz
---------	-------------

To:

*INRUSH	**0.1 A 60 Hz
---------	---------------

Change:

Frequency (337)	5 A 10 Hz
	10 A 300 Hz

To:

Frequency (337)	0.1 A 10 Hz
	0.2 A 300 Hz