

Supplement

Title: 373 Users Supplement Issue: **2**
Part Number: 3612805 Issue Date: 10/10
Print Date: July 2010 Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.



Tequipment
.NET

USA

205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET

Change #1, 55883

On page 15, under *DC Voltage*, add flag note [1] to Accuracy and add:

- [1] In an RF field greater than 1 V/m between 130 MHz and 350 MHz, add 65 counts to specified accuracy.

Change #2, 56248

On page 3, under **Warning** add the following:

Measure a known voltage first to make sure that the Meter operates correctly.

On page 17, under Safety Compliance change:

From: EN/IEC 61010-031:2002/A1:2008

To: EN/IEC 61010-031:2002

Add: CAN/CSA-C22.2 No. 61010-2-032:04