

Supplement

Title:	374, 375, 376 Users	Supplement Issue:	2
Part Number:	3608883	Issue Date:	11/10
Print Date:	July 2010	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above Users Manual.

Tequipment
equipment
.NET



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET

Change #1

On page 3, under **Warning** add the following:

- **When batteries are changed, ensure that the calibration seal in the battery compartment is not damaged. If damaged, the Meter may not be safe to use. Return the Meter to Fluke for replacement of the seal.**

Change #2, 56750

On page 3, under **Warning** add the following:

- **Reinstall the clear battery cover after the batteries are replaced.**

On page 27, under Operating Altitude:

Change: Operating Altitude 3000 meters

To: Operating Altitude 2000 meters