

Manual Supplement

Manual Title: 51,52,53,54 Series II Service Supplement Issue: 2
Part Number: 1276123 Issue Date: 6/05
Print Date: October 1999 Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:


1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET

Change #1

On page 8, under **Verification of IR Communications Port (Fluke 53 and 54)**, add step 3A as follows:

- 3A. On the Fluke 53 or 54, press  then **RECALL**.

Change #2

On page 8, Table 3, replace steps 4 and 5 with the following:

Step	Input Voltage (MV DC)	Equivalent Temperature		Display Reading Limits	
4	40.8853	990.0 °C		989.2 to 990.8	
5	48.8382		2192 °F		2190 to 2194