

Manual Supplement

Manual Title:	700G Series Calibration	Supplement Issue:	1
Print Date:	April 2012	Issue Date:	2/13
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	700G Series
CD Rev. & Date:	12/2011
CD PN:	4098593

Change #1, 61061, 64658, 64721

On page 3, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 4, replace the **Environmental**, section with:

Environmental

Operating Temperature	-10 °C to +55 °C (14 °F to 131 °F)
Storage	-20 °C to +70 °C (-4 °F to +158 °F)
Storage (without batteries)	-40 °C to +70 °C (-40 °F to +158 °F)
Humidity	10 % to 95 % RH Non-condensing
Pollution Degree	2
IP Rating.....	64 (with boot and serial-port plug installed)
Electromagnetic Environment.....	IEC 61326-1, Portable
Electromagnetic Compatibility	Applies to use in Korea only. Class A equipment (Industrial Broadcasting & Communication Equipment) ^[1] [1] This Product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.