



205 Westwood Ave  
Long Branch, NJ 07740  
1-877-742-TEST (8378)  
Fax: (732) 222-7088  
salesteam@Tequipment.NET

# Supplement

Title: 700PEX I.S.	Supplement Issue: 4
Part Number: 2106534	Issue Date: 8/10
Print Date: May 2004	Page Count: 2
Revision/Date:	

---

---

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

## Change #1

Under ***Product Compliance Markings***, change the following:

Replace:

C E  II 1 G Eex ia IIC T4  
**0344**

With:

C E  II 1 G EEx ia IIC T4  
**0344 KEMA 04ATEX1102 X**

Replace:

Ex Certification by Mensor Corporation,  
San Marcos, TX USA

With:

Manufactured by Mensor Corporation, 201  
Barnes Drive, San Marcos, TX USA 78666

## Change #2

Under ***Product Compliance Markings***, add the following:



Conforms to Australian directives.

## Change #3

Under **Safety Information**, add the following to the **Warning** section:

- Entity parameters for intrinsic safety can be found in Table 3.

Under **Faults and Damage**, add the following to the **Warning** section:

- Substitution of components may impair intrinsic safety.

## Change #4

Replace the **Box Contents** with:

### **Box Contents**

Each pressure module comes with an installed strap and instruction sheet. All pressure modules, with the exception of the 700P29Ex, come, with 1/4 NPT F to ISO7-R 1/4 (Metric) M adapters.



205 Westwood Ave  
Long Branch, NJ 07740  
1-877-742-TEST (8378)  
Fax: (732) 222-7088  
salesteam@Tequipment.NET