



Supplement

Title: 705 Inst.Sht. Supplement Issue: **2**
Print Date: August 1998 Issue Date: 3/08
Revision/Date: 2, 6/03 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 705 & 707
CD Rev. & Date: 1, 2/2008
CD PN: 2088974

Change #1

Under **Specifications**, change the **Input/Output Protection**, to:

Input/output protection on mA range: self resetting, nonreplaceable 0.1 A fuse.


Change #2, 41512

Under **General Specifications**, following the **Weight** specification, add:

Protection Class: Pollution Degree II

Change #3

On the second panel, under **International Symbols**, add the following:

 N10140	Conforms to Australian C-Tick directives
--	--

Tequipment
_____.NET



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET