

Manual Supplement

Manual Title: 718Ex 30G/100G/300G Users Supplement Issue: **3**
Print Date: May 2004 Issue Date: 3/10
Revision/Date: 2, 5/09 Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 718Ex 30G/100G/300G
CD Rev. & Date: 2, 5/2009
CD PN: 2097427

Tequipment
NET



**205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET**

Change #1, 51120

On page 27, add the following section under the **Entity Parameters**:

Pressure Module Output Circuit:

Vo, Uo	Io	Po	Co			Lo		
			IIC	IIB	IIA	IIC	IIB	IIA
7.14 V	102 mA	182 mW	13.5 μ F	240 μ F	1000 μ F	3.4 mH	10.2 mH	27.2 mH

Change #2, 52975

On page 9, under **Certification Information**, change the following:

From: CE  II 1 G EEx ia IIC T4
0344

To: CE  II 1 G Ex ia IIC T4
0344

On page 27, under **Product Compliance Markings**, change the following:

From: CE  II 1 G EEx ia IIC T4
0344 **KEMA 04ATEX1601 X**

To: CE  II 1 G Ex ia IIC T4
0344 **KEMA 04ATEX1601 X**

Change #3, 54529

On page 26, under **General Specifications** add the following:

Operating altitude: 3000 meters maximum

On page 27, under **Certification Information**, replace the fourth bullet with the following:

- Manufactured by Martel Electronic, Inc.
3 Corporate Park Drive
Derry, NH USA



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET