

Manual Supplement

Manual Title:	77 Series IV Calibration	Supplement Issue:	1
Print Date:	September 2006	Issue Date:	3/07
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.



Change #1

On page 12, Table 2, under **Meter Response** Beeper Off, change step 8,

From:

0.47 V ac	0.53 V ac
-----------	-----------

To:

0.047 V ac	0.053 V ac
------------	------------

Change #2

On page 13, Table 2, make the following changes:

Under **5500 Output**, step 25 change:

From:

Apply 900 nF

To:

900 nF

Add a new step 26 and renumber the remaining steps accordingly:

26	➔ Diode Test ^[1]	2.0 V	1.978 V dc	2.022 V dc
----	-----------------------------	-------	------------	------------