

Instrument Security Procedures

Model:

Fluke 77-IV, Fluke 77-BN

Product Name:

Digital Multimeter

Instrument Description:

Battery-powered, average responding-rms indicating multimeter with a 6000-count, 3 3/4-digit display, and a bar graph.

Memory Description:

The Fluke 77IV has one memory device: U4, micro controller, TI MSP430F437, 32kB Flash, 1024B RAM, contains operating code for the product and calibration constants.

Memory Cleaning Instructions:

The user controlled storage operation, other than calibration constants, is the storage of Min, Max and Average. Min, Max and Average is cleared when the function switch is changed.