

# Manual Supplement

Manual Title: 77 Series IV Users  
Part Number: 2695884  
Print Date: September 2006  
Revision/Date:

Supplement Issue: 1  
Issue Date: 10/06  
Page Count: 1

---

This supplement contains information necessary to ensure the accuracy of the above manual.



**Tequipment**  
**.NET**

**USA**

205 Westwood Ave  
Long Branch, NJ 07740  
1-877-742-TEST (8378)  
Fax: (732) 222-7088  
salesteam@Tequipment.NET

## Change #1 - 38525

On page 4, under **MIN MAX AVG Recording Mode**, under the note replace the last part of both sentences with the following:

*.....changes longer than 350 ms in duration.*

On page 13, under **Function**, replace the MIN MAX AVG, row with the following:

MIN MAX AVG	For dc functions, accuracy is the specified accuracy of the measurement function $\pm 12$ counts for changes longer than 350 ms in duration.  For ac functions, accuracy is the specified accuracy of the measurement function $\pm 40$ counts for changes longer than 350 ms in duration.
-------------	--

## Change #2

On page 5, under **Manual Ranging and Autoranging**, replace the second paragraph under the note with the following:

*If you press **RANGE** while in MIN MAX\_AVG, the meter beeps twice, indicating an invalid operation, and the range does not change.*