



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET

Supplement

Manual Title:	80PK-3A I. S.	Supplement Issue:	1
Print Date:	Sept. 1989	Issue Date:	3/10
Revision/Date:	4, 4/02	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1, 54379

Under **Specifications**, change the following:

From: **Junction Accuracy:** (With respect to ANSI MC96.1):
 ± 2.2 °C (3.96 °F) over the range of 0 °C to 260 °C
(32 °F to 500 °F)

To: **Accuracy:** (With respect to ANSI MC96.1): ± 4.0 °C
(7.2 °F) over the range of 0 °C to 260 °C (32 °F to 500 °F)



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET