

# Manual Supplement

Manual Title: 8808A Calibration  
Print Date: October 2008  
Revision/Date:

Supplement Issue: 1  
Issue Date: 6/10  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 8808A  
CD Rev. & Date: 3, 12/2009  
CD PN: 2713002



205 Westwood Ave  
Long Branch, NJ 07740  
1-877-742-TEST (8378)  
Fax: (732) 222-7088  
salesteam@Tequipment.NET

**Change #1**

On page 3-11, replace Table 3-8 with the following:

**Table 3-8. Frequency Verification Steps**

Nominal Input		Range	90-Day Test Limit		1-Year Test Limit	
Ampl.	Freq.		High	Low	High	Low
0.1 V	20 Hz	2000 Hz	20.04 Hz	19.96 Hz	20.06 Hz	19.94 Hz
0.1 V	50 Hz	2000 Hz	50.02 Hz	49.98 Hz	50.03 Hz	49.97 Hz
0.1 V	1900 Hz	2000 Hz	1900.24 Hz	1899.76 Hz	1900.26 Hz	1899.74 Hz
0.1 V	19 kHz	20 kHz	19.0023 Hz	18.9977 Hz	19.0025 Hz	18.9975 Hz
0.1 V	195 kHz	200 kHz	195.024 Hz	194.977 Hz	195.026 Hz	194.975 Hz
0.1 V	1000 kHz	1000 kHz	1000.14 Hz	999.86 Hz	1000.16 Hz	999.84 Hz