

Manual Supplement

Manual Title: 8845A/8846A Programmers Supplement Issue: 2
Print Date: September 2006 Issue Date: 9/10
Revision/Date: 2, 6/08 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 8845A/8846A
CD Rev. & Date: 5, 10/2009
CD PN: 2453193



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET

Change #1

Under the **LIMITED WARRANTY AND LIMITATION OF LIABILITY** replace the second sentence with the following:

The warranty period is three years limited and begins on the date of shipment.

Change #2

On page 12, under **Configuring the Ethernet (LAN) Port**, in the second paragraph,

Change: Net Mask
To: netmask

On page 13, under **Setting a Static Internet Address**, replace the first sentence with the following:

The Meter comes from the factory with 169.254.1.1 in the static IP address register.

On page 15, under **Configuring the General Network Socket Port**, replace the first sentence with the following:

In order to communicate with each other, the host computer must know the socket port number of the Meter.

On page 16, under **Establishing an Ethernet Connection**, replace the last sentence in the last paragraph with the following:

This prevents multiple computers from trying to control the Meter at the same time.