

Manual Supplement

Manual Title: 88 V Users
Part Number: 2166623
Print Date: August 2004
Revision/Date: 1, 1/06

Supplement Issue: **2**
Issue Date: 8/10
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.



**205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET**

Change #1, 39785

On page 73, Table 19, replace the Diode row with the following:

→	1000 V rms	< 7.9 V dc	3.000 V dc	1.0 mA typical
---	------------	------------	------------	----------------

Change #2, 55742


On page 1, replace the **Safety Information** with the following:

The Meter complies with:

- Measurement Category III, 1000 V, Pollution Degree 2
- Measurement Category IV, 600 V, Pollution Degree 2

See Specifications for a list of compliant standards.

On page 4, remove the following from the Table 1:

	Inspected and licensed by TÜV Product Services.
---	---

On page 67 replace the **Safety** section and add the following:

Safety:.....Complies with ANSI/ISA 61010-1 (82.02.11):2004, CAN/CSA C22-2 No.61010-1-04, and UL 61010-1:2004 to 1000 V CAT III/600 V CAT IV. Pollution Degree 2

EMC:.....Complies with EN/IEC 61326-1:2006