

Instrument Security Procedures

Model:

Fluke 902

Product Name:

HVAC Clamp Meter

Instrument Description:

Battery-powered, true-rms indicating multimeter with a 6000-count 3 3/4-digit display.

Memory Description:

The Fluke 902 has one memory device: U1, micro controller, TI MSP430F427, 32kB Flash, 1024B RAM, contains operating code for the product and calibration constants.

Memory Cleaning Instructions:

The user controlled storage operation, other than calibration constants, is the storage of Min, Max and Hold. Min, Max and Hold is cleared when the function switch is changed.



205 Westwood Ave
Long Branch, NJ 07740
1-877-742-TEST (8378)
Fax: (732) 222-7088
salesteam@Tequipment.NET