

Instrument Security Procedures

Model:

Fluke-985

Product Name:

Fluke 985 Airborne Particle Counter

Instrument Description:

Airborne Particle Counter that counts particles that range in size from 0.3 µm to 10.0 µm, can store 10k samples and has a rechargeable Lithium Ion battery. Supplied with charging base that is used for connectivity to PC.

Memory Description:

Microprocessor flash: 512 kB for bootloader and executable code

Microprocessor SRAM: 98 kB for system use

External NOR flash: 8 mB ROM program data

External SDRAM: system use

External EEPROM: factory and calibration settings

Memory Cleaning Instructions:

From the Buffered data screen, use arrow keys to select and erase stored data, then confirm.

