

Manual Supplement

Manual Title:	P3100/P3200 Users	Supplement Issue:	1
Part Number:	3952224	Issue Date:	10/12
Print Date:	November 2010	Page Count:	2
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.



FLUKE®

Calibration

Change #1

On page 2-2, replace the first sentence in the **Warning** with:

DO NOT use PTFE tape on these connections, as this will prevent correct sealing.

On page 2-5, Table 2-1, replace Items 1 and 3 with:

Item	Description	Part
1	O Ring	3865163
3	O Ring	3865195

On page 6-1, replace the first sentence in the **Caution** with:

O-ring seals (where fitted) are fluoroelastomer rubber, and should not be immersed in solvents, as they will become damaged.

On page 6-4, Table 6-1, replace Items 3 and 6 with

Item	Description	Part
3	O Ring	3865277
6	O Ring	3864802

On page 6-4, replace the first sentence in the first **Caution** with:

O-ring seals (where fitted) are fluoroelastomer rubber, and should not be immersed in solvents, as they will become damaged.

On page 6-6, Table 6-2, replace Items 7 and 9 with:

Item	Description	Part
7	O Ring	3865195
9	O Ring	3922070

On page 6-7, replace the first sentence in the first **Caution** with:

O-ring seals (where fitted) are fluoroelastomer rubber, and should not be immersed in solvents, as they will become damaged.

On page 6-9, Table 6-3, replace Item 8 with:

Item	Description	Part
8	O Ring	3922070

On page 6-12, Table 6-4, replace Items 2 and 6 with:

Item	Description	Part
2	O Ring	3865332
6	O Ring	3867920

On page 6-14, Table 6-5 replace Items 10 and 13 with:

Item	Description	Part
10	O Ring	3922070
13	O Ring	3865359

On page 6-15, Table 6-6, replace Items 2, 3 and 9 with:

Item	Description	Part
2	O Ring	3865174
3	O Ring	3865163
9	Check Valve	3966453

On page 6-18, Table 6-7, replace Items 6 and 12 with:

Item	Description	Part
6	O Ring	3865326
12	O Ring	3865163

On page 9-1, replace the second paragraph with:

The P5521 has a fluoroelastomer diaphragm and seals, and a maximum working pressure range of 10,000 psi / 700 bar.

