

# Fluke i30s

## AC/DC Current Clamp

**FLUKE**®

### Technical Data



The i30s current clamp is based on Hall effect technology for use in measurement of both dc and ac current. The i30s may be used in conjunction with oscilloscopes and other suitable recording instruments for accurate non-intrusive current measurement.

#### Electrical specifications

**Specified current range:** 30 mA to 30 A DC,  
30 mA to 20 A AC rms  
**Usable current range:** 5 mA to 30 A DC,  
30 mA to 20 A AC rms  
**Crest factor:** 1.4  
**Output sensitivity:** 100 mV/A  
**Accuracy (at +25 °C):**  $\pm 1\%$  of reading  $\pm 2$  mA  
**Resolution:**  $\pm 1$  mA  
**Load impedance:**  $> 100$  k $\Omega$   
**Conductor position sensitivity:**  $\pm 1\%$  relative to  
centre reading  
**Frequency range:** DC to 100 kHz ( $-0.5$  dB)  
**Phase shift below 1 kHz:**  $< 2$  degrees  
**Temperature coefficient:**  $\pm 0.01\%$  of reading/ $^{\circ}\text{C}$   
**Power supply:** 9 V Alkaline, MN1604/PP3,  
30 hours, low battery indicator  
**Working voltage (see Safety Standards section):**  
300 V ac rms or dc

#### General specifications

**Maximum conductor size:** 19 mm (.748 in) diameter  
**Output connection:** Safety BNC connector,  
supplied with safety 4 mm (.157 in) adapter  
**Output zero:** Manual adjust via thumbwheel  
**Cable length:** 2 m (6.56 ft)  
**Operating temperature range:** 0 °C to +50 °C  
( $-32$  °F to 122 °F)  
**Storage temperature range (with battery removed):**  
-20 °C to +85 °C ( $-4$  °F to 185 °F)  
**Operating humidity:** 15 % to 85 % (non-condensing)  
**Weight:** 250 g (.55 lb)

**Safety standards**

BS EN 61010-1: 2001  
 BS EN 61010-2-032: 2002  
 BS EN 61010-031: 2002

300 Vrms, Category III, Pollution Degree 2


Use of the probe on uninsulated conductors is limited to 300 V acrms or dc and frequencies below 1 kHz.

**EMC Standards**

EN 61326: 1998 +A1, A2, & A3

**Dimensions (HxWxD)**

183 mm x 71 mm x 25 mm (7.2 in x 2.8 in x 1 in)



**Ordering information**  
 i30s AC/DC Current Clamp



i30s connected to a Fluke 199C ScopeMeter.





**205 Westwood Ave  
 Long Branch, NJ 07740  
 1-877-742-TEST (8378)  
 Fax: (732) 222-7088  
 salesteam@Tequipment.NET**