



The Value Leader
www.tpi-thevalueleader.com

Combustible Gas Leak Detector

To find combustible gas leaks with this unit follow these simple instructions:


Press the ON/OFF key to power on the 725. All LED's will flash as the unit initializes. LED's will turn on sequentially as the 725 auto zeroes. When the solid green LED above the ON/OFF switch stays lit, the 725 is ready.

Remove the sensor cap from the top of the 725

Trace fuel lines, tanks and containers for gas leaks. Once a leak is detected, the 725 will sound an alarm and the red LED's will light showing level of leak.

To turn off the 725, press and hold the ON/OFF button for five seconds until the 725 turns off.

INDICATED LEAK LEVELS



Level 4: > 5000ppm Methane
Level 3: > 4000ppm Methane
Level 2: > 2000ppm Methane
Level 1: > 1000ppm Methane
Power: Green = ready for use
Red = replace batteries

Press to power on and zero. Hold down for 5 seconds to power off.

Optional A779 soft pouch available

Specifications:

Sensitivity > 1,000ppm combustible gases
Operating Temp 32° to 104°F or 0° to 40°C
Warm-up 7 to 10 seconds
Response Time Instantaneous
Battery 1.5 volt, AAA Alkaline (2 each)
Power Consumption DC 200mA
Battery Life 3 hours continuous use

THIS INSTRUMENT is in compliance with EN 50 081-1:1992, EN 50 082-1:1992, IEC 801-2:1984, IEC 801-3:1984, IEC 801-4:1988



Test Products International, Inc.
503-520-9197 Fax: 503-520-1225



Test Products International, Ltd.
905-693-8558 Fax: 905-693-0888



Test Products International Europe
+44 (0) 1293 561212 Fax: +44 (0) 1293 813465

email: info@tpi-thevalueleader.com



6 37623 00725 5

MADE IN KOREA