




ResearchIR 4

User's Guide



Document Number: 29354-000

Revision: 11

Issue Date: January 26, 2018

Copyright

© FLIR Systems, Inc., 2001-2018. All rights reserved worldwide. No parts of the software including source code may be reproduced, transmitted, transcribed or translated into any language or computer language in any form or by any means, electronic, magnetic, optical, manual or otherwise, without the prior written permission of FLIR Systems.

This manual must not, in whole or part, be copied, photocopied, reproduced, translated or transmitted to any electronic medium or machine readable form without prior consent, in writing, from FLIR Systems.

Names and marks appearing on the products herein are either registered trademarks or trademarks of FLIR Systems and/or its subsidiaries. All other trademarks, trade names or company names referenced herein are used for identification only and are the property of their respective owners.

FLIR Systems, Inc.

Customer Support: 800-GO-INFRA (464-6372)

flir.custhelp.com

Service: 866-FLIR-911 (354-7911)

www.flir.com

Table of Contents

Revision History	7
1 Introduction	8
2 Installation.....	10
2.1 Software and GigE Driver Installation.....	10
2.2 Camera Link and CoaXPress (CXP) frame grabbers	13
2.3 Software License Activation	14
2.3.1 License Activation Process	14
2.3.1.1 Online Activation	14
2.3.1.2 30-day Trial (Evaluation)	14
2.4 Installation Issues.....	14
2.4.1 Activation Issues	14
2.4.2 Using the RIR 4 “RUS” license tool	16
2.4.2.1 Offline Activation	16
2.4.2.2 Transferring a license from one PC to another	17
2.4.3 Firewalls.....	18
2.4.4 Adding Firewall Exceptions	18
2.4.5 Troubleshooting steps for USB-Video cameras.....	20
2.5 Configuring Pleora Drivers	20
2.6 PC’s with multiple network interfaces	21
3 Getting Started.....	22
3.1 Connecting to a camera	22
3.2 Recording a file	23
3.3 Playing back the recorded file	24
4 User Interface	26
4.1 Title Bar	27
4.2 Main Image Window.....	27
4.3 Image Enhancement Tool	28
4.3.1 Scaling Mode	28
4.3.2 AGC Modes	29
4.4 Color Bar.....	29
4.4.1 Isotherm ROIs.....	30
4.5 Analysis Toolbar	31
4.6 Units and display controls	32
4.6.1 Units Selector.....	32
4.6.2 “Info” Button	33

4.6.3	Image Invert/Revert [Max]	33
4.6.4	Saturation/Clipping On/Off	33
4.7	Plot and Acquisition Toolbar	34
4.7.1	Plot Tools	35
4.7.1.1	Statistics Viewer	36
4.7.1.1.1	Stats Toolbar	36
4.7.1.1.2	Stats Table	37
4.7.1.1.3	Measurement Functions [Max]	38
4.7.1.1.4	Adding and Editing Measurement functions	39
4.7.1.2	Profile Plot [Max]	39
4.7.1.3	Temporal Plot	40
4.7.1.4	Histogram Plot [Max]	41
4.7.1.5	Oscilloscope Plot [Max]	41
4.7.1.6	Opening a plot	42
4.7.1.7	Managing docked plots	44
4.7.1.8	Plot Toolbar	44
4.7.1.8.1	Plot Cursors [Max]	46
4.7.1.8.2	Plot Legend	46
4.7.1.9	Plot Properties	47
4.7.2	Bad Pixel Tool [Max]	47
4.7.3	Acquisition Tools	48
4.7.3.1	Record Settings	48
4.7.3.2	Record Mode	49
4.7.3.2.1	Record to Memory Mode [Max]	49
4.7.3.2.2	Record to Disk Mode	50
4.7.3.2.3	Periodic Recording	50
4.7.3.3	Record Options	50
4.7.3.4	Pre-Trigger Recording Options [Max]	51
4.7.3.5	Record Conditions	51
4.7.3.6	File Naming Options	51
4.7.4	Preset Selector [Max]	52
4.8	Toolboxes [Max]	53
4.8.1	Object Parameters Toolbox	54
4.8.2	Camera Control Toolbox	54
4.8.3	Source Information Toolbox	54
4.8.4	Minimap Toolbox	54
4.8.5	Segmentation Toolbox [Max]	54
4.8.6	Filters Toolbox [Max]	55
4.8.6.1	MATLAB Filter	56
4.8.6.2	Passing metadata between ResearchIR and MATLAB	57

4.8.7	HSDR Toolbox [Max+]	57
4.8.7.1	First-time Setup.....	57
4.8.7.2	Pairing.....	58
4.8.7.3	HSDR Triggered recording (DVR express Core only).....	58
4.8.7.4	Movie Recording	58
4.8.7.5	Movie Playback.....	59
4.8.7.6	Movie Extraction	59
4.8.8	File Toolbox	59
4.9	Main Menu Bar.....	60
4.9.1	File Menu	60
4.9.1.1	Export Dialog	61
4.9.1.2	Extract Dialog [Max+].....	66
4.9.1.3	Batch Extract Dialog [Max+].....	67
4.9.2	Edit Menu.....	67
4.9.2.1	Spatial Calibration [Max]	67
4.9.2.2	Preferences Dialog	68
4.9.3	Camera Menu	71
4.9.3.1	Camera Selector	72
4.9.3.2	Camera Control.....	73
4.9.3.3	User Calibration [Max].....	73
4.9.3.3.1	Load Calibration	73
4.9.3.3.2	Save Calibration.....	74
4.9.3.3.3	Perform Calibration	74
4.9.3.3.3.1	Wizard Modes.....	74
4.9.3.3.3.2	Standard Calibration Wizard.....	75
4.9.3.3.3.3	Review Calibration Coefficients	84
4.9.3.3.3.4	Calibration Plots.....	85
4.9.3.4	PC Side Correction (Max Only)	86
4.9.3.5	Creating a PC-Side NUC in ResearchIR	87
4.9.4	View Menu	91
4.9.4.1	Zoom Menu.....	91
4.9.4.2	Palette Menu.....	91
4.9.4.3	Edit Palette Dialog	92
4.9.4.4	Startup Dialog	92
4.9.5	Tools Menu	93
4.9.5.1	Color Bar.....	95
4.9.5.2	Flying Spot Meter.....	95
4.9.5.3	File Operation	95
4.9.5.4	Make Self Viewing File [Max]	96

4.9.6	Hotkeys.....	96
4.9.7	Help Menu	97
4.10	Metadata display	97
4.10.1	Time in ResearchIR	98
4.10.1.1	Time Source	98
4.10.1.2	Time Reference	98
5	Camera Control.....	100
6	Other Utilities	102
6.1	REdit.....	102
6.2	CalibratIR.....	102
6.3	File AbatIR	102
6.4	CNUC Manager	103
7	Acknowledgements	104

Revision History

Version	Date	Initials	Notes
A	FEB 14 2014	RIM	V4.0 beta release
1	MAR 05 2014	RIM	V4.0 release
2	JUL 1 2014	RIM	V4.10 release
3	FEB 10 2015	RIM	V4.20 release
4	FEB 19 2015	RIM	Added HSDR triggered recording information
5	APR 08 2015	RIM	Updated for 4.20.1. Added updated info on troubleshooting activation issues.
6	MAY 05 2015	RIM	Fixed section numbering errors
7	MAY 12 2015	RIM	Updated for 4.20.2, added RUS tool info
8	NOV 20 2015	RIM	Updated for 4.30.0
9	JUN 30 2016	RIM	Updated for 4.30.2, and localization
10	JAN 12 2017	JAT	Updated for 4.40.2, copyright only
11	JAN 26, 2018	AAR	Updated with frame grabber info

1 Introduction^[taw1]

ResearchIR is a thermographic data acquisition and analysis application designed for the R&D environment.

Some of the key features of ResearchIR include:

- Easy to use, customizable workspaces
- Supports factory and user calibrations
- Supports factory and user Non-Uniformity Correction (NUC)
- Supports all FLIR R&D cameras
- Supports high-speed data acquisition (HSDR)
- Supports SEQ, SAF, PTW, ATS, FCF and RJPG file formats
- Supports Windows Vista/Win7/Win8 (32-bit and 64-bit)
- A combination of traditional and new capabilities designed to provide a complete solution for the R&D user

Recommended Minimum Hardware/Software:

OS: Windows7 (32-bit or 64-bit) or higher

Processor: Intel Pentium 4 3.0 GHz or faster recommended

RAM: 8GB or greater is recommended.

Hard Disk Free Space: 1GB (free space)

Network Adapter: Gigabit Ethernet

USB: One USB 1.1 (or higher) port for the license dongle, USB 2.0 required for cameras that support USB video. Note: a USB 3.0 port is required for HSDR Core2 operation.

End User License Agreement (EULA)

Computer Software License Agreement for FLIR Systems, Inc.

NOTICE TO USER: This is a Contract.

FLIR is willing to license the Software identified as ResearchIR to You only on the condition that You accept all of the terms and conditions contained in this License Agreement. Please read the terms and conditions carefully before installing or using the software. By installing or using the software, You are indicating Your acceptance of this Agreement. If You do not agree to the terms of this Agreement, FLIR is unwilling to license the software to You. In that event, You should not install or use the Software.

1. **Scope of Permitted Use:** In this Agreement, "Software" shall mean the machine readable software programs and associated files identified as ResearchIR, any modified version, upgrades or other copies of such programs and files, and any associated media and printed materials. You have the nonexclusive right to install and use the Software on a single central processing unit. Only one copy of Your licensed Software may be made by and for You for backup purposes. You may not make or distribute any other copies of the Software, nor does this license permit You to allow multiple users access to a single copy of the Software via a computer network. You may not resell or redistribute the Software.

2. **Proprietary Rights:** The Software is owned by FLIR and/or its suppliers and You acknowledge that the structure, organization and/or code to the Software is/are the valuable trade secret(s) of FLIR. You agree to hold such trade secret(s) in confidence. You further agree not to translate, decompile, disassemble, modify, reverse engineer, or otherwise attempt to discover the source code of the Software, in whole or in part. The Software is protected by United States Copyright Law and International Treaty provisions.

3. **No Other Rights:** FLIR retains title and ownership to any and all copies of the Software whether transmitted electronically or recorded as diskette copies and all subsequent copies of the Software, regardless of the form or media in or on which the original and other copies may exist. Except as expressly stated, this Agreement does not grant You any rights to patents, copyrights, trade secrets, trademarks or any other rights with respect to the Software.

4. **Term:** The Agreement is effective until terminated. This Agreement shall terminate automatically upon failure by You to comply with any of its terms. FLIR also has the right to terminate Your license immediately at any time. Upon any such termination, You must destroy the original and any copies of the Software. However, any terms of this Agreement which can reasonably be construed to protect FLIR will survive its termination.

2 Installation



FULL Administrator privileges are required to install ResearchIR and associated drivers. If you install ResearchIR on a computer managed by an IT department, please contact your System Administrator for assistance. After installation, ResearchIR can be run as a normal user.

2.1 Software and GigE Driver Installation

The installation program will install the ResearchIR application. In addition, it will install necessary support software if it detects that they are not installed on your system. The support software includes:


- Pleora GigE (eBUS) drivers
- FLIR camera drivers
- Vircam SDK
- SafeNet software license support

In general, you can accept all defaults for any dialog you see during the install process. Some drivers may cause Windows to display a message that the driver is “unsigned”. This is normal. Choose the option to continue with the installation.




Important Note: If you have an ResearchIR dongle, please do not insert it until after you have installed the software.


Step 1: Launch the downloaded .EXE file. (If you downloaded the .ISO file, that can be used to create an install CD using most CD burner applications.) You will be guided through the installation wizard.



Step 2: Accept the license agreement.



Step 4: Select any additional options.




If this is an initial installation of ResearchIR, then you will have the option to install additional components. Components that are recommended for all users will be checked by default. If the component is already installed, it will be grayed out.

If you have used older versions of FLIR software (or have other GigE cameras, you may already have an existing installation of the Pleora driver software. ResearchIR supports many versions of the Pleora GigE drivers. However, not all versions are compatible with all cameras. The installer may indicate that certain cameras are not supported by your current version. If you need compatibility with these cameras you will need to uninstall the older Pleora software using Control Panel. Then run the ResearchIR installer again. If no Pleora software is installed you will have the option to install the latest supported version.

Uninstalling older Pleora Driver software


STEP 1. Revert to Manufacturer’s NIC driver. Use the Driver Installation Tool to remove the High Performance or eBUS driver and revert back to the Manufacturer’s NIC driver. This will require a reboot.

STEP 2. Use Control Panel → Add/Remove Programs to uninstall the older Pleora software (see figure below). If you have Pleora 2.2 or newer you should remove iPORT Vision Suite, eBUS Driver Suite, and GEV Suite components from the Control Panel list. Then run the ResearchIR installation again and the latest supported version will be installed for you.



Step 5: ResearchIR will install drivers and libraries. The items displayed will vary depending on what components are already installed on the PC. A green check mark will appear next to each successfully installed component. If you see any red “x” indicators, contact FLIR support at <http://flir.custhelp.com>.

Step 6: Click “Finish”. ResearchIR installation is now complete.



2.2 Camera Link and CoaXPress (CXP) frame grabbers

ResearchIR supports a limited selection of frame grabber models that can be used to connect to cameras with Camera Link and CXP interfaces. ResearchIR can use these interfaces for both control and digital video. For help installing these devices, please refer to the FLIR ResearchIR Frame Grabber Installation Manual, document 4213512. The PDF download is available here:

http://flir.custhelp.com/app/answers/detail/a_id/1658.

PLEASE NOTE: When using a Camera Link or CXP frame grabber with ResearchIR, it will be necessary to install very specific driver and runtime libraries that FLIR provides on this folder on Box.com:

<https://flir.app.box.com/folder/29645234728>.

You cannot just install the drivers and runtime libraries that came with the frame grabber – they will not work!

The following frame grabber models are supported:

Camera Link

- Silicon Software microEnable IV AD4-CL (Base/Medium/Full)
- Dalsa Xcelera-CL PX4 (Base/Medium/Full)
- IO Industries Core2 CL (HSDR)

Note: Other frame grabbers in the DALSA X64, DALSA Xcelera, and Silicon Software microEnable IV product lines may work, but are not tested. The HSDR accessory recorder system acts like a frame grabber and brings image data into ResearchIR through an eSATA to USB 3.0 converter cable connected to the PC. The HSDR is available in either Camera Link or CoaXpress varieties.

CoaXPress

- Active Silicon Firebird (FDB-2XCXP6, FDB-4XCXP6)
- Bitflow Cyton CXP-4

- Euresys CXP Coaxlink Quad G1
- IO Industries Core2 CXP (HSDR)

Software License Activation

ResearchIR 4 requires a license key in order for the software to run. Prior to version 4.20, the license key was in the form of a USB dongle. Starting in version 4.20, the software can also be activated online by entering a provided product key. This method will require an internet connection for the activation process. (After the software is activated, an internet connection is no longer required.) Going forward, online activation will be the default method for license activation. (However, existing USB keys will continue to work and can be ordered if needed.)

2.2.1 License Activation Process

When ResearchIR is first launched it will automatically search for a valid license. If multiple licenses are found (for example both software and USB key are found) then the highest level license will be used). If no license is detected, then the user will see the Activation dialog. This window will present the user with three options:

1. Enter a Product Key
2. Insert a Hardware (USB) dongle
3. Start a 30-day trial (evaluation)

2.2.1.1 Online Activation

If Option 1 is chosen, the user will be presented with a screen where product can be registered. User must enter the following:

1. Product Key
2. User's name **OR** Company Name
3. Email address

2.2.1.2 30-day Trial (Evaluation)

If you choose the 30-day trial, you will have the option to try either the Max or Max+HSDR editions. The 30-day trial does not require an internet connection and the trial will start immediately. If you later activate a product key for a lower edition, the trial will continue to run until it expires and then the product key will take over. Trial periods cannot be renewed or extended. The trial time remaining can be viewed under the Help>>Gather Information menu item.

2.3 Installation Issues

2.3.1 Activation Issues

If you encounter issues with product activation the following steps should help to isolate the problem.

ResearchIR will return an error code if activation fails. Please provide the error code when requesting technical support.


E1. Invalid Product Key: Please make sure you entered your product key correctly. Then try again.

E2. No more activations available: By default, all product keys can be activated 5 times. Please contact your local representative if you need more activation codes. Then try again.

E3. Cannot find master dongle: This indicated that a network connection was established to the license server, but the system is down. Please try again later. If the situation persists please contact FLIR Customer Support.


E4. Cannot connect to server. Possibly because of blocked port but cannot distinguish between that or other connection issues: Your product failed to successfully activate. Please make sure you are connected to the internet and that port 8080 is not blocked by a software or hardware firewall. Then try again. If the situation persists please contact FLIR Customer Support.

Another way to troubleshoot activation issues is to bypass the activation Wizard within ResearchIR and use the web-based activation portal. To reach the portal, enter <http://researchir.flir.com:8080/ems/customerLogin.html> into your web browser. If you get see the following on your screen you have Internet connectivity to the license activation system (so a blocked port is not the problem.)



Enter your Product Key. If your key has been previously activated, then you will see information about the number of activation and the amount remaining. If you see a screen asking for registration information this indicates that the product has never been activated. You can enter your registration information on the web portal. If the activation succeeds, then the license on your PC will be updated and the next time ResearchIR is launched it should skip the activation wizard.

If after following these steps your activation issue is still not resolved, please contact FLIR Customer Support at <http://flir.custhelp.com>. Be sure to include information in your problem description about any error codes and also include the “Gather Information” text file which can be generated from the menu item under the Help menu. This will provide the support agent with more details that can be used to troubleshoot your problem.



2.3.2 Using the RIR 4 “RUS” license tool

The RIR4 RUS utility is designed to handle situations where internet access is not available for online license activation. The RUS tool can be used for the following situations:


- Offline activation via web customer portal
- Moving a license from one PC to another

For both cases, the user can complete the entire process without the need to submit a ticket to FLIR customer support. The RUS tool can be downloaded using the following link: <https://flir.box.com/s/n7cu7kkizcvee1qwbsndakkwpzq3ewy>.


2.3.2.1 Offline Activation

If a customer PC with RIR4 does not have internet access, but they have another PC that does, the RUS tool can be used to create a file that can be used to generate a activated license file through the web customer portal.

- Launch the RUS tool



- On the “Collect Status Information” tab, click the “Collect Information” button. You will be prompted to save a “C2V” file.
- Go to the ResearchIR customer portal at <http://researchir.flir.com:8080/ems/customerLogin.html>. You should see the following screen in your browser. Enter your product key and click Login.



- d) If the key has not been used before, you will see a screen asking to register. Enter your registration information and click Save.
- e) On the next screen, click the “Offline Activation” button in the upper right corner.
- f) A “Generate License” dialog box will appear. In the middle is a link to download the RUS tool if you have not already downloaded it. At the bottom is a field called “Upload C2V”. Click the “...” button to browse and select the file you created with the RUS tool. Then click “Generate”.
- g) Under “Activation Details” there will now be a link to “Download V2C File”

Activation Details			
Key ID	Key Type	Activation Date	Comments
586098184648146402	SL-AdminMode	2015-05-11	Download V2C File

- h) In the RUS tool, go to the “Apply License File” tab. Next to the “Update File”, click the “...” to browse for the V2C file you just downloaded, then click “Apply Update”.
- i) RIR4 should now see the license when launched.

2.3.2.2 Transferring a license from one PC to another

You will need to install and run the RUS tool on both the source and recipient PCs. On the recipient PC, launch the RUS tool and click on the “Transfer License” tab. Follow the steps presented in the dialog box.

Collect Status Information | Apply License File | **Transfer License**

To transfer (rehost) a license from one computer to another, you run the RUS program on both computers. On each computer, select the **Transfer License** tab and perform the appropriate step.

Collect information about the recipient computer

Step 1: On the computer to which you want to transfer the license (the "recipient computer"), collect and save information about the computer.

Save recipient information to: ...

Collect and Save Information

Generate the license transfer file

Step 2: On the computer that currently contains the license (the "source computer"), select the license to transfer, read the recipient information file and generate a license transfer file.

Key Type	Key ID	Products
SL-AdminMode	586098184648146402	ResearchIR Products

Read the recipient information file from: ...

Generate the license transfer file to: ...

Generate License Transfer File

- a) Next to “Save recipient information to”, click the “...” button to assign a name to the .ID file.
- b) Take the created file to the “source” PC. Run the RUS tool and go to the “Transfer License” tab. In the “Generate license transfer file” section, click the “...” next to “Read the recipient information file” to browse for the .ID file created in step 5.
- c) Click the “...” next to the “Generate the license transfer file to” to assign a name for the H2H file.
- d) Click the “Generate License Transfer File” button at the bottom of the window.
- e) Take the generated H2H file to the recipient PC. In the “Apply License” tab of the RUS tool, click the “...” button next to the “Update File” field and browse for the

H2H file creates in step 8. Click the “Apply Update” button. The license file should be updated and RIR 4 should see the license when launched.

2.3.3 Firewalls


Firewalls will generally block connectivity to a camera. The ResearchIR installer will attempt to configure a firewall exception to the Windows firewall service. If your computer is using a third party firewall, this service may be disabled. This could cause the ResearchIR installation to abort and it will appear as if nothing was installed. To work around this issue, you will need to browse the installation CD and look for the “MSI” folder. Inside there are two msi install files (one for 32-bit and one for 64-bit). Run the appropriate file. This will run the install in a “verbose” mode. If the firewall service is disabled, you will see the error below. Click the “Ignore” button and the installation will continue normally. (NOTE: You may need to manually configure a firewall exception if you have trouble detecting and connecting to your camera.)



2.3.4 Adding Firewall Exceptions

If you cannot turn off the firewall then you may need to manually add some “exceptions”. These exceptions will allow specific programs to get past the firewall. The “Vircam” exception applies to SC2500/5000/7000 series cameras. The “EBservice” exception applies to SC2500/2600/5000/7000 series cameras.

The firewall is generally accessed through the Windows Control Panel.



Choosing the option to allow a program through the firewall will present you with a list of current exceptions. If the “Allow another program...” button is disabled then you will need to click the “Change settings” button. If your computer is controlled by an IT dept

you may not have permission to change these settings and you will have to get your IT person involved.

Allow programs to communicate through Windows Firewall

To add, change, or remove allowed programs and ports, click Change settings.

What are the risks of allowing a program to communicate?

[Change settings](#)

For your security, some settings are managed by your system administrator.


Allowed programs and features:

Name	Domain	Home/Work (Pri...	Public	Group Policy
<input checked="" type="checkbox"/> 500 IPSEC	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Yes
<input checked="" type="checkbox"/> A320/325	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Yes



[Details...](#) [Remove](#)

[Allow another program...](#)

Once you get past this point you will see the “Add a Program” dialog.



Click the “Browse...” button. The two exceptions you should add are for Vircam and EBservice. The screenshots below show you the path and file that should be selected. Enable all the check boxes for Domain/Work/Public.




2.3.5 Troubleshooting steps for USB-Video cameras

Items 1-3 should be done before connecting a camera. Anything that needs to be disabled can be enabled once the camera connection is established.


1. Make sure the camera is set to USB Video mode. This has a faster frame rate than Network mode and is not blocked by Firewalls.
2. Switch off the wireless radios (Wi-Fi and Bluetooth.)
3. Remove all USB devices, including the mouse, before connecting the camera. You can usually plug the mouse back in after a successful connection.
4. Try using a different USB port. Some ports run through an internal hub. In general don't use front USB hubs.

2.4 Configuring Pleora Drivers


Launch the Driver Install Tool located in the Pleora Technologies group.



Highlight the network adapter you want to configure and use the drop down list to select the eBUS Universal Driver. Click the Install button.



Once install is complete, the “Current Driver” field will be updated. Press the “Close” button to exit and reboot if prompted to do so.



If you need to configure more than one network adapter, it is recommended that you only configure one port at a time. See Sec 2.5 for more information on configuring PCs with multiple network interfaces.

2.5 PC's with multiple network interfaces

If your PC or laptop has multiple network interfaces (wired or wireless) there are some things to watch for that can cause issues with connecting to cameras.


It is important that both network adapters not be in the same network subnet. This is one reason that we recommend using static IP addressing because a DHCP server will often give two consecutive addresses to the PC's NICs. When this happens, the network routing can get confused in the Pleora driver, and traffic intended for the camera can get routed to the wrong port.

This can happen with static IP addressing if both addresses are in the same subnet. So if you have two ports, don't set one to 192.168.10.1 and the other to 192.168.10.2. Those are in the same subnet and will cause problems. Instead use 192.168.10.1 and 192.168.11.1. Or use 192.168.10.1 for the wired port that is connected to the camera and set the wireless port to DHCP to work with your corporate network.

3 Getting Started

3.1 Connecting to a camera


1. Connect the camera to the computer and power the camera on.
2. Run ResearchIR, the startup dialog should be shown. If you have it set to not show on startup or if ResearchIR is already running, you can show it by selecting View→Startup Dialog in the file menu. Click the camera you want to connect to.



NOTE: If your camera isn't shown, you can click "Search Again" to look for it.

NOTE: It may take a few seconds to connect to the camera, especially if this is the first time ResearchIR has connected to this camera.

3. The camera image should now be displayed.




3.2 Recording a file


1. Move your mouse over the record toolbar button and the click on the configure icon.



2. Set the recording and file naming options you want and click “OK”.



3. Move your mouse over the record toolbar button and click the record button (or use F5 if that option was enabled).




4. ResearchIR will record a file. Wait for the status to display “Done” and the “Close” button to appear then click “Close”.




3.3 Playing back the recorded file


1. Make sure the File toolbox is visible by expanding the toolbox window and the File toolbox if they are not already expanded. (NOTE: If you don't have a dongle and are running ResearchIR Standard then you won't have access to the Toolboxes. You will have to use the File Menu to access your recorded data.)



2. Click the show record folder button to display the record folder in the file toolbox then double click on the file you just recorded.




3. The file should now be displayed.



NOTE: Files can also be played back by opening them with File→Open in the File menu.

4 User Interface



The ResearchIR main window is comprised of ten main regions:

1. Title Bar (Sec. 4.1)
2. Main image window (Sec. 4.2)
3. Image Enhancement Tool (Sec. 4.3)
4. Colorbar (Sec. 4.4)
5. Analysis Toolbar (Sec. 4.5)
6. Unit and Display controls (Sec. 4.6)
7. Plot and Acquisition Toolbar (Sec. 4.7)
8. Toolboxes. (ResearchIR Max Only). (Sec. 4.8)
9. Main Menu Bar (Sec. 4.9)
10. Metadata display (Sec. 4.10)

4.1 Title Bar



Indicates what ResearchIR edition is running. There are three options:

- ResearchIR : indicates that the Standard edition is running.
- ResearchIR Max: indicates that all add-on modules except HSDR are enabled
- ResearchIR Max + HSDR: indicates that all add-on modules are enabled.

NOTE: This manual describes features for all three editions. Some features are not supported in all editions. Throughout the user manual a feature section heading may have a notation in [] brackets. This notation will indicate the minimum edition level that will support this feature. Your license key will indicate and control what edition level you have access to. If you wish to upgrade to a higher edition, please contact your sales representative for more information.

4.2 Main Image Window

The Main Image Window displays either live imagery from a camera or stored images and movies. By default, the image size stretched as the application window is resized. This can be changed by selecting a different zoom factor in the View→Zoom menu. When a previously recorded movie is loaded, the movie player controls will be displayed at the bottom of the image window.




Icon	Name	Description
	Advance/Rewind a single frame	These arrow buttons at the ends of the playback slider show the previous or next frame in the movie for each press.
	Current frame position	Indicates the currently displayed frame number. The current frame position can be changed by dragging the handle with the mouse or can be entered directly by double clicking on the number.
	Playback limits	The blue bar indicates the playback limits. If the user only wishes to playback a subset of the data, the black handle can be dragged to the desired start/stop frame positions. The start/stop position can be entered directly by double clicking on the number.
	Rewind Controls	Rewind with normal speed, fast speed, and jump to beginning.
	Forward Controls	Forward with normal speed, fast speed, and jump to end.
	Stop	Stops playback at current frame vertex points can be moved individually.

NOTE: The Start and End frame for the play limits are also used for the Extract feature discussed in Section 4.9.1.2.

4.3 Image Enhancement Tool

This tool allows the user to control the scale limits and how the color palette is mapped to the data. A real-time histogram of the data is display for each image frame so that the user can see how the data values are distributed. This powerful tool can allow the user to see amazing detail even in low contrast imagery.



Icon	Name	Description
	Histogram	Shows distribution of data against current scale. Color bar palette is used for Histogram colors.
	Scale limits	The numerical fields to the left and right of the histogram indicate the lower and upper scale limits. Can be adjusted using the options described in Sec. Error! Reference source not found. The upper and lower scale limits are also reflected in the colorbar limits to the right of the image.
	Scaling Mode	Choose Scaling mode using the drop down list. Options are described in Sec. 4.3.1
	Automatic Gain Control (AGC).	Choose AGC mode from drop down list. Modes are described in Sec. 4.3.2. Slider to the right adjusts the “intensity” of the AGC.
	DDE Level	This controls the DDE “intensity”. This control is disabled for other AGC modes.
	Stop	Stops playback at current frame vertex points can be moved individually.

4.3.1 Scaling Mode

ResearchIR offers four choices for determining the scale limits.

Scale Limits From	Description
Image	ResearchIR looks at the entire image to determine the min and max values for the scale
Active ROI [Max]	ResearchIR looks only at the values within the selected ROI to determine min and max values for the scale

Segmentation [Max]	ResearchIR uses the limits in the Segmentation Toolbox to determine min and max values for the scale
Manual	The user manually sets the limits either by: <ul style="list-style-type: none"> (1) typing numbers in the boxes at the left and right corners of the data histogram (2) using the mouse to drag the edges of the shaded box on the histogram


4.3.2 AGC Modes

Although the camera data is 14-bit/16-bit, the color palettes used for display are 8-bit. Automatic Gain Control (AGC) algorithms are used to control how the colors are applied to the data. ResearchIR offers three choices for AGC algorithms. These algorithms can improve image detail and contrast.

Algorithm	Description
Linear	Simple linear mapping. Evenly distributes the color in the palette across all count values.
Plateau Equalization (PE)	Non-linear, histogram based algorithm. Provides good contrast for almost all scenes. The slider, next to the drop down controls the aggressiveness of the algorithm
Digital Detail Enhancement (DDE) [Max]	FLIR proprietary algorithm. Runs in addition to PE. Enhances image detail. The drop down next to the PE slider controls the aggressiveness of the DDE algorithm.
Advanced Plateau Equalization (APE) [Max]	FLIR proprietary algorithm. Enhances image detail. The drop down next to the PE slider controls the aggressiveness of the APE algorithm.

4.4 Color Bar

The Color bar shows the relationship between the color palette and the data values in the currently selected units. The palette can be changed by using the View→Palette menu. The scale limits and the color distribution are controlled by the Image Enhancement Tool.






The current segmentation levels are also displayed on the color bar as full width shaded regions in the currently selected segmentation colors.

Active Isotherm ROIs are displayed as half width shaded regions. Isotherm limits can be adjusted by dragging the ROI on the color bar. Isotherms ROIs can be deleted by clicking the isotherm on the color bar and pressing the Delete key.

4.4.1 Isotherm ROIs

Isotherm ROIs are added by using the buttons above the color bar.

Icon	Type	Description
	Interval	User sets range by dragging on color bar
	Above	User click on color bar to set lower limit
	Below	User clicks on color bar to set upper limit

Endpoints can be adjusted by dragging the limits on the color bar. Right click to set properties.

Although there are tools to create Above/Below/Interval isotherms, any isotherm can be converted to any type using the min/max properties. Right click the isotherm and choose Properties from the pop up menu.

















Above isotherms have the max box checked. Below isotherms have the min box checked. Interval isotherms have neither box checked.




The Track By option tells ResearchIR how to display the isotherm as the scene change over time.

Isotherm Track By:	Description
Value	Isotherm stays locked to the value limits. This is the most common setting.
Color	Isotherm is locked to the color bar values. As the scene changes and the AGC adjusts the color assignments the isotherm numerical limits can change.

4.5 Analysis Toolbar

The Analysis Toolbar allows the user to draw a Region of Interest (ROI) using a variety of drawing tools.


Icon	Name	Description
	Select/Edit ROI	With this tool selected, the user can mouse over an ROI. The ROI can be dragged to move it or the user can grab a “handle” to resize the ROI
	Box ROI	Draw an arbitrary rectangular box.
	Ellipse ROI [Max]	Draw an arbitrary ellipse
	Line ROI [Max]	Draw a straight line
	Bendable Line ROI [Max]	Draw a multi-segment line. Each left click adds a vertex. Right click to stop adding points. Using the Edit ROI tool the vertex points can be moved individually.
	Polygon ROI [Max]	Draw a closed polygon. Left click to add a vertex. Right click to close the polygon. Using the Edit ROI tool the vertex points can be moved individually.
	Freehand ROI [Max]	Draw a freehand shape. Hold down the left mouse button and drag to draw. Release the left button to close the ROI.
	Spot Cursor ROI [Max]	This ROI reads the value of a single pixel.
	Measurement Cursor ROI	Reads the average of a 3x3 box of pixels. Standard practice in thermography is to use a 3x3 block of pixels for measurements.
	Add/Remove Point [Max]	Add/remove vertex from Bendable Line or Polygon. Click on a vertex to deleted it or click on a line segment to add a vertex.
	Rotate ROI [Max]	With this tool selected, mouse over an ROI to activate the “handles”. Drag the ROI to rotate it.
	Select Next ROI	Select next ROI. Use to select very small or overlapping ROIs
	Delete current ROI	Select an ROI and then click this button or press Delete key to remove it
	Delete All ROIs	Clears all ROIs
	Load ROIs [Max]	Load an ROI descriptor file
	Merge ROIs [Max]	Load ROIs from file and add to existing ROIs

Icon	Name	Description
	Save ROIs [Max]	Save an ROI descriptor file
	Show/Hide ROIs	Allows user to turn on/off the display of ROIs without having to delete and add them
	Isotherm ROI [Max]	This tool is located above the color bar. See section 4.4.1

4.6 Units and display controls

4.6.1 Units Selector

This dropdown allows the user to choose the data units. The units available depend on the type of calibration done with the camera.



Factory calibrated cameras typically allow:

- Counts [Max]
- Object Signal (N/A for SCx000 camera) [Max]
- Temperature (factory)
- Radiance (factory) [Max]


User calibrated cameras typically allow:

- Counts
- Radiance (user) [Max]
- Temperature (user) [Max]


Unit	Description
Counts [Max]	Digital value of pixel. Often referred to as “raw data” but can be NUC corrected. Cameras typically have a 14-bit A-D converter so the resulting values will be in the range of 0-16383
Object Signal [Max]	Typically available only on handheld cameras. This NUC corrected but is proportional to radiance. Typically not of use to the user but used internally for diagnostic purposes.
Temperature	If the camera is calibrated, the raw pixel counts can be converted to units of temperature.
Radiance [Max]	Measure of IR power per unit area of the target. Units of Watts/(cm ² -steradian).

4.6.2 “Info” Button


The “Info” button gives the user a quick preview of file naming and acquisition settings. Hovering the mouse over the “i” will show the info panel. Clicking the “i” button will make the info panel a floating box that stays open until you close it.




4.6.3 Image Invert/Revert [Max]

The  buttons will cause the image to be flipped vertically or horizontally. This flipping is done in software. Some cameras also have image flipping controls, but activating these will require the NUC to be updated. Flipping the image on the computer side will not affect the NUC.

4.6.4 Saturation/Clipping On/Off




With the , control “on”, the software will display color indicators on the image and colorbar to show if data in the image is exceeding calibration limits. Turning this control off suppress these visual indicators.

4.7 Plot and Acquisition Toolbar




This toolbar provides quick access to plotting and acquisition functions. The icon for the currently selected tool is displayed. When the mouse is placed over the current tool a drop down menu is displayed to show other available tools.

Icon	Tool	Description
	Plot Tools	User can choose from a selection of plots for each ROI. Only the Stats table and Temporal plot are available in ResearchIR Basic.
	Bad Pixel Tool [Max]	Allows user to manually add a bad pixel to the Bad Pixel Map (BPM). Click the reticle tool and use the mouse or arrow keys to place the reticle over the pixel to be marked. Click the +/- button to add or remove the pixel from the BPM
	Acquisition Tools	Allows user to start or stop the acquisition, record a snapshot or movie and to set acquisition options.
	Preset Selector [Max]	Allow user to select the displayed preset. Only available if camera supports Preset Sequencing (PS) and PS or Superframing must be active.
	Arm Recording [Max]	This is optional for pre-trigger recordings. This button pre-allocates memory; otherwise, memory is allocated at the time the recording is triggered. When requesting a large memory buffer this can add a delay to the start of recording. If you are triggering off a header flag, then it is best to Arm the recording so you can be sure the first flagged frame is captured
	Export Bitmap	This is a quick export option. It will capture a full image with no border and write a PNG image to the Picture>>FLIR Systems folder. IT DOES NOT USE THE NORMAL EXPORT OPTIONS
	Toggle File Operation [Max]	Use this button to easily turn the File Operation feature On and Off.

Icon	Tool	Description
	UltraMax File detected	This button will appear when an UltraMax file is opened. Click the button will show two options. (1) Upgrade and replace current file with UltraMax image or (2) Backup the existing file before upgrading the file. The file will be backed up to a folder within the current folder. [NOTE: once a file is upgraded it cannot be downgraded again.]
	Apply PC-side NUC [Max]	Turns PC-side NUC on or off. Button only active is PC-side NUC has been done.
	Apply PC-Side Bad Pixel Map [Max]	Turns PC-side BPM on or off. Button only active if PC-side BPM has been done.

4.7.1 Plot Tools

This toolbar allows the user to choose from various plots. You can have one of each plot type per Region of Interest (ROI). Each of these plots can be free-floating or they can be “docked”. The available plot types include:



Icon	Description
1	Statistics viewer
2	Profile Plot
3	Temporal Plot
4	Histogram Plot
5	O-scope Plot
6	Tool Description





4.7.1.1 Statistics Viewer

The Statistics Viewer allows the user to see a variety of statistic for all the ROIs currently drawn. It is divided into three main areas.



1. Stats table toolbar
2. Stats table
3. Measurement functions

4.7.1.1.1 Stats Toolbar

	Pause the update of the stats. (While the image keeps updating). Click again to restart the update.
	Save the current stats to a text file
	Turn on/off the Image ROI. Creates a column in the table with stats for the whole image.
	ROI Subtraction. Opens a dialog that allows the user to create an new column in the stats table that is a subtraction of two current ROIs.

4.7.1.1.2 Stats Table

There will be one column in the table for each active ROI. The top line of each column shows the ROI name and its color. A "*" next to a temperature value indicated that the value has exceeded the limit of the calibration range but is still below the hard clip limit. This data may not be within the normally specified accuracy. A "<" or ">" next to a temperature value indicates that the value has hit the hard clipping limit and should not be considered accurate. Statistics for areas and lengths will be "N/A" unless the distance checkbox is enabled.

Statistic [units]	Line 1
Mean [°C]	17.2
Std. Dev. [°C]	0.0
Center [°C]	(150.5, 329.0) 17.2
Maximum [°C]	(156, 325) 17.2
Minimum [°C]	(151, 329) 17.1
Number of Pixels	12
Single Pixel Area [cm ²]	N/A
Area [cm ²]	N/A
Length [cm]	N/A
<input type="checkbox"/> Emissivity	1
<input type="checkbox"/> Distance [m]	2


ROI attributes can be changed by right clicking on the ROI or its column in the table and choosing "Properties" from the menu.

The ROI Properties dialog box contains the following fields and controls:








- Name:** Line 1
- Name Position:** Top (dropdown menu)
- Draw Name:**
- Color:** Blue (dropdown menu)
- Emissivity:** 1 (text input) with a "calc" button
- Distance:** 1 (text input)
- Buttons:** Cancel and Ok

NOTE: Emissivity must be a value between 0 and 1. If you enter a value outside this range the box will turn yellow to indicate an invalid entry.

4.7.1.1.3 Measurement Functions [Max]




The righthand side of the stats window is used to define measurement functions. This tool allows the user to define mathematical expressions using ROI statistics. Any number can be defined. Once a function is defined you can view the expression by mousing over the function name. Measurement functions are managed using the toolbar at the top of the pane.

	Add Function	Opens the function definition dialog. User can create a mathematical expression using constants, object parameters, ROI stats, and other measurement functions. An arbitrary name can be assigned to the function and will be displayed in the table.
	Edit Function	Opens the function definition dialog with the current expression. Edit the expression and click OK to save.
	Delete Function	Deletes the highlighted function.
	Delete All Functions	Clears the Measurement Function table
	Load Functions	User can load a previous set of function from disk
	Save Functions	User can save a set of functions for later use
	Temporal Plot	User can create a temporal plot of a defined function.

4.7.1.1.4 Adding and Editing Measurement functions


Choosing to add or edit a function will bring up the Measurement Function Editor.




The left area (1) allows the user to choose an input variable. Possible choices include existing ROIs, data from the camera image header, or even other measurement functions. Click the arrows to expand the lists. The top area (2) allows the user to set the name and reference color for the function. Area (3) shows the complete expression as a “formula”. These formulas can be a combination of inputs and math functions from the “calculator” area. Boolean functions (True, False, etc) can be used to evaluate the function and this state can be used to trigger the start of data recoding. (See Recording Conditions, Sec 4.7.3.5)

4.7.1.2 Profile Plot [Max]

The Profile Plot creates a linear graph of the data along an ROI. If the ROI is a line (straight or bendable), this will be a traditional linear profile, where the actual data values (in the currently selected units) are plotted on an X-Y graph. If the ROI is an area (box, circle, polygon, freeform, or isotherm) then plot will be a “thick profile”. Think of a thick profile as drawing a line across an ROI and then computing a statistic (Mean, Max, etc). That will be one data point in the thick profile. You then scan the line across the ROI one pixel at a time. This will produce a series of data points. As shown in the picture below, the thick profile can horizontal or vertical.




If you open the plot tools by clicking the “<<” icon in the upper right corner of the plot, and selecting the legend, you can choose the direction and statistic for the thick profile .



4.7.1.3 Temporal Plot


The Temporal plots allows the user to plot an ROI statistic versus time

As with other plots types, the plot tools can be displayed by clicking the << icon in the upper right of the plot window. The common tools are described in Sec 4.7.1.8. The “Plot vs:” option is used to select the units for the X-axis. The width field is used to set the plot time span. This can also be done with the slider bar below the plot. If the span is less than the total data, the dark gray part of slider can be moved to display earlier data while keeping the same span.




4.7.1.4 Histogram Plot [Max]

The Histogram Plot shows the user the distribution of values within an ROI. The common plot tools are described in Sec 4.7.1.8. The Classes field is used to set the number of “bins” that the data is group into.




4.7.1.5 Oscilloscope Plot [Max]

The Oscilloscope Plot (O-scope) gives the user an “edge-on” view of the image. The O-scope can be thought of as a line profile with a profile for all the rows overlaid on top of each other. This allows the user to quickly identify regions of the image that are saturated. This plot can be very useful when setting up a camera for a user calibration to make sure that the integration time is set correctly.




4.7.1.6 Opening a plot


1. Select an ROI by clicking on it. The ROI color will change slightly and sizing handles will appear to indicate that the ROI is selected.




2. If you want to open the plot to its default location, just click the plot button without dragging.




3. If you want to open the plot and drag into a specific dock location, click and drag the plot button into a dock location. The plot will be displayed as a translucent preview.



4. When a plot can be docked into a dock location the dock location will highlight blue.




5. Release the left mouse button to dock the plot.



NOTE: Plots can be docked to the right and bottom of the initial ResearchIR layout. When a plot is docked, it creates a plot container. New plots can be docked to the top, left, bottom, right, and inside plot containers. In addition, there are two options for docking a plot inside a plot container. As shown in the figure below, selecting the left-

center docking handle creates a plot in a new tab. Selecting the right-center dock handle adds a new graph to the existing plot (called a multi-plot).




4.7.1.7 Managing docked plots

Docked plots can be closed or undocked by clicking the “x” on the plot tab or by dragging the tab. Undocking a plot turns it into a floating window that can be docked again by dragging its title bar into a dock location. The currently displayed plot in a plot container will have a darker colored tab.














4.7.1.8 Plot Toolbar

When the mouse cursor is hovering over the plot, the plot toolbar is activated and can be seen in the upper right corner of the plot window.









Clicking the “<<” opens the toolbar with all the available plot tools. Mousing over a tool will display a tooltip with the tool name. Not all tools are available for all plot types

The plot tools include:

Icon	Tool	Description
	Reset	Reset plot option to default values
	Lock Scales	Do not auto scale. This option may be activated in conjunction with other tools such as zoom.
	Pan	Use mouse to move plot
	Horizontal Box Zoom	Use mouse to click and drag horizontal. Captures full y-axis.
	Vertical Box Zoom	Use mouse to click and drag vertical. Captures full x-axis.
	Arbitrary Box Zoom	Use mouse to draw an arbitrary box
	Horizontal Magnify	Only zoom x-axis. Left click to zoom in. Right click to zoom out.
	Vertical Magnify	Only zoom y-axis. Left click to zoom in. Right click to zoom out.
	Arbitrary Magnify	Zoom both x and y-axis. Left click to zoom in. Right click to zoom out.
	Cursor select On/Off	When enabled (blue), user can set cursor location.
	Print	Send plot capture to printer device
	Save (BMP or CSV)	Save plot data to disk
	Clear all plots	Clear plots (temporal plot only)

4.7.1.8.1 Plot Cursors [Max]




Clicking on the  button in the toolbar activates the plot cursor. Use the mouse to point to the desired cursor location and click. This will place an “X” on all displayed plots. Multiple cursors will all share the same x-axis location. When mousing over the plot, a  button will also appear in the upper left corner. Clicking this button activates the cursor readout window. If you don't want the window to block the upper left portion of the plot, then click the  button to “pin” the display to the left of the plot. The   buttons, will move the cursor to the next/previous data point. With the mouse, you can point to any position on the plot. If there is no actual data point you will see an interpolated value.



4.7.1.8.2 Plot Legend


Under the toolbar is the interactive legend. The legend can be used to select the y-axis scale or delete a graph from a plot. To show a mini (non-interactive) legend when the tool bar is closed, right click on the plot and choose “Properties” from the menu. On the Legend tab there is an option to draw the legend.



Icon	Description
 Box 1	Name and color of ROI
	Clicking this icon switches the Y1 and Y2 axis
	Deletes the ROI from the plot


4.7.1.9 Plot Properties

Right-clicking on any plot window will bring up the plot properties dialog. This dialog allows the user to customize the plot appearance.




4.7.2 Bad Pixel Tool [Max]

The Bad Pixel Tool allows the user to manually add bad pixels to the Bad Pixel Map (BPM) that was created when doing a PC-Side Non-Uniformity Correction (PC-NUC). Place the cursor over the bad pixel and click the add/remove button (or the Spacebar) to mark the pixel as bad. Pressing the button again will unmark the pixel. Once the pixel has been marked it will turn blue. Clicking the Bad Pixel tool again (see cursor in picture below) will turn off the tool and all blue pixels will be replaced with a nearest good pixel.




1. Enable bad pixel selection cursor
2. Add/remove selected pixel from map





The program will search for a good replacement pixel using the pattern indicated below. It will start with the pixel to the upper left of the bad pixel. If that pixel is bad or invalid (pixel is in a corner) the algorithm will search other adjacent pixels. If none of those pixels are good, then the algorithm will expand the search ring by one pixel.



4.7.3 Acquisition Tools


The Acquisition Tools allow the user to start and stop the live display, record single images or movies and set recording options.



Icon	Description
	Play/pause live image stream
	Record a single image
	Record a movie
	Edit record settings

4.7.3.1 Record Settings

Selecting the “Edit Recording Settings” Tool displays the following dialog. This dialog allows the user to choose the Record Mode (To Memory or To Disk) as well as Record Options and the location to store recorded data. Each section of this dialog is explained in more detail in the sections that follow.



4.7.3.2 Record Mode


Allows user to choose option for recording to memory or disk

4.7.3.2.1 Record to Memory Mode [Max]

Recording to PC memory is the fastest recording mode, but the time is limited by available physical RAM (does not use virtual RAM). When recording to PC memory the user can set either a specific number of frames to record or a time to record, in minutes or seconds.

When specifying time it is still necessary to specify a max frame limit. The frame limit can be much higher than the number of frames needed to record the specified time.

The thing to keep in mind for all recordings to memory is that the frame limit is used to allocate a RAM buffer. RAM allocated to this buffer will not be available to other applications running on the PC. If the number specified is more than can be from continuous PC RAM, the acquisition will not start properly. You will see the acquisition status dialog but the acquisition will terminate with zero frames acquired. If this happens, lower number of frames to record.




4.7.3.2 Record to Disk Mode

Recording to disk can accommodate longer recording times, but at a slower rate. In addition to setting a number of frames or length of time for recording, the user can stop or start the recording with the F5 key or use the PC clock to start or stop the acquisition.




4.7.3.2.3 Periodic Recording

Periodic recording involves setting a pattern for recording frames. The user also specifies how the pattern should be repeated, when it should stop, and how the data should be saved. This mode is currently only available to recording to PC memory.



4.7.3.3 Record Options

These additional options may be used with any recording mode. The limit record rate is typically used with uncooled cameras because this type of camera may not have a frame rate control. ResearchIR tries to achieve the selected rate by dropping frames.



4.7.3.4 Pre-Trigger Recording Options [Max]


User can setup a circular buffer where data is continually captured. When a recording is initiated the frames in the buffer allow the user to see data before the recording was started.



The user can set the buffer size (a tool tip will show the max value for a given computer). The rate limiter options allow the user to slow down the rate of data recorded prior to the trigger. This is useful for extending the time span of the pre-trigger buffer.

4.7.3.5 Record Conditions

The Record Conditions allow the user to choose additional options for starting and stopping a recording. Some camera models have an external trigger input and can set a bit in the image header to indicate a trigger was detected. ResearchIR can use this image header information to start a recording. It is also possible to use a measurement function with a Boolean result to trigger a recording, or to use the PC clock to trigger a recording.




The “Stop Recording” option is only available when recording to disk and using the “Start/Stop” recording mode.

4.7.3.6 File Naming Options


ResearchIR provides the user with a number of options for automatically generating filenames. A separate set of options can be specified for both movies and snapshots. Options include:




- **Prefix:** This is an alphanumeric code that is placed at the beginning of the filename. These codes can be used for anything but are often used to refer to a specific camera. The table of prefixes can be edited by the user and are not restricted to two characters.
- **Text:** This is an arbitrary text string
- **Count:** This is a numeric counter that is automatically incremented for each acquisition. The starting number and increment can be specified.
- **Timestamp:** This will place the timestamp from the first image in the file name. The time will be extracted from the image header on the first frame. If the camera does not support time in the header then PC clock time will be used.




4.7.4 Preset Selector [Max]


For cameras that support Preset Sequencing (also called Superframing or Multi-IT), the Preset Selector control allow the user to choose which preset to display. This feature is currently only available in SCx000 and RSx000 cameras.



Control	Description
	Tells ResearchIR to display all active presets in sequence. For general display purposes this mode is not very helpful because it can be quite “flashy” as the AGC adjusts from frame to frame. This mode is useful if you are trying to do a PC-side NUC with multiple presets. With this mode ResearchIR will NUC all active presets at the same time, using the same NUC scenes. Depending on the integration times being used this may or may not produce optimal results.
	Tells ResearchIR to filter out a particular preset for display. If a chosen preset is not active in the camera, ResearchIR will display a message “Frame Not Available” in the image window.
	Enables real-time Dynamic Range Extension (DRX). If a camera is calibrated (factory or user), with a different temperature range loaded in each preset, this option will apply the DRX algorithm. Using Preset Sequencing, DRX will take the best pixel data from each preset and combine the data to form a new image that spans all of the available calibration ranges. This allows the user to span a much larger dynamic range than could typically be covered with one integration time. DRX works best for static scenes.



4.8 Toolboxes [Max]



Toolboxes are only available in ResearchIR Max. Toolboxes allow users to access advanced program functions. Some toolboxes are available only if a certain add-on module is active, such as User Calibration, Export, or Dynamic Range Extension. Other toolboxes are only available for certain camera types. For example, the Object Parameters Toolbox is only available if a camera has a factory calibration.

The toolbox view can be customized to show only what the user wants to see. The program will remember the last settings the next time the program is started.

4.8.1 Object Parameters Toolbox

Object Parameters are used to more accurately compute the temperature of an object. The object parameters toolbox displays the current object parameters as downloaded from the camera and allows you to edit them. Editing the values in the toolbox overrides the values coming from the camera but does not change the values for these parameters in the camera. (NOTE: For ResearchIR Basic users, this toolbox is available under the Tools menu.)

4.8.2 Camera Control Toolbox

This toolbox gives the user a mini version of the camera controller where the common functions like temperature range can be selected without having to bring up the full camera controller.

4.8.3 Source Information Toolbox

The Source Information toolbox displays information about the camera or file being viewed. This information differs from the metadata viewer in that this information is not frame dependent. Source information includes data like camera type, lens used, filters, calibration ranges, GPS information, frame rate, integration time, spatial calibration data, etc. The information shown will vary depending on the data source.

4.8.4 Minimap Toolbox


The Minimap allows the user to see a thumbnail version of the current image. If the image is digitally zoomed then a box appears on the minimap to indicate which part of the image is currently displayed in the main image window. The zoom box on the minimap can be moved with the mouse to easily position it over the area of interest.

4.8.5 Segmentation Toolbox [Max]

The Segmentation toolbox controls if and how segmentation is applied to the image data. Segmentation defines a range of values that are considered valid in the image. For instance, if the segmentation min and max are 7000 counts and 9000 counts respectively then only the pixels in the image that have a value between 7000 and 9000 are considered valid. All other pixels are segmented out (ignored). Pixels that are segmented out are not included when computing statistics. The Number of Pixels statistic will reflect the number of valid pixels in the ROI. Pixels below the segmentation minimum are show as blue and pixels above the segmentation maximum are show as red. The segmentation range can be defined in terms of counts, radiance, or temperature units.

4.8.6 Filters Toolbox [Max]

The filters toolbox allows the user to apply a variety of operators to the entire image. Click “Add Filter” will display dialog where the user can choose available filters from a list. Multiple filters can be chosen and they will be applied sequentially.



Icon	Description
	Enable or disable a single filter or all filters
	Change the order that filters are applied
	Set filter parameters
	Delete filter

Filter Name	Operates on	Description
Gain	Pixel	Multiply each pixel by the gain value
Offset	Pixel	Add the offset value to each pixel
Exponential	Pixel	Compute $\exp(\text{pixel value})$
Natural Logarithm	Pixel	Compute $\ln(\text{pixel value})$
Square root	Pixel	Compute square root (pixel value)
Gaussian	Pixel	Compute a Gaussian blur (smoothing) to the image
Window Average	Pixel	Make each pixel the average of the selected kernel
MATLAB	Image	Passes an image to a MATLAB script. See Sec 4.8.6.1 for more details.
Median	Pixel	Make each pixel the median of the selected kernel
Min – Continuous	Pixel	Make each pixel the temporal minimum until reset
Min – Frame interval	Pixel	Make each pixel the temporal minimum over the last n frames
Max – Continuous	Pixel	Make each pixel the temporal maximum until reset

Max – Frame interval	Pixel	Make each pixel the temporal maximum over the last n frames
Frame Average	Image	Make current image the average of the last n frames. Relative mode shows the actual delta values. If this output will be fed into another filter that cannot accept negative numbers, the absolute mode will add the min value of the image in order to make all pixels >0
Sliding subtraction	Image	Subtract the previous nth frame from the current frame. Relative mode shows the actual delta values. If this output will be fed into another filter that cannot accept negative numbers, the absolute mode will add the min value of the image in order to make all pixels >0
HSM Mode	Image	Emulates the HSM mode found in GF-series cameras

4.8.6.1 MATLAB Filter

This filter allows a user to pass an image to an external MATLAB script for further processing. This filter will only be available if MATLAB is installed (not included with ResearchIR). The data that is passed will already have NUCs and calibrations applied (if enabled) and will be in the units selected in the upper right corner of the ResearchIR window. Below is an example of how to make a MATLAB script that is useable by the filter.

```
h = size(rir_filter_input, 1);
w = size(rir_filter_input, 2);

if rir_filter_reset
    rir_filter_output = zeros(h, w, class(rir_filter_input));
end

for y = 1:h
    for x = 1:w
        rir_filter_output(y,x) = max(rir_filter_output(y,x), rir_filter_input(y,x));
    end
end
```

`rir_filter_input` is a matrix with the input image

`rir_filter_reset` is a boolean scalar that is set to true when the image size changes or the user disables and enables the filter or the user clicks reset

`rir_filter_output` is the output image matrix, the filter is expected to create this matrix

the input will be 16-bit unsigned integer or 32-bit floating point depending on the unit selected in RIR

the output is also expected to be 16-bit unsigned integer or 32-bit floating point, if you need to you can promote from a 16-bit unsigned integer to a 32-bit floating point in the filter.

4.8.6.2 Passing metadata between ResearchIR and MATLAB

Most cameras have some metadata embedded in the image stream. The specific information will depend on camera model but it will typically include data like time stamp, and some camera settings. This data can be viewed in the metadata viewer under the image window. It is possible to pass these parameters to MATLAB along with the image. If MATLAB is being used to calculate a numeric result, rather than an image, it is also possible to pass this result back to ResearchIR and see it in the metadata viewer. Data in the metadata viewer can be plotted versus time using a temporal plot. Below is an example of a script that passes the average, min, and max values of an image back to ResearchIR.

```
rir_filter_metadata_output.Average = mean(rir_filter_input,'double');
rir_filter_metadata_output.Max = max(max(rir_filter_input));
rir_filter_metadata_output.Min = min(min(rir_filter_input));
```

4.8.7 HSDR Toolbox [Max+]

This toolbox will only be available if you have a High Speed Data Recorder (HSDR). ResearchIR can work with CL-160, CL-SAS, pHSDR versions of HSDR.


4.8.7.1 First-time Setup

Before the HSDR can be used with ResearchIR, the HSDR core software drivers must be installed. If one is using the HSDR Core2 hardware, the host PC must have a USB 3.0 port on it for the eSATA to USB 3.0 converter cable. If you are installing ResearchIR on an existing CL-160 or CL-SAS HSDR (with RTools), this software should already be installed, and you can proceed to Sec 4.8.7.2. If you have a pHSDR, this software can be found on a CD that came with the unit. Just run the “core.exe” program. Reboot the PC.


After installing the drivers, go to the “Tools” menu and select “HSDR Setup...” You will see the dialog box shown below. Select the type of HSDR you have and check the “Enable HSDR” box



NOTE: For pHSDR systems, you must connect the HSDR unit to the PC using the USB 3.0 to eSATA cable BEFORE launching ResearchIR, otherwise the software will not detect the HSDR. Refer to the HSDR User Guide for more information and troubleshooting help. It is FLIR document number 4212427.



Once this step has been completed, you will see the HSDR Toolbox and File / HSDR tabs in the File Toolbox.



4.8.7.2 Pairing


Each time a new camera is connected to an HSDR, the two devices must be “paired”. Connect to camera using the GigE interface. Once you have a live image, click the




button to pair the camera to the HSDR. Once complete, the “Not Paired” indicator will be replaced with the Camera information (Model # / SN)


4.8.7.3 HSDR Triggered recording (DVR express Core only)

It is possible to trigger the start of a recording on the HSDR providing an external signal to the AUX connector on the rear of the HSDR. A TTL signal can be applied to pin 18 (pin 26 for ground).



In ResearchIR, there is a  button that will “arm” the recording and set the program to look for the HSDR trigger signal.

4.8.7.4 Movie Recording

Use the  button in the HSDR Toolbox to set recording option for the HSDR. These options are independent of the options set for non-HSDR recording (see Sec 4.7.3.3).



The Record Mode options are similar to those available for non-HSDR recording. There

is an additional “Start/Stop” mode. Here you can use the  and  buttons to start and stop the recording when desired. After a recording is completed, a message will be displayed in the toolbox indicating the number of frames recorded, frame rate, recording time, and number of dropped frames (should be zero).



File name options are similar to those described in Sec. 4.7.3.6.

4.8.7.5 Movie Playback


The HSDR tab in the File Toolbox allows the user to access data that has been recorded to the HSDR but has not been extracted to the PC hard drive. Just double-click the movie you wish to play.

4.8.7.6 Movie Extraction

Use the Extract feature (Sec 4.9.1.2) to copy files from the HSDR drives to the regular PC hard drives. Once extracted, the movie will be displayed in the File tab of the File Toolbox. It is IMPORTANT that you extract data you want to keep as soon as possible. Certain actions such as pairing a new camera or changing the camera window size can trigger a reformat of the HSDR drives, possibly causing this data to be permanently lost.

4.8.8 File Toolbox

The File toolbox displays files in the current directory that ResearchIR can open.



Icon	Description
	Current displayed directory
	Set current directory
	Set current directory to record directory.
	Set current directory to export directory

In addition to clicking the open button, you can double click on a file to open it.

NOTE: The “Show Record Folder” and “Show Export Folder” buttons only change the folder being viewed. It does not set the location for recording and exporting. To set the recording location use the “Edit Recording Settings” button described in Sec. 4.7.3.


4.9 Main Menu Bar

File Edit Camera View Tools Help

This set of menus allow the user to access functions pertaining to camera connectivity, display parameters and plot tools.


4.9.1 File Menu

	Menu Option	Description
	Open	Open a file
	Close	Close the current file
	Open Recent	Shows a list of recently used files
	Save Bad Pixel Map	Updates PC-side NUC bad pixel map
	Export	Shows options for exporting data
	Extract	Allows the user to save a portion of the current file

	Batch Extract	Extract data from multiple files
	Page Setup	Options for setting up page for printing
	Print Main Screen	Allows user to print the current screen
	Exit	Exits the program

4.9.1.1 Export Dialog

The Export Dialog allows the user to export images and data from ResearchIR for use in reports or other programs. Data values can be exported to CSV files or images and movies can be exported to a number of standard formats. ResearchIR provides a number of options to customize the exported objects.



There are three basic export types:

- **Current Image:** Exports only the currently displayed image frame
- **Movie:** Exports the selected range of frames as a video
- **Multiple Images:** Exports the selected range of frames as a series of individual files.

Each type has a number of format options. (Only CSV, BMP, JPG, PNG, and WMV are available in Standard)

Export Type	Export Format	Description
Current Image	Comma Separated Variable (CSV)	Export ASCII text file that contains the data for each image pixel. Can be imported to other programs like Excel.
	Windows Bitmap (BMP)	Export Image capture stored in BMP format
	Portable Network Graphics (PNG)	Export Image capture stored in PNG format
	JPEG	Export Image capture stored in JPG format
	TIFF	Standard 24-bit TIFF (8-bit RGB)
	TIFF (16-bit counts)	Monochrome. Each pixel value stored in counts. Can be imported to Matlab
	TIFF (32-bit floating point)	Monochrome. Used for storing calibrated data. Can be imported to Matlab.
	FITS	Export binary data file in FITS format. This is an open file format.
Movie	SAF Image	Export binary data file in SAF file format. This is an open file format.
	Windows Media Movie (WMV)	Export a movie file in WMV format
	FITS	Export binary data file in FITS format. This is an open file format.
	SAF	Export binary data file in SAF file format. This is an open file format.
	TIFF (32-bit floating point)	Monochrome. Used for storing calibrated data. Can be imported to Matlab.
	YUV	This is a near lossless format that can be read by many high-end video editing packages. If you

Export Type	Export Format	Description
		need to encode you data in a specific codec, you can export in YUV and then use the video editing software to encode in the codec of your choice
Multiple Images	Comma Separated Variable (CSV)	Export ASCII text file that contains the data for each image pixel. Can be imported to other programs like Excel.
	Windows Bitmap (BMP)	Export Image capture stored in BMP format
	Portable Network Graphics (PNG)	Export Image capture stored in PNG format
	JPEG	Export Image capture stored in JPG format
	TIFF	RGB, 16-bit counts, or 32-bit floating point
	FITS	Export binary data file in FITS format. This is an open file format.
	SAF	Export binary data file in SAF file format. This is an open file format.

The **JPEG Quality** control adjusts the amount of compression. The further the slider is moved to the right, the better the image quality (larger the file size).

The **Rate (Hz)** control allows the user to set the playback rate. This is useful when the source data rate is much slower or much fast than normal.

The **Output** field allows the user to designate the location for exported files to be stored. There are also a number of additional export options that can be enables by checking the boxes. Options that do not apply to the chosen export settings are greyed out.

- **Export Entire Image:** If unchecked, the image is exported using the current size (as displayed). This could be a zoom factor less than 1x or greater than 1x, depending on the window layout and monitor resolution. If this option is checked then the exported image will be full size (as if you were set to a 1x zoom factor.)
- **Include Image Border:** Adds a border to the image.

- **Include color bar:** Adds the color bar and scale to the exported image.
- **Show ROIs:** Include the ROIs in the exported image.
- **Show ROI names:** Include the ROI name in the exported image.
- **Show ROIs:** Include the ROIs in the exported image.
- **Show ROI crosshairs:** Include the ROI center marker in the exported image.
- **Show min/max crosshairs:** Includes the markers showing the min/max pixels in the image.
- **Show Classification:** This option allows the user to add labels to the top and bottom of the image. These labels can be used for any purpose in addition to classification markings.
- **Show Timestamp:** This option adds the image timestamp to the exported video. The time can be appended to the bottom of the image, overlaid on top of the image, near the bottom of the frame, or merged with the classification text labels.
- **Export ROI bitmasks:** In addition to the image, additional files are exported with the ROI bitmasks. The files have the same base name with the ROI name appended. These are stored in the location specified in the Output field.
- **Export stats:** In addition to the image, an additional text file is exported with the ROI statistics.
- **Display images while exporting:** When checked, ResearchIR will display the frame being exported.
- **Include header when exporting to CSV.** Adds a block of header data to the beginning of the CSV file.
- **Use frame skipping:** When checked the user can specify a pattern of frames to keep and skip.


File Type	Description
CSV (Comma Separated Variable)	<ul style="list-style-type: none"> • Imager data. • Can be read into excel, but excel has some row/column count limitations. • Is an easy format to read if you want to look at the raw data. • This format does not describe what units the data is in.
BMP (Windows Bitmap)	<ul style="list-style-type: none"> • Uncompressed rendered image. • Common format for images, using this over TIFF is a matter of preference.
PNG (Portable Network Graphic)	<ul style="list-style-type: none"> • Lossless compressed rendered image. • Smaller file size than TIFF/BMP, but without any loss of information.
JPEG	<ul style="list-style-type: none"> • Lossy compressed rendered image.

	<ul style="list-style-type: none"> • Usually has the smallest file size, but still looks relatively good. • Because it's a lossy compression method, the resulting image is not a perfect representation of the rendered frame.
TIFF	<ul style="list-style-type: none"> • Uncompressed rendered image. • Common format for images, using this over BMP is a matter of preference.
TIFF (16-bit)	<ul style="list-style-type: none"> • Imager data. • Only available in counts mode. • Truncates data to an integer. • Outputs a 16-bit grayscale TIFF file with each pixel being a 16-bit counts value • This format does not describe what units the data is in. • Many image editing applications can read 16-bit TIFFs.
TIFF (32-bit)	<ul style="list-style-type: none"> • Imager data. • Doesn't suffer from data truncation like a 16-bit TIFF. • This format is standard, but much less common than 16-bit TIFF. • If your a fan of TIFF and want to get reduced data then this is a good format, otherwise use SAF. • This format does not describe what units the data is in.
FPF (FLIR Public Format)	<ul style="list-style-type: none"> • Imager data. • Only available for temperature / object signal. • Only factory calibrated cameras can produce complete FPF files. • Currently broken, currently only available for factory calibrated cameras. • This is a good format for factory calibrated cameras in temperature or object signal mode. • The format does have a units field.
FITS	<ul style="list-style-type: none"> • Imager data. • Does not truncate data (can be floating point or integer). • Format is standard and fairly easy to read. • Can be used for movies. • This format does not describe what units the data is in.

	<ul style="list-style-type: none"> This is a good choice if you have existing software that uses FITS files, otherwise use SAF.
SAF (Standard Archive Format)	<ul style="list-style-type: none"> Imager data. Does not truncate data (can be floating point or integer). Format is standard and fairly easy to read. Can be used for movies. The format does have a units field. Compatible with RTools. This is the recommended format for imager data.
WMV (Windows Media Video)	<ul style="list-style-type: none"> Lossy compressed rendered image. Only rendered movie format available. Can be viewed from almost any windows computer. Produces relatively small files with relatively good quality.

4.9.1.2 Extract Dialog [Max+]


The Extract feature allows the user to save a portion of a long video into another file to make a smaller file. If ResearchIR has recorded data to a High Speed Data Recorder (HSDR), the Extract dialog allows the user to move the data from the data recorder to the PC hard drive and to trim a file and save that portion to a new file.



When first opened this dialog displays the last folder used for extraction and the current filename. If you have been using the original data directory, you must change the location or filename as ResearchIR will not let you overwrite the original file.

4.9.1.3 Batch Extract Dialog [Max+]

If you have a lot of data that needs to be extracted from the HSDR, this can be a time consuming process. The batch extract dialog allows the user to setup a list of files to extract. Only one file can be specified per row in the table. (It is currently not possible to add a list of files from a list or folder in a single operation, but this will be added in a future version).



4.9.2 Edit Menu

	Menu Option	Description
	Spatial Calibration...	Enter data for computing lengths and areas of ROIs.
	Preferences	Set program preferences

4.9.2.1 Spatial Calibration [Max]

A spatial calibration allows ResearchIR to compute lengths and areas of ROIs drawn on an image. The Spatial Calibration dialog allows the user to enter the data necessary to compute the camera's Instantaneous Field of View (IFOV). IFOV is the field-of-view of a single pixel. ResearchIR supports independent values for horizontal and vertical IFOV; however, modern cameras have square pixels so these values will be the same.

The Spatial calibration dialog presents the user with four options for computing IFOV values. The results will be displayed in microradians.

Spatial Calibration

Calculate from FOV

FPA Width: FPA Height:

FOV: degrees ▾

Calculate from focal length

Pixel Pitch: microns

Focal Length: millimeters

Enter manually

IHFOV: IVFOV: milliradians ▾

Measure

Distance: meters ▾

Line: Length: meters ▾


IHFOV: IVFOV: (microradians)

OK Cancel

Spatial Calibration Method	Description
Calculate from FOV	Enter the FPA height and width in pixels and the Field-of-view (FOV) of the optics
Calculate from focal length	Enter the pixel pitch (size), and lens focal length
Manual	If you know the IFOV, just enter it manually
Measure	Draw a line ROI on an object of known length in the image and enter the distance to the object from the front of the lens.


4.9.2.2 Preferences Dialog

The Preferences dialog allows you to edit settings that affect how ResearchIR behaves.



Preference	Description
Metadata	Metadata are items that can be extracted from the frame header and can be shown in the metadata control on the main window. The available items will vary depending on the camera type. Multiple items can be selected by holding down the CTRL key while selecting with the left mouse button.
ROIs	Check to display ROI names and center crosshair Smooth ROIs: (Off) ROIs are pixel accurate when rendered. Allows user to see exactly which pixels are included. Useful when trying to compute accurate lengths, areas, and intensity (On) ROIs are not pixel accurate but rendering is not affected by native resolution of the image.
Level and Span	Attempt to decrease jitter or wobbling in the level and span control often caused by twinkling pixels. This is done by ignoring a few of the pixels on the top and bottom of the range of values. Enter the number of pixels at the top and bottom of the range to ignore
Fast Forward Skip	Enter number of frames to skip in fast forward/reverse playback
Target Frame Rate	Sets a limit for displaying live data. This is not related to the camera frame rate. The default value is 16 (Hz). It is sometimes useful to lower this value when trying to log temporal plots for very long periods of time (many hours) because the plot buffer can get full
Settings for all Users	If unchecked, preferences will be per user
Draw Min/Max	Check to display a “+” or arrow at the location of the min /max value in each ROI
Loop Movie Playback	Check this to play movies in a continuous loop
Image Interpolation	For lower resolution cameras (320x240 and below), this option can make the displayed image look less pixelated
Enable ATS-US USB remote start or Serial-USB remote start	ResearchIR supports two variants of USB Remote Start boxes. The older FLIR proprietary box or you can use a standard USB to RS-232 converter (must provide your own button) for switch closure only. Short pins 7/8 (CTS/RTS)
Remote Start Controls	Remote Start can be set to trigger recording of Live (currently connected camera) or HSDR or both
Stop Recording	When Hotkey / remote start can stop recording
Theme	Two color themes are supported: Light or Dark
Reset	Sets all Preferences to default values

4.9.3 Camera Menu

	Menu Option	Description
	Connect	Connects to last selected camera
	Disconnect	Disconnect from camera
	Select	Select a camera to connect to from a list of available cameras
	Control	Show camera controller for current connected camera
	User Calibration	Functions for performing, saving, and loading user calibrations
	PC Side Correction	Function for performing, loading, and saving PC side NUCs

4.9.3.1 Camera Selector


ResearchIR has two methods of selecting a camera: Auto detection, and manual selection. If you select a camera with the Startup dialog (shown at startup or View->Startup Dialog menu), ResearchIR will try to auto detect the control interface. This is generally the recommended method. If you need to manually specify the connection interface you can use the Camera Selector.

The Camera Selector has two view modes: Standard and Custom. The Standard Mode automatically scans the computer and displays available cameras. The cameras are grouped by interface (Firewire or Ethernet). The Custom Mode allows you to manually choose the camera and connection interface



NOTE: You can run more than one instance of ResearchIR on the same PC. If one instance of ResearchIR is connected to a camera, then that camera will not show up in the Camera Selector in other instances of the program. Once the camera has been disconnected, it will show up as an available camera.


Camera Selector (Standard View)



Select the camera, but do not connect and show live image

select and immediately connect and show live image from camera

Camera Selector (Custom View)



4.9.3.2 Camera Control

ResearchIR provides a camera controller for all supported cameras. The controller is customized for each camera and provides access to all camera functions.


4.9.3.3 User Calibration [Max]

The user calibration menu manages user radiance/temperature calibrations .

Menu Option	Description
Load	Load a previous user cal from disk
Save	Save the current user cal to disk
Load Pack	Pack file contains both INC and CAL file
Save Pack	Pack file contains both INC and CAL file
Perform	Start user cal wizard
Edit	Make changes to an existing user calibration
Clear	Unload current user cal from memory

4.9.3.3.1 Load Calibration

Brings up a file dialog box and allows user to select a user calibration to be loaded into memory. The user calibration can then be applied to the data by choosing it with the Units Selector.



4.9.3.3.2 Save Calibration

Brings up a file dialog box and allows user to save the current user calibration to disk. This will create two files: a .INC and a .CAL. These two files are part of the SAF format, which is fully described in a separate document.


4.9.3.3.3 Perform Calibration

The Perform option starts the User Calibration Wizard. This wizard will walk the user through the steps required to calibrate a camera. You will need a calibrated blackbody source to complete this procedure. This section is only a discussion of how to input user calibration data into ResearchIR.

It is important to note that as a general rule, the User Calibration Wizard will generate a calibration in terms of apparent effective radiance. This means that the calibration is mapping the digital counts from the camera to what radiance the camera sees, not necessarily to radiance at the source. If you use only ideal response curves for spectral response and atmosphere, ResearchIR will compute theoretical source radiance based on Planck's function and map the camera counts to that. Therefore, when you include data for spectral response and atmospheric path you are trying to reduce the theoretical radiance of the calibration source to the energy that the camera actually sees.

4.9.3.3.3.1 Wizard Modes

The user will be presented with a dialog asking them to choose the Wizard Mode, from Simple to Expert. The more advanced the mode, the fewer question will be asked. Also, the process is less “step-by-step”, allowing the user to customize the process according to their needs. It is recommended that beginners use the Simple wizard until they are comfortable with the process.




Mode	Description
Simple	<ul style="list-style-type: none"> • For temperature only calibration • Assumes a Blackbody emissivity of 0.98. • Assumes a top hat spectral response based on camera type. • No MODTRAN input • No Additional Responses • No reflected radiance compensation • Must follow the wizard steps • No access to calibration coefficients
Standard	<ul style="list-style-type: none"> • Best option for users new to User Calibration • Appropriate for Radiance and Temperature calibrations • Allows input of Spectral Response and MODTRAN data • Allows input of Additional responses • User can input blackbody emissivity • Allows for compensation of reflected radiance • Walks user through Wizard, but then user can review/edit data without Wizard • User can review/modify calibration coefficients
Advanced	<ul style="list-style-type: none"> • Same inputs as Standard • Limited wizard steps
Expert	<ul style="list-style-type: none"> • No wizard • User has full control of calibration data entry

4.9.3.3.2 Standard Calibration Wizard


This section will outline the steps in the Standard Calibration Wizard. The other Wizard modes will be variations of this mode.

STEP 1: Getting Started. Before starting a user calibration, you will need to choose the desired integration time for the camera and you will need to perform a

Non-Uniformity Correction (NUC). The NUC can either be done on the camera side or on the PC side. Most FLIR cameras support on-camera NUCs. (The SC8000 is an example of a camera that does not.) PC-Side NUCs are discussed in the next section. See your camera's user's manual for more information on camera-side NUCs.



STEP 2: Input Camera Spectral Response. The user can choose to use either an ideal “top hat” response or provide a file with an actual response curve. The graph will show the response curve that will be used.




An actual response file should be a **peak-normalized power spectral response** (not a photon response). A response file is a simple tab-delimited ASCII file with the wavelength in microns and normalized response values. Below is an example.



STEP 3: Input Atmospheric Path. This screen provides a place to input Atmospheric path data from MODTRAN. MODTRAN is a widely accepted model


used to predict atmospheric transmission. The MODTRAN model has several output files. **ResearchIR is setup to read the TOTAL TRANSMISSION and PATH THERMAL data from the MODOUT2 files.**

Once the MODTRAN data has been loaded, ResearchIR can plot the data.



STEP 4: Input Additional Responses: Additional responses can be used to account for other factors that can affect the path between the cal target and the camera that are not already accounted for by camera spectral response or atmospheric modeling. Such things might be a mirror reflectance curve or an additional filter. The values in this file are multiplied against the data.

More than one response curve can be added using the  Add Response button. Unwanted responses can be deleted using the  Remove Selected Response button



The file format for an additional response file is a simple tab-delimited ASCII file. The first column should be the wavelength in microns and the second column should be the transmission (value from 0 to 1). The data increment does not have to be the same as the spectral response or atmospheric files. ResearchIR will automatically interpolate the values.

STEP 5: Set default blackbody emissivity and reflected radiance compensation options. This dialog allows the user to set the default emissivity value for the blackbody. If you are using more than one blackbody or the emissivity change with temperature, it is possible to override this value when you are taking the actual data.

Calibration Wizard

What is the emissivity of your blackbody? 21.0

Emissivity is the ratio of how much energy your blackbody emits to the amount of energy a blackbody theoretically emits. Emissivity ranges from 0 to 1 and blackbodies tend to have emissivities close to 1.

Emissivity: 0.98

Compensate for reflected radiance?

Reflected radiance is the ambient energy reflected by the blackbody. If the blackbody's emissivity is not 1 it will be reflective and ambient energy will be reflected off the blackbody and into the camera.

Do not compensate for reflected radiance **In general, choose this option.**

Compensate for a known amount of radiance

Reflected Radiance: 1.327067e-005

Compensate for radiance from the ambient temperature

Ambient Temperature: 23


Ambient Emissivity: 0.40

Computed Reflected Radiance: 1.327067e-005


<< Previous [Cancel] [Hide] Next >>

The user can also set options to try and compensate for reflected energy in the scene that is not coming from the calibration source. Typically, this extra source is reflected solar energy. In almost all cases, you should not attempt to compensate for reflected radiance unless you have no other option. If at all possible you should arrange your calibration setup to provide clean, uncorrupted view of the calibration target. This will give you the best and most repeatable calibration.

STEP 6: Setup calibration source. This dialog box prompts the user to setup the camera to see the blackbody. Once the blackbody is setup, click “Next”.




STEP 7: Draw an ROI on the blackbody. This dialog box prompts the user to draw a Region of Interest (ROI) on the blackbody emitter. The ROI should be drawn so that it is large enough to get as many good pixels as possible but should not be drawn too close to the edge of the source. ResearchIR will take the spatial average of all the pixels in the ROI so any pixels that are off of the blackbody will cause errors. Not all blackbody sources are created equal and uniformity at the edges is usually the worst so it is better to have fewer pixels that you know are good. Sometimes it is hard to tell from looking at the image if the edges of the blackbody are uniform. Using the O-scope plot is a good way to tell if all of the pixels in the ROI are uniform.




STEP 8: Ready to Start Calibrating. The user is now ready to start adding points to the calibration table. The calibration wizard uses a linear curve fit so at least two points are required. However, more points are better. This dialog gives instructions and helpful tips for collecting good calibration data. Keep these in mind:

1. A calibration must have at least two points.
2. The calibration should cover the entire range of temperature you want to measure. Extrapolation is not a good practice. A good calibration will be very linear. If your curve has nonlinear characteristics there is generally something wrong with the calibration data.


Click “Next” to continue.




STEP 9: Add calibration point. This dialog is used to collect and plot the calibration data. The settings chosen here only affect this calibration point.



1. Enter the temperature of the calibration target
2. Set the emissivity of the calibration target
3. Check the presets for which this calibration point will be acquired. By default, all active presets are checked.
4. The current ROI value for each preset is displayed here
5. Click the “Add” button to record the calibration point.

STEP 10: Review data and add more points. Once you have added a data point you will see a table summarizing all of the data taken. To add an additional point, click the  button. Below is a sample of a complete calibration. The summary table and graph allow the user to modify the data and assess the quality of the calibration. The top section shows information about each cal point. The check boxes can be used to remove a particular point from a preset. The bottom section shows a plot of the counts versus radiance so that the user can see if the calibration has the expected linearity.




STEP 11: Edit Misc Tab. The Misc tab allows the user to edit tags that will be part of the INC file. The top section has several commonly used tags. The bottom section allows the user to add any tag desired.

Commonly used and recommended tags are located at the top of the dialog box. Check the boxes to enable the tags. These include:

Tag	Description
TPFACT	This parameter is a Multiplier to the computed radiance. It is generally a number >1. It is most often used to compensate for the effects of atmospheric attenuation or an ND filter that was not part of the original calibration.
Slant Range (m)	This is the direct distance from the camera to the target. This parameter is used in conjunction with IFOV to compute the size and areas of objects in the image. Area is need to compute target intensity
IFOV	This is the field-of-view of a single pixel. If a Spatial Calibration (see Sec 4.9.2.1) has already been performed, the IFOV data will be imported to this dialog. The user cal also run the Spatial Calibration calculator by clicking the “Calculate” button.
Bandpass	This is an option tag and is superseded by the spectral response data. If a spectral response is loaded, these fields will be populated with the 50% points of the response curve.


Additional (optional) tags can be entered at the bottom of the form. These can be tags that are defined as part of the SAF file format or they can be user-defined. Placing them in the “Global” tab means they are assigned to all presets; or they can be assigned only to a single preset. All tags entered on this tab will become part of the default INC file. As long as this calibration is loaded this

default INC will be used as a template to generate the INC file for each SAF movie that is recorded.

Click “Finish” to close the calibration wizard. To save the calibration data for future use click the  button. To apply the calibration to the live data choose either a Radiance (user) or Temperature (user) from the Units Selector.

4.9.3.3.3 Review Calibration Coefficients

The “Coeffs” tab allows the user to review/edit the coefficients for computing radiance and temperature. The four ain sections are described in the table below.





Counts to Radiance	Lookup Table	Radiance to Temperature	Check
Order: 1	<input checked="" type="checkbox"/> Auto Start/End	Order: 6	Counts: <input type="text"/>
C0: -9.33486e-002	Start: 15.0	C0: -3.10147e+001	Radiance: <input type="text"/>
R ² : 1.0000	End: 40.0	C1: 6.85220e+005	Temperature: <input type="text"/>
<input type="checkbox"/> Override	Increment: 1.0	<input type="checkbox"/> Override	
<input type="button" value="Calculate"/>	Emissivity: 1.00	<input type="button" value="Calculate"/>	
C1: 8.90943e-006		C2: -4.06922e+009	
C2: <input type="text"/>		C3: 1.80198e+013	
C3: <input type="text"/>		C4: -5.03730e+016	
C4: <input type="text"/>		C5: 7.92702e+019	
		C6: -5.34635e+022	

Section	Description
Counts to Radiance	Displays the coeffs used to convert digital counts to radiance. Order one is typical but higher orders can be selected using the dropdown. Order “-1” or “-2” can be chosen if doing an “offset correction”. R2 is a figure of merit for the quality of the curve fit. A value > 0.9995 is typical for a good calibration.
Lookup table	Controls the generation of the lookup table that converts radiance to temperature. In “Auto” mode, the table will use the min and max temperatures used in the table at the top of the dialog box. If “Auto” is unchecked, the user can manually set the limits and increment. It is generally best to leave the Emissivity field set to 1.
Radiance to Temperature	Displays the coefficients used to convert radiance to temperature using a polynomial. This section can generally be ignored as this is a legacy method. The Lookup Table is the preferred method for computing temperature.

Section	Description
Check	Allows the user to compute a value of counts, radiance, or temperature using the coefficients. Enter any values and the other two will be computed. Temperature is computed using the lookup table.

4.9.3.3.4 Calibration Plots

By default, the calibration wizard will build a plot of counts vs. radiance. The user can build additional plots using the **Add Graph** and **Edit Graph** buttons. This will launch the Add/Edit Graph dialog.




This dialog (shown above), allows the user to select from a number of graph type. As the graph are selected using the “>>” button they will appear in the preview pane to the right. Once a graph is selected, other graph type that are not compatible with the the current axes will be disabled. To use one of these the user would need to create another new graph. One the graph is created in the preview the user can give it a name. In the box below the name the user can delete graph or change their order. By mousing over the preview you can access the plot toolbar with functions for pan/zoom and saving the plot to CSV or BMP.

4.9.3.4 PC-Side Correction (Max Only)


PC-Side NUCs are similar to the camera-side NUCs. Both provide good image correction but there are some differences in functionality. The following table compares the two types of NUCs.

NUC Feature	PC-Side NUC	Camera-side NUC
1-point correction (Compute offset, Gain =1)	✓	✓
2-point correction (Compute Gain and Offset)	✓	✓
Update offset only (keep current gain, compute new offset)	✓	✓
Bad pixel detection	✓	✓
Use factory bad pixel map (eliminates more bad pixels and twinklers)		✓
Can be applied to camera analog output		✓
Can use camera internal NUC flag		✓
NUC data stored separately from raw digital data (NUC data can be changed in post-processing)	✓	
Manual bad pixel tool	✓	
NUC storage space	unlimited	limited


 If desired, both types of NUCs can be used simultaneously. However, if you are using a factory calibration it is **STRONGLY** recommended that you not use a PC-Side NUC as this can affect the calibration accuracy.

One exception to this recommendation is using the Bad Pixel Tool to mark additional bad pixels that are not masked by the automatic bad pixel detection algorithm.

ResearchIR keeps track of the last PC-Side NUC done for each camera it connects to. If a PC-Side NUC is available, then the NUC icons will appear next to the Units Selector.




A blue shading over the icons indicates that the NUC and Bad Pixel Map (BPM) are enabled.




4.9.3.5 Creating a PC-Side NUC in ResearchIR

STEP 1: Choose the NUC Type. This dialog allows the user to select the NUC action to perform. Click Next to continue.




NUC Type	Description
One Point Correction	Compute Offset, set Gain=1. No automatic bad pixel detection.
Two Point Correction	Compute both Offset and Gain. Optional automatic bad pixel detection.
Update Offset	Keep current gain. Compute new offset.
Bad Pixel Detection	Keep current Gain and Offset. Compute new Bad Pixel map.
Default	Remove current PC-side NUC. Set Gain=1 and Offset=0.

STEP 2: Set NUC Options. This dialog allows the user to set the number of frames to average, and whether to detect bad pixels and twinkling pixels. The default values are good for most situations.




STEP 3: Collect NUC data. User will be prompted to fill the field of view with a uniform scene. One or two scenes (hot and cold) will be needed depending on the type of correction being performed. Once you have filled the field of view of the camera with a uniform scene, click “Next”. It does not matter which scene (hot or cold) is used first.




For the best results, the two scenes should be far enough apart in temperature to span as much of the A/D range as possible. If the real scene spans a large range of the A/D than the scenes used to generate the NUC then the NUC algorithm will attempt to extrapolate. This will generally lead to undesirable artifacts in the image.

NOTE: When choosing the hot and cold scenes for the NUC it is not always necessary or desirable to span the same temperature range as the scene. The NUC source is generally placed right up to the camera so in many cases a temperature that would not saturate the camera at a normal measurement distance may saturate the camera when NUCing. This is caused by the difference in atmospheric transmission. For this reason, when choosing scenes for a NUC it is better to think in terms of how many A/D counts the source produces rather than its temperature.


STEP 4: Accept or Discard Correction. A dialog will appear showing the number of bad pixels detected. Generally this will be a very low number. A high number usually indicates that one of the scenes was not uniform. If accepted, the new correction will be stored as the default PC-side NUC.




STEP 5: Enable the PC-Side NUC. Once you have accepted the NUC, you can apply it to the live image by using the NUC controls next to the Units Selector.




Below is the original scene before the NUC and with the NUC and bad pixel map applied.




4.9.4 View Menu

	Menu Option	Description
	Zoom	Select image zoom level
	Palette	Choose image color palette
	Preserve Aspect Ratio	Maintain image aspect ratio when resizing application window
	Fullscreen	Show image only in fullscreen mode
	Startup Dialog	Show the initial startup screen for selecting recent files and cameras

4.9.4.1 Zoom Menu


	Menu Option	Description
	Stretch	Stretch the image to fill the space available.
	1/4	Display the image at ¼ (25%) of its original size.
	1/2	Display the image at ½ (50%) of its original size.
	1x	Display the image at its original size.
	2x	Display the image at twice (200%) its original size.
	4x	Display the image at four times (400%) its original size.

4.9.4.2 Palette Menu

	Menu Option	Description
	Invert	Invert the colors in the current palette.
	Load	Load an external palette.
	Edit	Configure a given palette's settings
	1234, Blue, Glowbow, Grayscale, Green, Ironbow, Rainbow, Red, Sepia	Use the selected built-in palette.

4.9.4.3 Edit Palette Dialog

The Edit Palette dialog allows the user to customize and create palettes. You cannot permanently modify a default palette but you can save a new file. New palette files do not show up in the menu list but can be loaded using the Palette>>Load menu option.




Control Points tell ResearchIR where to start and stop gradients. You can set the color and location of as many control points as you like. ResearchIR will make a gradient between those points. Control points do not have to be evenly spaced.


You can also specify some special purpose colors for Segmentation and Clipping. Clipping colors will be applied when a pixel falls outside of a calibration range.

4.9.4.4 Startup Dialog

The Startup dialog allows you to quickly open a recent file or connect to an available camera. The startup dialog is shown when ResearchIR starts up by default. It can be set to not show when ResearchIR starts up with an option in the dialog.



4.9.5 Tools Menu




Menu Option	Description
Statistics Viewer	Toggles to show or hide the stats viewer.
Profile Plot	Shows a plot of the data values along a profile of the selected ROI. If ROI is a Line, the graph will be a profile along the line. If the ROI is an area, the graph will be a “thick” profile. Y-axis units correspond to units selector.
Temporal Plot	Toggles to show a time history plot for selected ROI. User can choose statistic to plot vs. time.
Histogram Plot	Toggles to show a histogram plot of selected ROI.
Oscilloscope Plot	Toggles to show an edge-on view of the entire image
Color Bar	Toggles to turn the color scale on or off
Flying Spot Meter	Toggles to show a flying cursor when the mouse cursor is over the image
File Operation	Shows a dialog that allows a user to specify a reference image that can be added, subtracted, multiplied or divided by all images
Make Self Viewing File	Shows dialog to create a Self Viewing File
Hotkeys...	Define Hotkeys
Setup HSDR	Initial Configuration of HSDR for use with ResearchIR
Export/Import HSDR Database	Allows user to save and reload the HSDR database when swapping media shuttles



Although the HSDR contains all of the recorded imagery, there is a database kept by ResearchIR on the PC that is an index of what movies have been recorded. If you wish to swap media shuttles on the HSDR is **EXTREMELY CRITICAL** that you use the Export function to save a copy of the data base before swapping the media. This database file must be kept with the SSD shuttle and imported once again when the shuttle is installed. The Database is reset when the “Setup HSDR” function is used. Once the database has been overwritten it is IMPOSSIBLE to recover the data on the SSD drives..

4.9.5.1 Color Bar

The Color Bar shows the relationship between colors in the selected color palette and the data values in the currently selected units.




The current segmentation levels are also displayed on the color bar as full width shaded regions in the currently selected segmentation colors.

Active Isotherm ROIs are displayed as half width shaded regions. Isotherm limits can be adjusted by dragging the ROI on the color bar. Isotherms ROIs can be deleted by clicking the isotherm on the color

bar and pressing the Delete key.


4.9.5.2 Flying Spot Meter



The Flying Spot Meter is a real-time data cursor that appears as the user runs the mouse cursor over the image. To the lower right of the cross-hairs is a tooltip that displays the cursor coordinate (in pixels) and the data value in the currently selected units.

4.9.5.3 File Operation

The File Operation dialog allows the user to designate an operation and a reference image. Once selected, the reference image is applied on a pixel basis to the current image using the operator. For example, this can be used to subtract one image from another and only show the differences. If the reference file is a movie, you can specify the frame within the movie to be used. Using the “Frame” option you can specify a frame from the current movie or from live data.



You can use the Extract feature (Sec. 4.9.1.2) to pull a frame from a movie and use as the reference or you can use the Snapshot record function (Sec. 4.7.3) to grab a single frame.

4.9.5.4 Make Self Viewing File [Max]

The Self Viewing File allows the user to share data with another person that does not have ResearchIR. Rather than just sharing a graphical format like JPEG, the SVF allow the secondary user to analyze the IR data with most of the features of the ResearchIR program. The main limitations of the SVF version of ResearchIR are:

- can only open the file to which it is linked
- max data size of 2GB
- cannot connect to cameras
- cannot export data


The Make Self Viewing File has two modes of operation: Single and Batch. The Single mode creates a single self viewing file from a single input (typically the currently open file). The Batch mode creates multiple self viewing files from multiple inputs.

4.9.6 Hotkeys

Hotkeys are keyboard shortcuts that can be configured by the user. A green check in the “en” field indicated the hotkey is enabled. Clicking the red x in the “rm” field will delete a hotkey assignment. To assign a hotkey, click on the command field to select the hotkey. Then click in the hotkey edit box at the bottom of the dialog box. Then press the key combination you wish to assign to this hotkey and click the “Assign” button.



4.9.7 Help Menu




Menu Option	Description
ResearchIR Help	Shows an online copy of the ResearchIR User's Guide.
About ResearchIR ...	Shows the program About Box. The About Box provides information about the program version.
Gather Information	This options is useful for troubleshooting when contacting FLIR Support. A text file is created that contains a variety of information that can be easily emailed to the support agent.

4.10 Metadata display

The metadata display allows the user to see data embedded in the image header. The data available is camera dependent but typically can include things like image timestamp, camera settings, and camera status flags.

Metadata			
Time	328:01:13:08.529478	Talo	00.000000
DeltaT	00.000000	FrameNumber	0
Preset	0	ActivePreset	0
FrameCounter	182582	FrameRate	60.00
HeaderID	20	IRIGLocked	F

The View → Preferences dialog allows the user to choose which header items are displayed in the table. The items available will vary, depending on the camera model.



4.10.1 Time in ResearchIR

The time displayed in the metadata can come from different sources, depending on the camera being used. Understanding the source time and the time reference are important when looking at the data at a future time

4.10.1.1 Time Source

Time can either come from the camera or the PC clock. The most accurate (and precise) time is camera time. If the camera inserts a timestamp in the image header, ResearchIR will use that time. If camera time is not available, ResearchIR will fall back on the PC clock. The following table shows what time source ResearchIR will use for a given camera.

NOTE: When viewing RGB, Visible, and Screen modes, PC time is always used.

Time Source	Camera
Camera Time	SC4000/SC6x00/SC8x00 (IRIG-B)
	SC5000/SC7000 (IRIG optional)
	SC620/640/660
	A40, S45, S65
PC Time (UTC)	A320/A325
	SC3xx
	T-series (T200-T400)
	i-series (i40, i60)

4.10.1.2 Time Reference

Time can be referenced to either a local time zone or UTC (Coordinated Universal Time – formerly known as Greenwich Mean Time (GMT))

When using camera time, ResearchIR has no way to know what the time zone reference is, so this is the responsibility of the user to keep track of. See your cameras User's Manual for details on setting the internal PC clock.

The SCx000 cameras generally offer an IRIG-B timestamp. IRIG is a precision time format developed by the US Government Test Ranges. The camera's IRIG clock can be set locally or it can be slaved to a master clock. IRIG does not have a time zone reference so the user will have to keep track of the time reference.

When ResearchIR falls back on using the PC clock for time, it will store the time referenced to UTC. In order to know the local time, the user will have to offset the time by the difference between local time and UTC. The table below gives some common time zones and their reference to UTC. These offsets are for Standard Time. Daylight Saving Time (or its international equivalent) would cause the offset to be one hour less (or more depending on the location).


Location	Time Offset from UTC	Daylight Time Offset
US Eastern Time (EST)	UTC-05	UTC-04
US Central Time (CST)	UTC-06	UTC-05
US Mountain Time (MST)	UTC-07	UTC-06
US Pacific Time (PST)	UTC-08	UTC-07
United Kingdom	UTC	UTC+01
France, Germany, Italy	UTC+01	UTC+02
Sydney, Australia	UTC+08	UTC+09

5 Camera^[taw4] Control


ResearchIR has a built-in camera controller for each of the supported cameras. Selecting the Camera→Control menu item will display the controller for the currently connected camera.

This section gives a summary of the controllers for some of the cameras supported. For detailed explanations of your cameras features please refer to your camera’s user manual.

For handheld cameras that have on-camera controls, the ResearchIR controller will have a basic configuration screen where you can set temperature range and update the NUC. You can also select the image mode to display the IR image, visible camera (if available) or the screen image. There is also a button interface where you roughly have the same button layout and labels as are on the camera. Using the button interface along with the screen display makes it possible to remotely navigate the on-camera menus.



Camera Controller




Button Interface

Below is an example of ResearchIR displaying the IR image, visible and screen image from an S65.

NOTE: Thermal analysis can only be done on the IR image. The visible and screen display are RGB data streams. The analysis tools will still work but you will be analyzing the RGB values. You can analyze the individual RGB channels or the average of the three.

Calibrated IR image data will be recorded in an SEQ file format. Visible and Screen images can also be recorded but they will be stored in a SAF RGB movie.



6 Other Utilities

In the Utilities folder there a number of external utilities provided. This section will explain their use and operation.

6.1 REdit

This utility has been carried over directly from RTools. It allows users to edit the headers of SAF files. This utility has its own online help which can be accessed from the “About...” menu.

6.2 CalibratIR

This is a standalone version of the User Calibration Wizard (see Sec 4.9.3.3.3.) With this utility a user can view/edit user calibration files with needing to run ResearchIR .

6.3 File AbatIR

The File AbatIR utility is used to apply the DRX algorithm to recorded data in post-processing. This is similar to the algorithm applied in Section 4.7.4. File AbatIR works with SAF, SEQ, and PTW files. The File AbatIR utility differs from the Realtime DRX feature in several key ways.

- Rather than computing the DRX data on-the-fly, File AbatIR generates a completely separate “ABMOV” file.
- The user can control the order in which the presets are processed and the count threshold at which the next preset is used.

The table below explains the various options and their uses. “Editable” output types allow the user to change “Object Parameters” (i.e. emissivity), even after processing. Also, “higher” unit types can be derived. For example a counts (editable) file be used to generate radiance or temperature data. These files can only be viewed in ResearchIR . Non-editable types have fixed values.

Output Type	Description
Counts (Editable)	Useful to “pre-process” data so that the real-time DRX feature can display data at a faster rate. Radiance and temperature units can be generated in ResearchIR
Radiance (User, Editable)	Output data is stored in radiance units but object parameters can still be changed in ResearchIR . Temperature units can be generated in ResearchIR
Radiance (User, Rescaled)	Data has been rescaled (normalized) to hide the absolute values.
Radiance (User)	Output stored in radiance units. Computations based on user cal data. (This is the traditional output produced by the “FPAbater” utility in

	RTools). Temperature units cannot be generated from this data.
Temperature (User)	Output stored in temperature units using user cal.
Radiance (Factory)	Output stored in radiance units. Computations based on factory cal data. Temperature units cannot be generated from this data.
Temperature (Factory)	Output stored in temperature units. Computations based on factory cal data. Radiance units cannot be generated from this data.

6.4 CNUC Manager

If you have an SC5000 or SC7000 series camera with multiple CNUCs, this utility can be used to manage which CNUCs are loaded on the camera. This functionality is not integrated in to the ResearchIR camera controller for these cameras.

7 Acknowledgements^[taw6]

The authors would like to acknowledge the Independent JPEG Group for their JPEG library used by this software for JPEG file support. The authors would also like to acknowledge the ZLib and LibPNG projects for their libraries used by this software for PNG file support.

The authors would also like to acknowledge the libtiff project; this software uses libtiff (<http://www.libtiff.org>) for TIFF file support. libtiff is subject to the following copyright:

Copyright (c) 1988-1997 Sam Leffler

Copyright (c) 1991-1997 Silicon Graphics, Inc.

Permission to use, copy, modify, distribute, and sell this software and its documentation for any purpose is hereby granted without fee, provided that (i) the above copyright notices and this permission notice appear in all copies of the software and related documentation, and (ii) the names of Sam Leffler and Silicon Graphics may not be used in any advertising or publicity relating to the software without the specific, prior written permission of Sam Leffler and Silicon Graphics.

THE SOFTWARE IS PROVIDED "AS-IS" AND WITHOUT WARRANTY OF ANY KIND, EXPRESS, IMPLIED OR OTHERWISE, INCLUDING WITHOUT LIMITATION, ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.

IN NO EVENT SHALL SAM LEFFLER OR SILICON GRAPHICS BE LIABLE FOR ANY SPECIAL, INCIDENTAL, INDIRECT OR CONSEQUENTIAL DAMAGES OF ANY KIND, OR ANY DAMAGES WHATSOEVER RESULTING FROM LOSS OF USE, DATA OR PROFITS, WHETHER OR NOT ADVISED OF THE POSSIBILITY OF DAMAGE, AND ON ANY THEORY OF LIABILITY, ARISING OUT OF OR IN CONNECTION WITH THE USE OR PERFORMANCE OF THIS SOFTWARE.