

P/N: T911987

Copyright

© 2021, FLIR Systems, Inc.

All rights reserved worldwide. Names and marks appearing herein are either registered trademarks or trademarks of FLIR Systems and/or its subsidiaries. All other trademarks, trade names or company names referenced herein are used for identification only and are the property of their respective owners.

Document identity

Publ. No.: T911987

Commit: 77572

Language:

Modified: 2021-06-24

Formatted: 2021-06-24

Website

<http://www.flir.com>

Customer support

<http://support.flir.com>

Disclaimer

Specifications subject to change without further notice. Camera models and accessories subject to regional market considerations. License procedures may apply. Products described herein may be subject to US Export Regulations. Please refer to exportquestions@flir.com with any questions.



General description

Use the Acoustic camera tester to check the accuracy of the FLIR Si124 Acoustic imaging camera. The tester generates a sound signal that is recognized and evaluated by the camera. The dB-level of the signal can be used to ensure a correct level of readings and also the condition of the microphone array.

The Acoustic camera tester is used together with the charger delivered with the FLIR Si124 Acoustic imaging camera.

Physical data

Weight	0.285 kg (0.63 lb)
Size (L x W x H)	150 x 80 x 50 mm (5.9 x 3.1 x 2.0 in)
Tripod mounting	Quick release mount

Shipping information

Packaging, type	Cardboard box
Packaging, contents	<ul style="list-style-type: none"> Acoustic camera tester Table tripod Documentation
Packaging, weight	1.70 kg (3.75 lb)
Packaging, size	390 x 290 x 300 mm (15.4 x 11.4 x 11.8 in)
EAN-13	7332558028612
UPC-12	845188025588
Country of origin	Finland

Compatible with the following products:

- T911970; FLIR Si124
- T912113; FLIR Si124-LD
- T912114; FLIR Si124-PD