

# Manual Supplement

Manual Title: 1750 Operators  
Print Date: October 2006  
Revision/Date: 4, 7/11

Supplement Issue: 5  
Issue Date: 10/14  
Page Count: 2

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 1750  
CD Rev. & Date: 13, 3/2012  
CD PN: 2774701

## Change #1

On page 4, move the first bullet item to the **Standard Accessories** table:

CS1750/1760 Case Polyester, for Fluke 1750 & 1760	2673962
---	---------

On page 4, Table 2:

Change:

Personal Digital Assistant (PDA) with USB cable and accessories	2386780
---	---------

To:

Android based tablet computer- manufacturer and model may vary.	Contact Fluke
---	---------------

## Change #2

On page 9, delete the **Installing the Software** and **Installing Power View on the PDA** sections and replace with:

For complete instructions on how to install the Android tablet computer, go to the **Power Quality Software Downloads** page at [www.Fluke.com](http://www.Fluke.com).

## Change #3

On page 4, Table 2, remove the following:

Sets of international ac power plug adapters for the PDA charger	2583479
--	---------


## Change #4

On page 48, under **Regulatory Information for Wireless Communication**, in the **Cautions**, replace the third bullet with:

- FCCID: See rear decal on Fluke 1750 for the appropriate ID.

## Change #5, 65922

On page 4, replace the **CAT III** and **CAT IV** with the following and add:

	Conforms to relevant South Korean EMC Standards
<b>CAT III</b>	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
<b>CAT IV</b>	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

On page 49, under **General Specifications** remove the **Conformance** section.

On page 52, under **Environmental and Safety Specifications** remove **Standards Conformance**, **EMC** and **Safety** and replace with:

Safety .....	IEC 61010-1: Installation 300 V CAT II, Measurement 600 V CAT IV / 1000 V CAT III, Pollution Degree 2
Altitude.....	2000 m
IP Rating.....	IEC 60529: IP30
Electromagnetic Environment.....	IEC 61326-1: Basic
Electromagnetic Compatibility .....	Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

## Change #6, 48

On page 5, replace the 5<sup>th</sup> bullet with:

- Flexible Current Transformers 3210-PR-TF-II 1000 Amp, Flex CT 24 inch, 3310-PR-TF-II 5000 Amp, Flex CT 24 inch, 3312-PR-TF-II 5000 Amp, Flex CT 48 inch.

On pages 4 and 5 remove the following bullets:

- 4006 RPM, Security CBL, Steel, 6 ft. Looped Lockable
- 4003 RPM, Shipping Caddy with Wheels Power Recorder
- 4005 RPM, Shipping Container, Reusable
- Interface and Extension Cables 3570 CT Cable