

Manual Supplement

Manual Title:	123B/124B/125B Users	Supplement Issue:	1
Print Date:	January 2016	Issue Date:	1/16
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 257

On page 54, replace step 9 with:

9. Push **ENTER** and use **▲▼** and **ENTER** to set the percentage of deviation for time (relative to 1 period) and the amplitude deviation (relative to peak-to-peak) for waveforms. Threshold can be used for signals up to 4 kHz.