

Manual Supplement


Manual Title: 175, 177 & 179 Users Supplement Issue: 1
Part Number: 2063257, 2063269, 2063233, 2063240 Issue Date: 7/10
Print Date: May 2003 Page Count: 2
Revision/Date: 1, 10/08

This supplement contains information necessary to ensure the accuracy of the above manuals.




Change #1, 55375

On page ii, remove the following from the Symbols table:

	Inspected and licensed by TÜV (Technischer Überwachungs Verein) Product Services
---	--

On page 11, under the **Specifications**, change:

From: **Certifications:** CSA, TÜV (EN61010), UL, **CS**,  VDE

To: **Certifications:** CSA UL, **CS**,  VDE

Change #2, 55420

On page 14, add the following table to the bottom of the page:

Frequency Counter Sensitivity					
Input Range ^{[1],[2]}	Typical Sensitivity (RMS Sine Wave)				
	2 Hz to 45 Hz	45 Hz to 10 kHz	10 kHz to 20 kHz	20 kHz to 50 kHz	50 kHz to 100 kHz
Volts AC	600 mV	80 mV	150 mV	400 mV	Unspecified ^[3]
	6 V	0.6 V	1.0 V	2.8 V	Unspecified ^[3]
	60 V	3.8 V	4.1 V	5.6 V	9.6 V
	600 V	50 V	39 V	45 V	58 V
	1000 V	500 V	320 V	380 V	NA
Volts DC	6 V	0.75 V	1.4 V	4.0 V	Unspecified ^[3]
	60 V	3.8 V	4.3 V	6.6 V	13 V
	600 V	40 V	39 V	45 V	58 V
	1000 V	500 V	320 V	380 V	NA
AC/DC Amps	mA	4 mA	4 mA	4 mA ^[4]	NA
	A	0.4 A	0.4 A	0.4 A ^[4]	NA

[1] Maximum input for specified accuracy = 10X Range or 1000 V.

[2] Noise at low frequency and amplitude may exceed the frequency accuracy specification.

[3] Unspecified but usable depending on quality and amplitude of signal.

[4] In mA and A ranges, frequency measurement is specified to 30 kHz.