



EU DECLARATION OF CONFORMITY

Product Identification:

First declared: 2012

Fluke 3000 FC - Meter Family

Including Models: 3000, v3000, v3001, i3000, a3001, a3002, t3000

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Directive 2014/53/EU, Radio Equipment (RED)

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits
EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors
EN 300 328 V1.7.1 Data transmission operation in 2.4GHz ISM band
EN 301 489-1 V1.9.2 Radio equipment and services
EN 301 489-17 V2.1.1 Broadband Data transmission systems

Special Conditions:

None

Manufacturer

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160707131710 - This declaration supersedes all previous declarations for this product.

07 July 2016