

Supplement















Title: 789 Safety Info. Supplement Issue: **2**
Print Date: August 2012 Issue Date: 6/15
Part Number: 4276668 Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 166

Under **Safety Specifications**, replace the **Symbols** table with:

Symbols

Symbol	Description
	WARNING. RISK OF DANGER.
	Conforms to European Union directives.
	Consult user documentation.
	Certified by TÜV SÜD Product Service.
	WARNING. HAZARDOUS VOLTAGE. Risk of electric shock.
	Conforms to relevant Australian EMC standards.
	Fuse
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.
	Conforms to relevant South Korean EMC Standards.
	Certified by CSA Group to North American safety standards.
	Double Insulation
	AC (Alternating Current)
	DC (Direct Current)
	Earth
	This product complies with the WEEE Directive marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste. Product Category: With reference to the equipment types in the WEEE Directive Annex I, this product is classed as category 9 "Monitoring and Control Instrumentation" product. Do not dispose of this product as unsorted municipal waste.