



CERTIFICATE

No. Z1 14 06 14544 107

Holder of Certificate: Fluke Corporation

6920 Seaway Boulevard Everett, WA 98203-9090

USA

Production Facility(ies):

14544, 56588

Certification Mark:



Product: Multimeters

(Digital Process Meter)

Model(s): 789

Parameters: Rated Input Voltage: 6 VDC (4 x 1.5V AA batteries IEC LR6)

Measuring Ratings:

1000V CAT III, 600V CAT IV

Protection Class: IP Protection: Reinforced Insulation

Protection:

IPX0

Temperature Ambient: -20°0

-20°C to +55°C

Tested according to: EN 61010-1:2010

EN 61010-2-033:2012

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

PI1305384103-100

Date, 2014-07-08

Page 1 of 1



