

CALIBRATION CERTIFICATE
PRESSURE RANGE AS LEFT

UNIT UNDER TEST (UUT)

Manufacturer : Druck
 Type Number : DPI 104 IS
 Serial Number : 3413543
 Sales Order Number : W5129
 Parameter Range : 0 to 300 psi g
 Calibration Date : 11/30/11
 Calibrated By : SJH

CALIBRATOR INFORMATION

Calibration Instrument : Ruska 7250xi
 Serial Number : PG-301g (s/n:62445)
 (*1) Calibrated Against : prs-003-7.6cert100820V-1368
 Pressure Medium : Nitrogen
 Uncertainty : 77.66 ppm

AMBIENT CONDITIONS

Ambient Temperature (°C) : 24.13

PERFORMANCE DATA

Actual Applied Value psi	Unit Under Test Reading psi (*2)	Unit Under Test Deviation (*3)	Permissible Deviation (± 0.05 % fs) (*4)
0.00	0.00	+0.000 % fs	± 0.050 % fs
60.00	60.01	+0.003 % fs	± 0.050 % fs
120.00	120.01	+0.003 % fs	± 0.050 % fs
180.00	180.01	+0.003 % fs	± 0.050 % fs
240.00	240.01	+0.003 % fs	± 0.050 % fs
300.00	300.01	+0.003 % fs	± 0.050 % fs
150.00	150.01	+0.003 % fs	± 0.050 % fs
0.00	0.00	+0.000 % fs	± 0.050 % fs

COMMENTS

Certified by:

Date:

NOTES

- (*1) Traceable to NIST. Calibrated in accordance with ISO9000 Quality Standards
- (*2) Actual recorded values. For specification, see Permissible Deviation column.
- (*3) Deviation calculated from UUT Reading minus Actual Applied Value.
- (*4) Non linearity, hysteresis and repeatability.