

Bibby Scientific Ltd Beacon Road Stone Staffordshire ST15 0SA United Kingdom

Tel +44 (0) 1785 812121 Fax +44 (0) 1785 810405 www.bibby-scientific.com

Registered office: 4 Felsted Gardens, Ferry Street London E14 3BS Registered No.6381141 BS EN ISO 9001:2001 registered company

Datasheet

Pure Water Conductivity Probe with ATC Jenway Part Number 027 900



This probe is manufactured from PVC and stainless steel and is designed for the measurement of purified water.

Specifications:

Parameter	Specification
Cell Constant	K = 0.086 ±10%
Recommended	0 - 200µS
Measurement Range	0-200μ3
Temperature Range	0-50°C
Temperature	Thermistor 10 K @ 25°C
Compensation	
Body Materials	PVC
Plate Material	316 stainless steel
Dimensions	120x12.7mm
Cable Length	1 metre (nominal)
Termination	7 pin DIN

This datasheet is meant for your guidance only.



Page 1 of 1 Datasheet Ref. 027 900 Dec 2010/REV B