

## **Bibby Scientific Ltd**

Beacon Road Stone Staffordshire ST15 0SA United Kingdom

Tel +44 (0) 1785 812121 Fax +44 (0) 1785 810405 www.bibby-scientific.com

Registered office: 4 Felsted Gardens, Ferry Street London E14 3BS Registered No.6381141 BS EN ISO 9001:2001 registered company

## **Datasheet**

## General Purpose Glass Bodied pH Electrode Jenway Part Number 924 015



This electrode is designed for standard measurements of general solutions. It is designed for general purpose pH measurement for accurate Q.C. and research work. A suitable reference electrode is required to perform pH measurements.

## Specifications:

Parameter	Specification
Туре	Non-combination pH electrode
Sealed/Refillable	Sealed
pH Range	0-14
Temperature Range	0-80°C
Body Material	Glass
Dimensions	120x12mm
Cable Length	1 metre (nominal)
Termination	BNC
Isopotential	0±20mV
Slope	56.0 to 62.7mV/decade @ 25°C
Response Time	95% of final reading in 10 seconds
Stability	<1mV(between 1 & 2 minutes)
Impedance	<300MΩ @25°C

This Datasheet is meant for your guidance only.

