

## **Bibby Scientific Ltd**

Beacon Road Stone Staffordshire ST15 0SA United Kingdom

Tel +44 (0) 1785 812121 Fax +44 (0) 1785 810405 www.bibby-scientific.com

Registered office: 4 Felsted Gardens, Ferry Street London E14 3BS Registered No.6381141 BS EN ISO 9001:2001 registered company

## **Datasheet**

## General Purpose Epoxy Bodied pH Electrode Jenway Part Number 924 067



This electrode is designed for standard measurements of general solutions. The epoxy body makes it especially suitable for field measurements or use in education.

## Specifications:

Parameter	Specification
Туре	Combination
Sealed/Refillable	Sealed
End Type	Bulb (with protective cover)
pH Range	0-14
Temperature Range	0-60°C
Reference Type	AgCl
Junction Type	Porous Teflon
Body Material	Epoxy
Dimensions	120x12mm
Cable Length	Not applicable
Termination	S7
Isopotential	0±20mV
Slope	59.16mV±3mV @ 25°C
Response Time	95% of final reading in 10 seconds
Stability	<1mV(between 1 & 2 minutes)
Impedance	<300MΩ @25°C

Incompatibilities – Tris Buffers

Low Ionic Strength Solutions - use 924 068

This electrode is supplied with Jenway combined pH/conductivity meter - 430.

This datasheet is meant for your guidance only.

