

ACCUBLOCK™ MINI DIGITAL DRY BATH

User Manual



D0100
D0100-230V



Lit M00076

Labnet

Rev 2, September 2016

About This Manual

This manual is designed to assist you in the optimal usage of your Accublock Mini Dry Bath. The manual is available in English, French, German, Italian, Portuguese, and Spanish on our website at:

www.labnetinternational.com/document-center

1.0 General Specifications

| | |
|--------------------|------------------------------------|
| Temperature range | Ambient +5° to 100°C |
| Temp. increments | 0.1°C |
| Temp. uniformity | ≤0.2°C |
| Temp. accuracy | ≤0.2°C |
| Timer | 0 -19h:59 m or continuous |
| Block | Anodized Aluminum |
| Block capacity | 12x 1.5mL tubes |
| Dimensions (WxDxH) | 5.5 x 4.8 x 2.4 in./14 x 12 x 6 cm |
| Weight | 3 lb/1.3 kg |
| Electrical | 18 VDC 1.66 A, 30 W max |

Electrical requirements:

100-240V, 50/60Hz, 0.7A – 30W Max

2.0 Safety Precautions:

- **Do not** use this product in an explosive environment
- **Do not** use in the presence of flammable or combustible material
- **Do not** heat substances that react violently when heated
- **Do not** touch area around block or block well when unit is hot.
- **Do not** spill liquids into the well area or into the unit
- **Connect unit** only to a properly grounded outlet

Caution: HOT Surfaces that can burn or cause injury.

Use of this product in any manner not specified by the manufacturer or modification of the product may cause injury and/or may void the warranty.

3.0 Installation

Upon unpacking Mini Digital Dry Bath, inspect for damages. Shipping damage is the responsibility of the carrier. Inspect that the following are present:

AccuBlock™ Mini with AccuRack™

Users manual

Switching Power Adapter

Line cord

Select a location that is dry and not subject to drafts or moving air from heating or air conditioning vents, or air blown by other equipment. Place the unit on a flat, preferably non-flammable surface. Allow sufficient room around the unit for access and cooling. Six inches minimum on all sides is suggested. Plug the unit into a properly grounded outlet. Using the lifter, insert the block(s) into the well. The unit is now ready for use.

4.0 Controls

| | |
|--------------------------|--|
| LCD Display | (Block Temperature, Set Temperature, Timer) |
| Heat LED | (Heater Function) |
| Main Power Switch | Turns entire unit on and off. |
| START/STOP Switch | Begins and ends the heating process |
| Mode Switch | Toggles the LCD display between Temperature and Timer Settings |
| Up Switch | Used to increment the set temperature. |
| Down Switch | Used to decrement the set temperature. |

5.0 Operation

1. Turn the unit on using the Main Power rocker switch on the rear of the unit.
2. The LCD will display a self test and the LED will turn on.
3. After two seconds the unit will enter normal operation and display the current temperature of the block.
4. Press the Up or Down switches to set the block to the desired temperature. The display will change to the current temperature set point. The temperature is set in 0.1 degree C increments. Holding the Up or Down key for more the eight seconds will increase or decrease the set point rapidly. The display will automatically return to current temperature four seconds after the switch is released. Note, the temperature may not be adjusted with the unit actively heating the block.
5. The temperature setting will be stored in memory when the display returns to the current block temperature. This setting will be retained if the power is turned off. Note the display must return to current temperature before the power is turned off or the new set temperature will NOT be saved.
6. Press the START/STOP to begin the temperature control. The “^” will be displayed on the LCD when the unit is in active temperature control.
7. The LED will turn on and off with the heating element.
8. The LED will be off when the unit is turned off and when the heater is not actively heating the block.
9. The maximum set temperature is 101C.
10. The Mode switch will scroll through current temperature, the elapsed time from the press of the START button, LCD off and countdown timer.

Heating & Countdown Timer

1. Turn the unit on using the Main Power rocker switch.
2. Press the Up or Down switches to set the block to the desired temperature. Holding the Up or Down key for more the eight seconds will increase or decrease the set point rapidly. The display will automatically return to current temperature four seconds after the switch is released. Note, the temperature may not be adjusted with the unit running.
3. Press the Mode button until “1:00” is displayed. This is the default Count Down timer setting. The maximum time is setting is 19:59.
4. Press the up or down key to set the desired count down time.
5. Press Start/Stop button to begin heating.
6. The LCD will alternate between actual temperature and time remaining.
7. When time runs out the LCD will alternate between the current temperature and “---“. The beeper will sound when “---“ is displayed.
8. Heating will NOT turn off when time expires.
9. To cancel the countdown timer press the Mode key one time to advance to the standard temperature only read out.

6.0 Calibration

1. Press and hold the Start/Stop then simultaneously powering up the unit with the ON/OFF switch.
2. Release the Start/Stop button when the LCD displays “---“.
3. The will beep 4 times and the display will show the calibration set point and the “:” will flash rapidly.
4. Press the UP or Down keys to set the desired temperature at which you want to calibrate the unit.

5. Press the Start/Stop to begin calibration. The “^” will be displayed on the LCD when the unit is in active temperature control.
6. Allow 10 – 40 minutes for the unit to heat up and reach equilibrium. A high temperature calibration set point will require more time.
- 7.
8. After the unit has equalized the LCD will flash “---“and the current temperature.
9. Read the thermometer (or sensor meter) and use the UP and Down keys to adjust the display to the thermometer or sensor reading.
10. Press the Start/Stop key to save the new settings.
11. Press the Start/Stop key again. The unit will adjust to the operating temperature with the new calibration factor. Allow sufficient time for the unit to re-equilibrate and then compare the thermometer or sensor reading to the calibrated display. They should closely match (2+/- 0.2 C). If not, repeat calibration procedure.

7.0 Troubleshooting Guide / Service (Labnet Service 1-888-LABNET-1)

Problem Explanation / Solution

Display / LEDs do not light up

1. Check power cord & outlet.
2. Check ON / OFF switch.
3. Check for Light on Power Adapter.
4. Call service.

Unit not heating

1. Is set point below room temp?
2. Is Heating LED turning Off and On?
3. Press “Start/Stop” key.
4. Is the “^”displayed on the LCD?
5. Call service.

Block or Sample temp not same as display temp.

1. Is unit in heating mode?
2. Is unit sitting in draft?
3. Check accuracy of thermometer.
4. Is reference thermometer making good contact?
5. Follow calibration procedure

8.0 Cleaning and Maintenance

Make sure that the dry bath and block are cool and the power cord is disconnected before performing any cleaning or maintenance. Repair or maintenance should only be performed by an authorized service technician.



The dry bath may be cleaned with a moist cloth containing a mild soap solution. Do not immerse the dry bath in water or any liquid.

Spills: In the event liquid is accidentally spilled into the block area, disconnect the plug from the outlet and turn the unit upside down to minimize liquid contact with the internal components. Allow the unit to dry. Have qualified service technician clean the unit and replace any damaged parts.

9.0 Appendix

9.1 Symbols and Conventions

The following chart is an illustrated glossary of the symbols that may be used in this manual or on the product.

| | |
|---|---|
|  | CAUTION This symbol refers you to important operating and maintenance (servicing) instructions within the product Instruction Manual. Failure to heed this information may present a risk of damage or injury to persons or equipment. |
|  | WARNING: To avoid accidental burning be very careful touching the heater block of the unit. It can be very hot after it is used at high temperatures. Allow the heater block to cool down before handling. |

EQUIPMENT DISPOSAL-EUROPEAN REGULATIONS



According to Directive 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE), Accublock Mini Dry Bath is marked with the crossed-out wheeled bin and must not be disposed of with domestic waste.

Consequently, the buyer shall follow the instructions for reuse and recycling of waste electronic and electrical equipment (WEEE) provided with the products and available at the following link:

www.corning.com/weee

LIMITED WARRANTY

Labnet International Inc. warrants that this product will be free from defects in material and workmanship for a period of one (1) year from date of purchase. This warranty is valid only if the product is used for its intended purpose and within the guidelines specified in the supplied instruction manual.

Should this product require service, contact Labnet's Service department at 732-417-0700 to receive a return authorization number and shipping instructions. Products received without proper authorization will be returned. All items returned for service should be sent postage prepaid in the original packaging or other suitable carton, padded to avoid damage. Labnet International Inc. will not be responsible for damage incurred by improper packaging. Labnet may elect for onsite service for larger equipment.

This warranty does not cover damage caused by accident, neglect, misuse, improper service, natural forces or other causes not arising from defects in original material or workmanship. This warranty does not cover motor brushes, fuses, light bulbs, batteries or damage to paint or finish. Claims for transit damage should be filed with the transportation carrier.

ALL WARRANTIES INCLUDING THE IMPLIED WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE LIMITED IN DURATION OF 12 MONTHS FROM THE ORIGINAL DATE OF PURCHASE.

LABNET'S SOLE OBLIGATION UNDER THIS WARRANTY IS LIMITED TO THE REPAIR OR REPLACEMENT, AT LABNET INTERNATIONAL INC.'S DISCRETION, OF A DEFECTIVE PRODUCT. LABNET INTERNATIONAL, INC. IS NOT LIABLE FOR INCIDENTAL OR CONSEQUENTIAL DAMAGE, COMMERCIAL LOSS OR ANY OTHER DAMAGES RESULTING FROM THE USE OF THIS PRODUCT.

Some states do not allow limitation on the length of implied warranties or the exclusion or limitation of incidental or consequential damages. This warranty gives you specific legal rights. You may have other rights which vary from state to state.

No individual may accept for, or on behalf of Labnet International Inc., any other obligation of liability, or extend the period of this warranty.

***Please register your product online at:
www.labnetinternational.com***

Warranty/Disclaimer: Unless otherwise specified, all products are for research use only. Not intended for use in diagnostic or therapeutic procedures. Labnet International makes no claims regarding the performance of these products for clinical or diagnostic applications.



31 Mayfield Ave. Edison, NJ 08837
P: +732-417-0700 F: +732-417-1750
Toll Free: 888-522-6381
labnetinfo@corning.com
www.labnetinternational.com

9300110000